

TENDER ENQUIRY DOCUMENT
FOR PURCHASE OF
MEDICAL EQUIPMENT

FOR AND ON BEHALF OF
EMPLOYEE STATE INSURANCE CORPORATION
UNDER THE ADMINISTRATIVE CONTROL OF
MINISTRY OF LABOUR AND EMPLOYMENT
GOVT. OF INDIA

HLL/PCD/ESIC-51/10-11



BY

HLL Lifecare Limited

(A GOVT. OF INDIA ENTERPRISE & Formerly HINDUSTAN LATEX LTD.)

Procurement & Consultancy Services Division

B-14A, Sector-62, Noida-201 307, Uttar Pradesh, India

Ph: 0120-4071500; Fax: 0120-4071513

URL: www.lifecarehll.com

Email: pcd@lifecarehll.com

INDEX

Section	Topic	Page No.
Section I	– Notice inviting Tender (NIT) -----	03
Section II	– General Instructions to Tenderers (GIT) -----	07
Section III	– Special Instructions to Tenderers (SIT) -----	24
Section IV	– General Conditions of Contract (GCC) -----	25
Section V	– Special Conditions of Contract (SCC) -----	40
Section VI	– List of Requirements -----	41
Section VII	– Technical Specifications -----	44
Section VIII	– Quality Control Requirements -----	91
Section IX	– Qualification Criteria -----	92
Section X	– Tender Form -----	94
Section XI	– Price Schedules -----	95
Section XII	– Questionnaire -----	99
Section XIII	– Deleted	
Section XIV	– Manufacturer’s Authorisation Form -----	100
Section XV	– Bank Guarantee Form for Performance Security /CMC Security -----	101
Section XVI	– Contract Form (A & B) -----	102
Section XVII	– Proforma of Consignee Receipt Certificate -----	106
Section XVIII	– Proforma of Final Acceptance Certificate by the Consignee -----	107
Section XIX	– Details of Shipping arrangement for Liner Cargoes in respect of C&F/CIF/ TurnkeyF.O.R. Contracts for Import-----	109
Section XX	– Check List for the Tenderers -----	113
Section XXI	– Consignee address -----	116

SECTION I**SHORT TERM TENDER NOTICE****For Tender from****HLL LIFECARE LIMITED**

(A GOVERNMENT OF INDIA ENTERPRISE)

Procurement & Consultancy Services Division

B-14A, Sector-62, Noida-201 307, Uttar Pradesh, India

Ph: 0120-4071500; Fax: 0120-4071513

Email: pcd@lifecarehll.com**1. Tender Enquiry No. HLL/PCD/ESIC-51/10-11****Date: 04.03.2011**

Procurement & Consultancy Services Division of HLL Lifecare Limited (Formerly Hindustan Latex Limited), for and on behalf of Director General of Employee State Insurance Corporation (ESIC), invites sealed tenders, from eligible and qualified tenderers for supply of following Medical Equipment to ESI Hospital at **Bapunagar (Ahmedabad)**.

Sl. No.	Item Description	Total Qty.	EMD Amount (Rs.)
1	Abdominal hysterectomy set	2	4,000
2	Ambulatory BP monitor	4	16,000
3	Automated Visual Field Analyzer with printer (Perimeter)	1	4,000
4	Automatic film processor	1	10,000
5	Baby warmer	4	4,000
6	Bilirubinometer	1	6,000
7	Bio-chemistry semi auto analyser	1	3,000
8	BIPAP	1	6,000
9	Biphasic Defibrillator	5	35,000
10	Blood cell counter-Fully Automated 3 Part Analyser	1	9,000
11	Blood gas analyser	1	16,000
12	Coagulation Analyser	1	10,000
13	Combined A & B Scan Machine	1	4,000
14	CPM	1	9,500
15	CR system (Computer Radiography System)	1	30,000
16	Delivery & caesarean set	2	12,000
17	Dental Unit Complete	1	4,000
18	Diagnostic X-ray Machine 300m.A	1	10,000
19	Diognostic X-ray Machine 500 m.A	1	18,000
20	ECG machine 12 channel	2	6,000
21	Electrolyte analyser	1	3,000
22	Emergency resucitation Kit	2	22,000
23	Fully automated ELISA reader with washer	1	18,000
24	Fully Automated random access biochemistry analyser	1	40,000
25	Holter Recorder & Analyser	1	16,000

Sl. No.	Item Description	Total Qty.	EMD Amount (Rs.)
26	Infant Resuscitator	1	10,000
27	Laryngoscope (Fibreoptic)	1	6,000
28	MOBILE C-ARM IMAGE INTENSIFIER WITH DSA	1	50,000
29	Multipara bed side monitors with Central station	10(1)	90,000
30	Neonatal monitor	2	16,000
31	Oesophagoscope Adult	1	5,500
32	Operating Zoom microscope for ophthalmology	1	40,000
33	Patient Warming system	1	10,000
34	Phaco Emulsification Machine	1	20,000
35	Pharmaceutical Refrigerator for OT	3	9,000
36	Portable ventilator	1	10,000
37	Portable X-ray machine (Mobile)	2	12,000
38	Pulse oximeter	7	7,000
39	Slit Lamp	1	6,000
40	Suction machine(portable)	2	6,000
41	Syringe infusion pump	5	6,000
42	TMT	1	24,000
43	Traction unit	1	18,000
44	UVB Chamber	1	10,000
45	Vacuum extractor	1	9,000
46	Vaginal hysterectomy set	2	10,000
47	Volumetric infusion pump	7	8,400
48	800mA X-ray machine with fluroscopy with IITV	1	32,000
49	O T Tables for General Surgery & Gynaecology	5	2,20,000
50	O T Tables for Orthopaedics	1	1,00,000

2. Tender No.: HLL/PCD/ESIC-51/10-11

Sl No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	05.03.2011 to 24.03.2011, in all working days, during 10:00am to 4:00pm
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Limited Procurement & Consultancy Services Divn. B-14A, Sector-62, Noida -201 307
iii.	Cost of the Tender Enquiry Document	Rs. 3,000.00/ USD 75.00
iv.	Time and date of Pre-bid meeting	11.03.2011, 11:00am
v.	Closing date & time for receipt of Tender	25.03.2011, 2:00pm

Sl No.	Description	Schedule
vi.	Time and date of opening of Techno-Commercial tenders	25.03.2011, 2:30pm
vii.	Venue for Pre-bid Meeting & Techno-Commercial Tender Opening	Same as given in 2 (ii)

3. Interested tenderers may obtain further information about this tender from the office of Head (P&CD), HLL Lifecare Ltd., Noida. Tender Enquiry Documents may be purchased on payment of non-refundable fee of Rs. 3,000.00/ USD 75.00 per set in the form of account payee Demand Draft/Pay Order/Cashier's Cheque/Banker's Cheque, drawn on a scheduled bank in India, in favour of "**HLL Lifecare Limited**" payable at New Delhi.
4. If requested, the Tender Enquiry Documents will be mailed by Registered Post/Speed Post to the domestic tenderers and by international airmail to the foreign tenderers, for which extra expenditure per set will be Rs 100.00 for domestic post and USD 50.00 for international airmail. The tenderer is to add the applicable postage cost in the non-refundable fee mentioned in Para 3 above. However, HLL Lifecare Ltd. shall not be responsible for any postal loss/delay.
5. Tenderer may also download the tender enquiry documents from the web site www.esic.nic.in or www.lifecarehll.com and submit its tender by utilizing the downloaded document, along with the required non-refundable fee as mentioned in Para 3 above.
6. All prospective tenderers may attend the Pre Tender meeting. The venue, date and time indicated in the Para 2 above
7. Tenderers shall ensure that their tenders, complete in all respects, are dropped in the Tender Box located at **HLL Lifecare Limited, Procurement and Consultancy Division, B-14A, Sector-62, Noida -201307, Uttar Pradesh** on or before the closing date and time indicated in the Para 2 above, failing which the tenders will be treated as late tender and rejected. The tenders sent by post/ courier must reach the above said address on or before the closing date & time indicated in Para 2 above, failing which the tenders will be treated as late tender and rejected.
8. In the event of any of the above mentioned dates being declared as a holiday / closed day for the purchase organisation, the tenders will be sold/received/opened on the next working day at the appointed time.
9. The Tender Enquiry Documents are not transferable.
10. All Tenders must be accompanied by EMD as mentioned against each item. Tenders without EMD shall be rejected.

Head (P & CD)
HLL Lifecare Limited,
Procurement and Consultancy Division
B-14A, Sector -62, Noida -201307,
Uttar Pradesh.

SECTION - II**GENERAL INSTRUCTIONS TO TENDERERS (GIT)****CONTENTS**

Sl. No.	Topic	Page No.
A	PREAMBLE	
1	Definitions and Abbreviations	9
2	Introduction	10
3	Deleted	--
4	Language of Tender	11
5	Eligible Tenderers	11
6	Eligible Goods and Services	11
7	Tendering Expense	11
B	TENDER ENQUIRY DOCUMENTS	
8	Contents of Tender Enquiry Documents	11
9	Deleted	--
10	Clarification of Tender Enquiry Documents	12
C	PREPARATION OF TENDERS	
11	Documents Comprising the Tender	12
12	Tender Currencies	13
13	Tender Prices	13
14	Indian Agent	15
15	Firm Price / Variable Price	16
16	Deleted	--
17	Documents Establishing Tenderer's Eligibility and Qualifications	16
18	Documents Establishing Good's Conformity to Tender Enquiry Document	16
19	Earnest Money Deposit (EMD)	17
20	Tender Validity	17
21	Signing and Sealing of Tender	17
D	SUBMISSION OF TENDERS	
22	Submission of Tenders	18
23	Late Tender	18
24	Alteration and Withdrawal of Tender	18

E	TENDER OPENING	
25	Opening of Tenders	19
F	SCRUTINY AND EVALUATION OF TENDERS	
26	Basic Principle	19
27	Preliminary Scrutiny of Tenders	19
28	Deleted	--
29	Discrepancies in Prices	20
30	Discrepancy between original and copies of Tender	20
31	Qualification Criteria	20
32	Conversion of Tender Currencies to Indian Rupees	20
33	Deleted	--
34	Comparison of Tenders	21
35	Additional Factors and Parameters for Evaluation and Ranking of Responsive Tenders	21
36	Tenderer's capability to perform the contract	21
37	Contacting the Purchaser	21
G	AWARD OF CONTRACT	
38	Purchaser's Right to Accept any Tender and to Reject any or All Tenders	21
39	Award Criteria	22
40	Variation of Quantities at the Time of Award	22
41	Notification of Award	22
42	Issue of Contract	22
43	Non-receipt of Performance Security and Contract by the Purchaser/Consignee	22
44	Return of EMD	22
45	Publication of Tender Result	23
46	Corrupt or Fraudulent Practices	23

A. PREAMBLE

1. Definitions and Abbreviations

1.1 The following definitions and abbreviations, which have been used in these documents shall have the meanings as indicated below:

1.2. Definitions:

- (i) “Purchaser” means the organization purchasing goods and services as incorporated in the Tender Enquiry document.
- (ii) “Tender” means Bids / Quotation / Tender received from a Firm / Tenderer / Bidder.
- (iii) “Tenderer” means Bidder/ the Individual or Firm submitting Bids / Quotation / Tender
- (iii) “Supplier” means the individual or the firm supplying the goods and services as incorporated in the contract.
- (iv) “Goods” means the articles, material, commodities, livestock, furniture, fixtures, raw material, spares, instruments, machinery, equipment, medical equipment, industrial plant etc. which the supplier is required to supply to the purchaser under the contract.
- (v) “Services” means services allied and incidental to the supply of goods, such as transportation, installation, commissioning, provision of technical assistance, training, after sales service, maintenance service and other such obligations of the supplier covered under the contract.
- (vi) “Earnest Money Deposit” (EMD) means Bid Security/ monetary or financial guarantee to be furnished by a tenderer along with its tender.
- (vii) “Contract” means the written agreement entered into between the purchaser and/or consignee and the supplier, together with all the documents mentioned therein and including all attachments, annexure etc. therein.
- (viii) “Performance Security” means monetary or financial guarantee to be furnished by the successful tenderer for due performance of the contract placed on it. Performance Security is also known as Security Deposit.
- (ix) “Consignee” means the Hospital/Dispensaries/Institute/Medical College/ person to whom the goods are required to be delivered as specified in the Contract. If the goods are required to be delivered to a person as an interim consignee for the purpose of despatch to another person as provided in the Contract then that “another” person is the consignee, also known as ultimate consignee.
- (x) “Specification” means the document/standard that prescribes the requirement with which goods or service has to conform.
- (xi) “Inspection” means activities such as measuring, examining, testing, gauging one or more characteristics of the product or service and comparing the same with the specified requirement to determine conformity.
- (xii) “Day” means calendar day.

1.3 Abbreviations:

- (i) “T E Document” means Tender Enquiry Document
- (ii) “NIT” means Notice Inviting Tenders.
- (iii) “GIT” means General Instructions to Tenderers
- (iv) “SIT” means Special Instructions to Tenderers
- (v) “GCC” means General Conditions of Contract
- (vi) “SCC” means Special Conditions of Contract

- (vii) “DGS&D” means Directorate General of Supplies and Disposals
- (viii) “NSIC” means National Small Industries Corporation
- (ix) “PSU” means Public Sector Undertaking
- (x) “CPSU” means Central Public Sector Undertaking
- (xi) “LSI” means Large Scale Industry
- (xii) “SSI” means Small Scale Industry
- (xiii) “LC” means Letter of Credit
- (xiv) “DP” means Delivery Period
- (xv) “BG” means Bank Guarantee
- (xvi) “ED” means Excise Duty
- (xvii) “CD” means Custom Duty
- (xviii) “VAT” means Value Added Tax
- (xix) “CENVAT” means Central Value Added Tax
- (xx) “CST” means Central Sales Tax
- (xxi) “RR” means Railway Receipt
- (xxii) “BL” means Bill of Lading
- (xxiii) “FOB” means Free on Board
- (xxiv) “FCA” means Free Carrier
- (xxv) “FOR” means Free On Rail
- (xxvi) “CIF” means Cost, Insurance and Freight
- (xxvii) “CIP (Destinations)” means Carriage and Insurance Paid up to named port of destination. Additionally the Insurance (local transportation and storage) would be extended and borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery.
- (xxviii) “DDP” means Delivery Duty Paid named place of destination (consignee site)
- (xxix) “INCOTERMS” means International Commercial Terms as on the date of Tender Opening
- (xxx) ”ESIC” means Employee State Insurance Corporation.
- (xxxi) “DG” means Director General of ESIC under the administrative control of Ministry of Labour and Employment
- (xxxii) “CMC” means Comprehensive Maintenance Contract (labour, spare and preventive maintenance)
- (xxxiii) “RT” means Re-Tender.

2. Introduction

- 2.1 The Purchaser has issued these TE documents for purchase of goods and related services as mentioned in Section – VI – “List of Requirements”, which also indicates, *interalia*, the required quantity, delivery schedule, terms and place of delivery.
- 2.2 This section (Section II - “General Instructions to Tenderers”) provides the relevant information as well as instructions to assist the prospective tenderers in preparation and submission of tenders. It also includes the mode and procedure to be adopted by the purchaser for receipt and opening as well as scrutiny and evaluation of tenders and subsequent placement of contract.
- 2.3 The tenderers shall also read the Special Instructions to Tenderers (SIT) related to this purchase, as contained in Section III of these documents and follow the same accordingly. Whenever there is a conflict between the GIT and the SIT, the provisions contained in the SIT shall prevail over those in the GIT.
- 2.4 Before formulating the tender and submitting the same to the purchaser, the tenderer should read and examine all the terms, conditions, instructions, checklist etc. contained in the TE documents. Failure to provide and/or comply with the required information, instructions etc. incorporated in these TE documents may result in rejection of its tender.

3. Deleted**4. Language of Tender**

4.1 The tender submitted by the tenderer and all subsequent correspondence and documents relating to the tender exchanged between the tenderer and the purchaser, shall be written in the English language, unless otherwise specified in the Tender Enquiry. However, the language of any printed literature furnished by the tenderer in connection with its tender may be written in any other language provided the same is accompanied by a notarised English translation and, for purposes of interpretation of the tender, the English translation shall prevail.

4.2 The tender submitted by the tenderer and all subsequent correspondence and documents relating to the tender exchanged between the tenderer and the purchaser, may also be written in the Hindi language, provided that the same are accompanied by notarised English translation, in which case, for purpose of interpretation of the tender etc, the English translations shall prevail.

5. Eligible Tenderers

This invitation for tenders is open to all suppliers who fulfil the eligibility criteria specified in these documents.

6. Eligible Goods and Services

All goods and related services to be supplied under the contract shall have their origin in India or any other country with which India has not banned trade relations. The term “origin” used in this clause means the place where the goods are mined, grown, produced, or manufactured or from where the related services are arranged and supplied.

7. Tendering Expense

The tenderer shall bear all costs and expenditure incurred and/or to be incurred by it in connection with its tender including preparation, mailing and submission of its tender and for subsequent processing the same. The purchaser will, in no case be responsible or liable for any such cost, expenditure etc regardless of the conduct or outcome of the tendering process.

B. TENDER ENQUIRY DOCUMENTS**8. Content of Tender Enquiry Documents**

8.1 In addition to Section I – “Notice inviting Tender” (NIT), the TE documents include:

- Section II – General Instructions to Tenderers (GIT)
- Section III – Special Instructions to Tenderers (SIT)
- Section IV – General Conditions of Contract (GCC)
- Section V – Special Conditions of Contract (SCC)
- Section VI – List of Requirements
- Section VII – Technical Specifications
- Section VIII – Quality Control Requirements
- Section IX – Qualification Criteria
- Section X – Tender Form
- Section XI – Price Schedules
- Section XII – Questionnaire
- Section XIII – Deleted
- Section XIV – Manufacturer’s Authorisation Form
- Section XV – Bank Guarantee Form for Performance Security/CMC Security
- Section XVI – Contract Forms A & B
- Section XVII – Proforma of Consignee Receipt Certificate

Section XVIII – Proforma of Final Acceptance Certificate by the consignee
 Section XIX – Details of Shipping arrangement for Liner Cargoes in respect of
 C&F/CIF/Turnkey/F.O.R. Contracts for Import
 Section XX – Check List for the Tenderers
 Section XXI – Consignee List

8.2 The relevant details of the required goods and services, the terms, conditions and procedure for tendering, tender evaluation, placement of contract, the applicable contract terms and, also, the standard formats to be used for this purpose are incorporated in the above-mentioned documents. The interested tenderers are expected to examine all such details before submission of the tender.

9. Deleted

10. Clarification of TE documents

A tenderer requiring any clarification or elucidation on any issue of the TE documents may take up the same with the purchaser in writing. The purchaser will respond to such request provided the purchaser receives the same minimum 1 (one) hour prior to the scheduled time of pre-bid meeting. Clarification for the same will be published on the websites as mentioned in para 5 of NIT (Section I) within 4 days after the prebid meeting.

C. PREPARATION OF TENDERS

11. Documents Comprising the Tender

11.1 The **Two Tender System**, i.e. “Techno – Commercial Tender” and “Price Tender” prepared by the tenderer shall comprise the following:

A) **Techno – Commercial Tender (Un priced Tender)**

- i) Earnest money furnished in accordance with GIT clause 19.
- ii) Tender Form as per Section X.
- iii) Documentary evidence, as necessary in terms of clauses 5 and 17 establishing that the tenderer is eligible to submit the tender and, also, qualified to perform the contract if its tender is accepted.
- iv) Tenderer/Agent who quotes for goods manufactured by other manufacturer shall furnish Manufacturer’s Authorisation in the prescribed format as per Section - XIV.
- v) Power of Attorney/Authorisation in favour of signatory of TE documents.
- vi) Documents and relevant details to establish in accordance with GIT clause 18 that the goods and the allied services to be supplied by the tenderer conform to the requirement of the TE documents.
- vii) Performance Statement as per section IX along with relevant copies of purchase orders and end users’ satisfaction certificate.
- viii) Price Schedule(s) as per Section XII filled up with all the details including Qty., Make, Model, Contry of origin, etc. of the goods offered with prices blank (without indicating any prices).
- ix) Certificate of Incorporation of the bidder.
- x) Checklist as per Section XX.
- xi) Statement of deviations parameter wise from tendered technical specifications, if any.

B) **Price Tender:**

The information given at clause no. 11.1 A)- viii) above should be reproduced with the prices indicated.

N.B.

1. All pages of the Tender should be page numbered and indexed.
2. It is the responsibility of tenderer to go through the TE document to ensure furnishing all required documents in addition to above, if any.

- 11.2 The authorized signatory of the tenderer must sign the tender duly stamped at appropriate places and initial all the remaining pages of the tender.
- 11.3 A tender, which does not fulfil any of the above requirements and/or gives evasive information/reply against any such requirement, shall be liable to be ignored and rejected.
- 11.4 Tender sent by fax/telex/cable/electronically shall be ignored.

12. Tender currencies

- 12.1 The tenderer supplying indigenous goods or already imported goods shall quote only in Indian Rupees.
- 12.2 For imported goods if supplied directly from abroad, prices shall be quoted in any freely convertible currencies say USD, Euro, GBP or Yen. As regards price(s) for allied services, if any required with the goods, the same shall be quoted in Indian Rupees only if such services are to be performed /undertaken in India. Commission for Indian Agent, if any and if payable shall be indicated in the space provided for in the price schedule and will be payable in Indian Rupees only.
- 12.3 Tenders, where prices are quoted in any other way shall be treated as non-responsive and rejected.

13 Tender Prices

- 13.1 The Tenderer shall indicate on the Price Schedule provided under Section XI all the specified components of prices shown therein including the unit prices and total tender prices of the goods and services it proposes to supply against the requirement. All the columns shown in the price schedule should be filled up as required. If any column does not apply to a tenderer, it should be clarified as "NA" by the tenderer.
- 13.2 The tenderer has the option to submit its quotation for any one or more item (s) in the List of Requirements. However, separate sealed cover to be used for each item for price bid.
- 13.3 The quoted prices for goods offered from within India and that for goods offered from abroad are to be indicated separately in the applicable Price Schedules attached under Section XI.

The price quoted by the tenderer for indigenous goods shall not be higher than the lowest price charged for the goods of the same nature, class or description to an individual/ firm/ organisation or department of Govt. of India.

For imported goods, the price quoted shall not be higher than the lowest price charged by the tenderer for the goods of the same nature, class or description to a purchaser, domestic or foreign or to any organisation or department of Govt. of India.

If it is found at any stage that the goods as stated have been supplied at a lower price, then that price, with due allowance for elapsed time will be applicable to the present case and the difference in cost would be refunded by the supplier to the purchaser, if the contract has already been concluded.

- 13.4 While filling up the columns of the Price Schedule, the following aspects should be noted for compliance:

13.4.1 For domestic goods or goods of foreign origin located within India, the prices in the corresponding price schedule shall be entered separately in the following manner:

- a) the price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like sales tax, CST VAT, CENVAT, Custom Duty, Excise Duty etc. already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc;
- b) any sales or other taxes and any duties including excise duty, which will be payable on the finished goods in India if the contract is awarded;
- c) charges towards Packing & Forwarding, Inland Transportation, Insurance, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Price Schedule;
- d) the price of Incidental Services, as mentioned in List of Requirements and Price Schedule;
- e) the prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Price Schedule; and
- f) the price of CMC, as mentioned in List of Requirements, Technical Specification and Price Schedule.

13.4.2 For goods offered from abroad, the prices in the corresponding price schedule shall be entered separately in the following manner:

- a) the price of goods quoted FOB port of shipment, as indicated in the List of Requirements and Price Schedule;
- b) the amount of freight and insurance and price of goods quoted CIP port of entry in India as indicated in the List of Requirements and Price Schedule;
- c) the **price of goods quoted should be on DDP basis at consignee site in India** as indicated in the List of Requirements, Price Schedule and Consignee List;
- d) wherever applicable, the amount of custom duty with CDEC applicable on CIP value on the goods to be imported; ESIC will issue Customs Duty Exemption Certificate (CDEC) where applicable.
- e) the charges for Loading/Unloading, Inland transportation, Insurance and other local costs, Incidental cost to delivery of the goods from the port of entry in India to Consignee Site, as specified in the List of Requirements and Price Schedule;
- f) the charges for Incidental Services, as in the List of Requirements and Price Schedule;
- g) the prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Price Schedule; and
- h) the price of CMC, as mentioned in List of Requirements, Technical Specification and Price Schedule.

13.5 Additional information and instruction on Duties and Taxes:

13.5.1 If the Tenderer desires to ask for excise duty, sales tax/ VAT, Service Tax, Works Contract Tax etc. to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

13.5.2 Excise Duty:

- a) If reimbursement of excise duty is intended as extra over the quoted prices, the supplier must specifically say so also indicating the rate, quantum and nature of the duty applicable. In the absence of any such stipulation it will be presumed that the prices quoted are firm and final and no claim on account of excise duty will be entertained after the opening of tenders.

- b) If a Tenderer chooses to quote a price inclusive of excise duty and also desires to be reimbursed for variation, if any, in the excise duty during the time of supply, the tenderer must clearly mention the same and also indicate the rate and quantum of excise duty included in its price. Failure to indicate all such details in clear terms may result in rejection of that tender.
- c) Subject to sub clauses 13.5.2 (a) & (b) above, any change in excise duty upward/downward as a result of any statutory variation in excise duty taking place within contract terms shall be allowed to the extent of actual quantum of excise duty paid by the supplier. In case of downward revision in excise duty, the actual quantum of reduction of excise duty shall be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.

13.5.3 Sales Tax:

If a tenderer asks for sales tax/ VAT, Service Tax and Works Contract Tax to be paid extra, the rate and nature of sales tax applicable should be shown separately. The sales tax / VAT, Service Tax and Works Contract Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction of sale is legally liable to sales tax / VAT, Service Tax and Works Contract Tax and is payable as per the terms of the contract. If any refund of Tax is received at a later date, the Supplier must return the amount forth-with to the purchaser.

13.5.4 Octroi Duty and Local Duties & Taxes:

Octroi duty, terminal tax and other levies of local bodies (like town body, municipal body etc.) if not exempted shall be reimbursed to the supplier on production of such proof of payment along with the final bill.

13.5.5 Customs Duty:

The tenderer shall quote the Customs duty wherever applicable, considering availability of CDEC. The applicable rates and amount of the Custom Duty and the corresponding Indian Customs Tariff number should be shown separately in the price schedule. In case of non-availability of CDEC, the supplier has to custom cleared the goods paying the applicable Custom Duty and any difference of duty from the applicable CDEC rate shall be reimbursed separately on production of such proof of payment alongwith the final bill.

- 13.6 For transportation of imported goods offered from abroad, relevant instructions as incorporated under GCC Clause 10 shall be followed.
- 13.7 For insurance of goods to be supplied, relevant instructions as provided under GCC Clause 11 shall be followed.
- 13.8 Unless otherwise specifically indicated in this TE document, the terms FCA, FOB, FAS, CIF, CIP, DDP etc. for imported goods offered from abroad, shall be governed by the rules & regulations prescribed in the current edition of INCOTERMS, published by the International Chamber of Commerce, Paris
- 13.9 The need for indication of all such price components by the tenderers, as required in this clause (viz., GIT clause 13) is for the purpose of comparison of the tenders by the purchaser and will no way restrict the purchaser's right to award the contract on the selected tenderer on any of the terms offered.

14. Indian Agent

- 14.1 If a foreign tenderer has engaged an agent in India in connection with its tender, the foreign tenderer, in addition to indicating Indian agent's commission, if any, in a manner described under GIT sub clause 12.2 above, shall also furnish the following information:

- a) The complete name and address of the Indian Agent and its permanent income tax account number as allotted by the Indian Income Tax authority.
- b) The details of the services to be rendered by the agent for the subject requirement.
- c) Details of Service outlets in India, nearest to the consignee(s), to render services during Warranty and CMC period.

15. Firm Price

- 15.1 Unless otherwise specified in the SIT, prices quoted by the tenderer shall remain firm and fixed during the currency of the contract and not subject to variation on any account.
- 15.2 However, as regards taxes and duties, if any, chargeable on the goods and payable, the conditions stipulated in GIT clause 13 will apply.

16. Deleted

17 Documents Establishing Tenderer's Eligibility and Qualifications

- 17.1 Pursuant to GIT clause 11, the tenderer shall furnish, as part of its tender, relevant details and documents establishing its eligibility to quote and its qualifications to perform the contract if its tender is accepted.
- 17.2 The documentary evidence needed to establish the tenderer's qualifications shall fulfil the following requirements:
 - a) in case the tenderer offers to supply goods, which are manufactured by some other firm, the tenderer has been duly authorised by the goods manufacturer to quote for and supply the goods to the purchaser. The tenderer shall submit the manufacturer's authorization letter to this effect as per the standard form provided under **Section XIV** in this document.
 - b) the tenderer has the required financial, technical and production capability necessary to perform the contract and, further, it meets the qualification criteria incorporated in the Section IX in these documents.
 - c) in case the tenderer is not doing business in India, it is duly represented by an agent stationed in India fully equipped and able to carry out the required contractual functions and duties of the supplier including after sale service, maintenance & repair etc. of the goods in question, stocking of spare parts and fast moving components and other obligations, if any, specified in the conditions of contract and/or technical specifications.
 - d) in case the tenderer is an Indian agent/authorized representative quoting on behalf of a foreign manufacturer for the **restricted item**, the Indian agent/authorized representative is already enlisted under the Compulsory Enlistment Scheme of Ministry of Finance, Govt. of India, operated through Directorate General of Supplies & Disposals (DGS&D), New Delhi.

18. Documents establishing Good's Conformity to TE document.

- 18.1 The tenderer shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully conform to the goods and services specified by the purchaser in the TE documents. For this purpose the tenderer shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the TE documents to establish technical responsiveness of the goods and services offered in its tender.
- 18.2 In case there is any variation and/or deviation between the goods & services prescribed by the purchaser and that offered by the tenderer, the tenderer shall list out the same in a chart form without ambiguity and provide the same along with its tender.

18.3 If a tenderer furnishes wrong and/or misleading data, statement(s) etc. about technical acceptability of the goods and services offered by it, its tender will be liable to be ignored and rejected in addition to other remedies available to the purchaser in this regard.

19. Earnest Money Deposit (EMD)

19.1 Pursuant to GIT clauses 8.1 and 11.1(A) the tenderer shall furnish along with its tender, earnest money for amount as shown in the List of Requirements. The earnest money is required to protect the purchaser against the risk of the tenderer's unwarranted conduct as amplified under sub-clause 19.7 below.

19.2 Deleted

19.3 The earnest money shall be denominated in Indian Rupees or equivalent currencies as per GIT clause 12.2. **The earnest money shall not be accepted in any other form except the following:**

- i. Account Payee Demand Draft or
- ii. Banker's cheque

19.4 The demand draft or banker's cheque shall be drawn on any commercial bank in India or country of the tenderer, in favour of the "HLL Lifecare Limited" payable at New Delhi.

19.5 Deleted.

19.6 Unsuccessful tenderers' earnest money will be returned to them without any interest, after expiry of the tender validity period, but not later than thirty days after conclusion of the resultant contract. Successful tenderer's earnest money will be returned without any interest, after receipt of performance security from that tenderer.

19.7 Earnest Money is required to protect the purchaser against the risk of the Tenderer's conduct, which would warrant the forfeiture of the EMD. Earnest money of a tenderer will be forfeited, if the tenderer withdraws or amends its tender or impairs or derogates from the tender in any respect within the period of validity of its tender or if it comes to notice that the information/documents furnished in its tender is incorrect, false, misleading or forged without prejudice to other rights of the purchaser. The successful tenderer's earnest money will be forfeited without prejudice to other rights of Purchaser if it fails to furnish the required performance security within the specified period.

20. Tender Validity

20.1 If not mentioned otherwise in the SIT, the tenders shall remain valid for acceptance for a period of **120 days (One hundred and twenty days)** after the date of opening of techno-commercial tenders prescribed in the TE document. Any tender valid for a shorter period shall be treated as unresponsive and rejected.

20.2 In exceptional cases, the tenderers may be requested by the purchaser to extend the validity of their tenders up to a specified period. Such request(s) and responses thereto shall be conveyed by surface mail or by fax/ email followed by surface mail. The tenderers, who agree to extend the tender validity, are to extend the same without any change or modification of their original tender and they are also to extend the validity period of the EMD accordingly. A tenderer, however, may not agree to extend its tender validity without forfeiting its EMD.

20.3 In case the day up to which the tenders are to remain valid falls on/ subsequently declared a holiday or closed day for the purchaser, the tender validity shall automatically be extended up to the next working day.

21. Signing and Sealing of Tender

21.1 The tenderers shall submit their tenders as per the instructions contained in GIT Clause 11.

- 21.2 Unless otherwise mentioned in the SIT, a tenderer shall submit two copies of its tender marking them as “Original” and “Duplicate”. Duplicate tenders may contain all pages including Technical Literature/Catalogues as in Original tenders.
- 21.3 The original and duplicate copies of the tender shall either be typed or written in indelible ink and the same shall be signed by the tenderer or by a person(s) who has been duly authorized to bind the tenderer to the contract. The letter of authorization shall be by a written power of attorney, which shall also be furnished along with the tender.
- 21.4 All the copies of the tender shall be duly signed at the appropriate places as indicated in the TE documents and all other pages of the tender including printed literature, if any shall be initialled by the same person(s) signing the tender. The tender shall not contain any erasure or overwriting, except as necessary to correct any error made by the tenderer and, if there is any such correction; the same shall be initialled by the person(s) signing the tender.
- 21.5 The tenderer is to seal the original and copy of the tender in separate envelopes, duly marking the same as “Original”, “Duplicate” and so on and writing the address of the purchaser and the tender reference number on the envelopes. The sentence “NOT TO BE OPENED” before _____ (The tenderer is to put the date & time of tender opening) are to be written on these envelopes. The inner envelopes are then to be put in a bigger outer envelope along with envelope containing EMD, which will also be duly sealed, marked etc. as above. If the outer envelope is not sealed and marked properly as above, the purchaser will not assume any responsibility for its misplacement, premature opening, late opening etc.
- 21.6 TE document seeks quotation following two Tender System, in two parts. First part will be known as ‘Techno - Commercial Tender’, and the second part ‘Price Tender’ as specified in clause 11 of GIT. Tenderer shall seal ‘Techno - Commercial Tender (along with envelope containing EMD)’ and ‘Price Tender’ separately and covers will be suitably super scribed. Both these sealed covers shall be put in a bigger cover and sealed and procedure prescribed in Paras 21.1 to 21.5 followed.

D. SUBMISSION OF TENDERS

22. Submission of Tenders

- 22.1 Unless otherwise specified, the tenderers are to deposit the tenders in the tender box kept for this purpose at **HLL Lifecare Limited, Procurement and Consultancy Division, B-14A, Sector - 62, Noida -201307, Uttar Pradesh**. In case of bulky tender, which can not be put into tender box, the same shall be submitted by the tenderer by hand to **Head (P&CD)** or his nominee, **HLL Lifecare Limited, Procurement and Consultancy Division, B-14A, Sector -62, Noida - 201307, Uttar Pradesh**. The officer receiving the tender will give the tenderer an official receipt duly signed with date and time.
- 22.2 The tenderers must ensure that they deposit their tenders not later than the closing time and date specified for submission of tenders. It is the responsibility of the tenderer to ensure that their Tenders whether sent by post or by courier or by person, are dropped in the Tender Box by the specified clearing date and time. In the event of the specified date for submission of tender falls on / is subsequently declared a holiday or closed day for the purchaser, the tenders will be received up to the appointed time on the next working day.

23. Late Tender

- 23.1 A tender, which is received after the specified date and time for receipt of tenders will be treated as “late” tender and will be ignored.

24. Alteration and Withdrawal of Tender

- 24.1 The tenderer, after submitting its tender, is permitted to alter / modify its tender so long as such alterations / modifications are received duly signed, sealed and marked like the original tender,

within the deadline for submission of tenders. Alterations / modifications to tenders received after the prescribed deadline will not be considered.

- 24.2 No tender should be withdrawn after the deadline for submission of tender and before expiry of the tender validity period. If a tenderer withdraws the tender during this period, it will result in forfeiture of the earnest money furnished by the tenderer in its tender.

E. TENDER OPENING

25. Opening of Tenders

- 25.1 The purchaser will open the tenders at the specified date and time and at the specified place as indicated in the NIT.

In case the specified date of tender opening falls on / is subsequently declared a holiday or closed day for the purchaser, the tenders will be opened at the appointed time and place on the next working day.

- 25.2 **Authorized representatives of the tenderers**, who have submitted tenders on time **may attend the tender opening** provided they bring with them **letters of authority from the corresponding tenderers**.

The tender opening official(s) will prepare a list of the representatives attending the tender opening. The list will contain the representatives' names & signatures and corresponding tenderers' names and addresses.

- 25.3 **Two-Tender system as mentioned in para 21.6 above will be as follows**. The **Techno-Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the TE document. During the Techno - Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, Earnest Money Deposit and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price Tenders of only the Techno-Commercially acceptable offers shall be opened at a latter date which will be notified to such tenderers. The prices, special discount if any of the goods offered etc., as deemed fit by tender opening official(s) will be read out.

F. SCRUTINY AND EVALUATION OF TENDERS

26. Basic Principle

- 26.1 Tenders will be evaluated on the basis of the terms & conditions already incorporated in the TE document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

27. Preliminary Scrutiny of Tenders

- 27.1 The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished, whether the documents have been properly signed stamped and whether the Tenders are generally in order.
- 27.2 Deleted.
- 27.3 Deleted

- 27.4 The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. as prescribed in the TE document. The tenders, which do not meet the basic requirements, are liable to be treated as non – responsive and will be summarily ignored.
- 27.5 The following are some of the important aspects, for which a tender shall be declared non-responsive and will be summarily ignored;
- (i) Tender form as per Section X (signed and stamped) not enclosed
 - (ii) Tender is unsigned.
 - (iii) Tender validity is shorter than the required period.
 - (iv) Required EMD have not been provided.
 - (v) Tenderer has quoted for goods manufactured by other manufacturer(s) without the required Manufacturer's Authorisation as per Format given in Section XIV.
 - (vi) Tenderer has not agreed to give the required performance security.
 - (vii) Goods offered are not meeting the tender enquiry specification.
 - (viii) Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry like terms of payment, liquidated damages clause, warranty clause, DDP clause, Delivery period clause, dispute resolution mechanism applicable law.
 - (ix) Poor/ unsatisfactory past performance.
 - (x) Tenderers who stand deregistered/banned/blacklisted by any Govt. Authorities.
 - (xi) Tenderer is not eligible as per GIT Clauses 5.1 & 17.1.
 - (xii) Tenderer has not quoted for the entire quantity as specified in the List of Requirements for the quoted item (s).

28. Deleted

29 Discrepancies in Prices

- 29.1 If, in the price structure quoted by a tenderer, there is discrepancy between the unit price and the total price (which is obtained by multiplying the unit price by the quantity), the unit price shall prevail and the total price corrected accordingly, unless the purchaser feels that the tenderer has made a mistake in placing the decimal point in the unit price, in which case the total price as quoted shall prevail over the unit price and the unit price corrected accordingly.
- 29.2 If there is an error in a total price, which has been worked out through addition and/or subtraction of subtotals, the subtotals shall prevail and the total corrected; and
- 29.3 If there is a discrepancy between the amount expressed in words and figures, the amount in words shall prevail, subject to sub clause 29.1 and 29.2 above.

30. Discrepancy between original and copies of Tender

In case any discrepancy is observed between the text etc. of the original copy and that in the other copies of the same tender set, the text etc. of the original copy shall prevail.

31. Qualification Criteria

Tenders of the tenderers, who do not meet the required Qualification Criteria prescribed in Section IX, will be treated as non-responsive and will not be considered further.

32. Conversion of tender currencies to Indian Rupees

In case the TE document permits the tenderers to quote their prices in different currencies, all such quoted prices of the responsive tenderers will be converted to a single currency viz., Indian Rupees for the purpose of equitable comparison and evaluation, as per the exchange rates established by the Reserve Bank of India for similar transactions, **as on the date of 'Price Tender' opening.**

33. Deleted

34. Comparison of Tenders

34.1 Unless mentioned otherwise in Section – III – Special Instructions to Tenderers and Section – VI – List of Requirements, the comparison of the responsive tenders shall be carried out on **Delivery Duty Paid (DDP)** consignee site basis. The quoted turnkey prices and CMC prices will also be added for comparison/ranking purpose for evaluation.

35. Additional Factors and Parameters for Evaluation and Ranking of Responsive Tenders

35.1 Further to GIT Clause 34 above, the purchaser's evaluation of a tender will include and take into account the following:

- i) In the case of goods manufactured in India or goods of foreign origin already located in India, sales tax & other similar taxes and excise duty & other similar duties, Customs Duties, Service Tax, Works Contract Tax etc which will be contractually payable (to the tenderer), on the goods if a contract is awarded on the tenderer; and
- ii) in the case of goods of foreign origin offered from abroad, customs duty and other similar import duties/taxes, which will be contractually payable (to the tenderer) on the goods if the contract is awarded on the tenderer.

35.2 The purchaser's evaluation of tender will also take into account the additional factors, if any, incorporated in SIT in the manner and to the extent indicated therein.

35.3 Deleted

36. Tenderer's capability to perform the contract

36.1 The purchaser, through the above process of tender scrutiny and tender evaluation will determine to its satisfaction whether the tenderer, whose tender has been determined as the lowest evaluated responsive tender is eligible, qualified and capable in all respects to perform the contract satisfactorily. If, there is more than one item in the List of Requirements, then, such determination will be made separately for each item.

36.2 The above-mentioned determination will, interalia, take into account the tenderer's financial, technical and production capabilities for satisfying all the requirements of the purchaser as incorporated in the TE document. Such determination will be based upon scrutiny and examination of all relevant data and details submitted by the tenderer in its tender as well as such other allied information as deemed appropriate by the purchaser.

37. Contacting the Purchaser

37.1 From the time of submission of tender to the time of awarding the contract, if a tenderer needs to contact the purchaser for any reason relating to this tender enquiry and / or its tender, it should do so only in writing.

37.2 In case a tenderer attempts to influence the purchaser in the purchaser's decision on scrutiny, comparison & evaluation of tenders and awarding the contract, the tender of the tenderer shall be liable for rejection in addition to appropriate administrative actions being taken against that tenderer, as deemed fit by the purchaser.

G. AWARD OF CONTRACT**38. Purchaser's Right to accept any tender and to reject any or all tenders**

The purchaser reserves the right to accept in part or in full any tender or reject any or more tender(s) without assigning any reason or to cancel the tendering process and reject all tenders at

any time prior to award of contract, without incurring any liability, whatsoever to the affected tenderer or tenderers.

39. Award Criteria

Subject to GIT clause 38 above, the contract will be awarded to the lowest evaluated responsive tenderer decided by the purchaser in terms of GIT Clause 36.

40. Variation of Quantities at the Time of Award/ Currency of Contract

40.1 At the time of awarding the contract, the purchaser reserves the right to increase or decrease by up to twenty five (25) per cent, the quantity of goods and services mentioned in the schedule (s) in the "List of Requirements" without any change in the unit price and other terms & conditions quoted by the tenderer.

40.2 If the quantity has not been increased at the time of the awarding the contract, the purchaser reserves the right to increase by up to twenty five (25) per cent, the quantity of goods and services mentioned in the contract without any change in the unit price and other terms & conditions mentioned in the contract, during the currency of the contract or within one-year from the date of Notification of Award.

41. Notification of Award

41.1 Before expiry of the tender validity period, the purchaser will notify the successful tenderer(s) in writing, by registered / speed post or by fax/ telex/cable (to be confirmed by registered / speed post) that its tender for goods & services, which have been selected by the purchaser, has been accepted, also briefly indicating therein the essential details like description, specification and quantity of the goods & services and corresponding prices accepted. The successful tenderer must furnish to the purchaser the required performance security within twenty-one (21) days from the date of dispatch of this notification, failing which the EMD will be forfeited and the award will be cancelled. Relevant details about the performance security have been provided under GCC Clause 5 under Section IV.

41.2 The Notification of Award shall constitute the conclusion of the Contract.

42. Issue of Contract

42.1 Promptly after notification of award, the Purchaser/Consignee will mail the contract form (as per Section XVI) duly completed and signed, in duplicate, to the successful tenderer by registered / speed post.

42.2 Within twenty one days from the date of the contract, the successful tenderer shall return the original copy of the contract, duly signed and dated, to the Purchaser by registered / speed post.

42.3 The Purchaser reserves the right to issue the Notification of Award consignee wise.

43. Non-receipt of Performance Security and Contract by the Purchaser/Consignee

Failure of the successful tenderer in providing performance security and / or returning contract copy duly signed in terms of GIT clauses 41 and 42 above shall make the tenderer liable for forfeiture of its EMD and, also, for further actions by the Purchaser/Consignee against it as per the clause 24 of GCC – Termination of default.

44. Return of E M D

The earnest money of the successful tenderer and the unsuccessful tenderers will be returned to them without any interest, whatsoever, in terms of GIT Clause 19.6.

45. Publication of Tender Result

The name and address of the successful tenderer(s) receiving the contract(s) will be mentioned in the notice board/bulletin/web site of the purchaser.

46. Corrupt or Fraudulent Practices

It is required by all concerned to observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy, the Purchaser: -

- (a) defines, for the purposes of this provision, the terms set forth below as follows:
 - (i) “corrupt practice” means the offering, giving, receiving or soliciting of any thing of value to influence the action of a public official in the procurement process or in contract execution; and
 - (ii) “fraudulent practice” means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Purchaser, and includes collusive practice among Tenderers (prior to or after Tender submission) designed to establish Tender prices at artificial non-competitive levels and to deprive the Purchaser of the benefits of free and open competition;
- (b) will reject a proposal for award if it determines that the Tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
- (c) will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a contract by the purchaser if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing the contract.

SECTION - III
SPECIAL INSTRUCTIONS TO TENDERERS
(SIT)

The following Special Instructions to Tenderers will apply for this purchase. These special instructions will modify/substitute/supplement the corresponding General Instructions to Tenderers (GIT) incorporated in Section II. The corresponding GIT clause numbers have also been indicated in the text below:

In case of any conflict between the provision in the GIT and that in the SIT, the provision contained in the SIT shall prevail.

SECTION - IV**GENERAL CONDITIONS OF CONTRACT (GCC)****TABLE OF CLAUSES**

Sl No.	Topic	Page
1	Application	26
2	Use of contract documents and information	26
3	Patent Rights	26
4	Country of Origin	26
5	Performance Security	26
6	Technical Specifications and Standards	27
7	Packing and Marking	27
8	Inspection, Testing and Quality Control	28
9	Terms of Delivery	28
10	Transportation of Goods	28
11	Insurance	29
12	Spare parts	29
13	Incidental services	30
14	Distribution of Dispatch Documents for Clearance/Receipt of Goods	30
15	Warranty	31
16	Assignment	32
17	Sub Contracts	32
18	Modification of contract	32
19	Prices	33
20	Taxes and Duties	33
21	Terms and mode of Payment	33
22	Delay in the supplier's performance	35
23	Liquidated Damages	36
24	Termination for default	36
25	Termination for insolvency	37
26	Force Majeure	37
27	Termination for convenience	37
28	Governing language	38
29	Notices	38
30	Resolution of disputes	38
31	Applicable Law	38
32	General/Miscellaneous Clauses	38

GENERAL CONDITIONS OF CONTRACT (GCC)

1. Application

The General Conditions of Contract incorporated in this section shall be applicable for this purchase to the extent the same are not superseded by the Special Conditions of Contract prescribed under Section V, List of requirements under Section VI and Technical Specification under Section VII of this document.

2. Use of contract documents and information

2.1 The supplier shall not, without the purchaser's prior written consent, disclose the contract or any provision thereof including any specification, drawing, sample or any information furnished by or on behalf of the purchaser in connection therewith, to any person other than the person(s) employed by the supplier in the performance of the contract emanating from this TE document. Further, any such disclosure to any such employed person shall be made in confidence and only so far as necessary for the purposes of such performance for this contract.

2.2 Further, the supplier shall not, without the purchaser's prior written consent, make use of any document or information mentioned in GCC sub-clause 2.1 above except for the sole purpose of performing this contract.

2.3 Except the contract issued to the supplier, each and every other document mentioned in GCC sub-clause 2.1 above shall remain the property of the purchaser and, if advised by the purchaser, all copies of all such documents shall be returned to the purchaser on completion of the supplier's performance and obligations under this contract.

3. Patent Rights

The supplier shall, at all times, indemnify and keep indemnified the purchaser, free of cost, against all claims which may arise in respect of goods & services to be provided by the supplier under the contract for infringement of any intellectual property rights or any other right protected by patent, registration of designs or trademarks. In the event of any such claim in respect of alleged breach of patent, registered designs, trade marks etc. being made against the purchaser, the purchaser shall notify the supplier of the same and the supplier shall, at his own expenses take care of the same for settlement without any liability to the purchaser.

4. Country of Origin

4.1 All goods and services to be supplied and provided for the contract shall have the origin in India or in the countries with which the Government of India has trade relations.

4.2 The word "origin" incorporated in this clause means the place from where the goods are mined, cultivated, grown, manufactured, produced or processed or from where the services are arranged.

4.3 The country of origin may be specified in the Price Schedule

5. Performance Security

5.1 Within twenty-one (21) days from date of the issue of notification of award by the Purchaser/Consignee, the supplier, shall furnish performance security to the Purchaser/Consignee for an amount equal to ten percent (10%) of the total value of the contract, **valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations.**

5.2 The Performance security shall be denominated in Indian Rupees or in the currency of the contract as detailed below:

- a) It shall be in any one of the forms namely Account Payee Demand Draft drawn from any Scheduled bank in India or Bank Guarantee issued by a Scheduled bank in India, in the

prescribed form as provided in section XV of this document in favour of the Purchaser/Consignee.

- b) In the case of Bank Guarantee furnished from banks outside India (i.e. foreign Banks), it should be authenticated and countersigned by any nationalised bank in India by way of back-to-back counter guarantee.

- 5.3 In the event of any failure /default of the supplier with or without any quantifiable loss to the government including furnishing of consignee wise Bank Guarantee for CMC security as per Proforma in Section XV, the amount of the performance security is liable to be forfeited. The Administration Department may do the needful to cover any failure/default of the supplier with or without any quantifiable loss to the Government.
- 5.4 In the event of any amendment issued to the contract, the supplier shall, within twenty-one (21) days of issue of the amendment, furnish the corresponding amendment to the Performance Security (as necessary), rendering the same valid in all respects in terms of the contract, as amended.
- 5.5 The supplier shall enter into Comprehensive Maintenance Contract as per the 'Contract Form - B' in Section XVI with respective consignees, 3 (three) months prior to the completion of Warranty Period. The CMC will commence from the date of expiry of the Warranty Period.
- 5.6 Subject to GCC sub – clause 5.3 above, the Purchaser/Consignee will release the Performance Security without any interest to the supplier on completion of the supplier's all contractual obligations including the warranty obligations & after receipt of Consignee wise bank guarantee for CMC security in favour of Head of the Hospital/ Institute/ Medical College of the consignee as per the format in Section XV.

6. Technical Specifications and Standards

- 6.1 The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications and quality control parameters mentioned in 'Technical Specification' and 'Quality Control Requirements' under Sections VII and VIII of this document.

7. Packing and Marking

- 7.1 The packing for the goods to be provided by the supplier should be strong and durable enough to withstand, without limitation, the entire journey during transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. As and if necessary, the size, weights and volumes of the packing cases shall also take into consideration, the remoteness of the final destination of the goods and availability or otherwise of transport and handling facilities at all points during transit up to final destination as per the contract.
- 7.2 The quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall strictly comply with the requirements as provided in Technical Specifications and Quality Control Requirements under Sections VII and VIII and in SCC under Section V. In case the packing requirements are amended due to issue of any amendment to the contract, the same shall also be taken care of by the supplier accordingly.
- 7.3 Packing instructions:

Unless otherwise mentioned in the Technical Specification and Quality Control Requirements under Sections VII and VIII and in SCC under Section V, the supplier shall make separate packages for each consignee (in case there is more than one consignee mentioned in the contract) and mark each package on three sides with the following with indelible paint of proper quality:

- a. contract number and date
- b. brief description of goods including quantity
- c. packing list reference number
- d. country of origin of goods

- e. consignee's name and full address and
- f. supplier's name and address

8. Inspection, Testing and Quality Control

- 8.1 The purchaser and/or its nominated representative(s) will, without any extra cost to the purchaser, inspect and/or test the ordered goods and the related services to confirm their conformity to the contract specifications and other quality control details incorporated in the contract. The purchaser shall inform the supplier in advance, in writing, the purchaser's programme for such inspection and, also the identity of the officials to be deputed for this purpose. The cost towards the transportation, boarding & lodging will be borne by the purchaser and/or its nominated representative(s).
- 8.2 The Technical Specification and Quality Control Requirements incorporated in the contract shall specify what inspections and tests are to be carried out and, also, where and how they are to be conducted. If such inspections and tests are conducted in the premises of the supplier or its subcontractor(s), all reasonable facilities and assistance, including access to relevant drawings, design details and production data, shall be furnished by the supplier to the purchaser's inspector at no charge to the purchaser.
- 8.3 If during such inspections and tests the contracted goods fail to conform to the required specifications and standards, the purchaser's inspector may reject them and the supplier shall either replace the rejected goods or make all alterations necessary to meet the specifications and standards, as required, free of cost to the purchaser and resubmit the same to the purchaser's inspector for conducting the inspections and tests again.
- 8.4 In case the contract stipulates pre-despatch inspection of the ordered goods at supplier's premises, the supplier shall put up the goods for such inspection to the purchaser's inspector well ahead of the contractual delivery period, so that the purchaser's inspector is able to complete the inspection within the contractual delivery period.
- 8.5 If the supplier tenders the goods to the purchaser's inspector for inspection at the last moment without providing reasonable time to the inspector for completing the inspection within the contractual delivery period, the inspector may carry out the inspection and complete the formality beyond the contractual delivery period at the risk and expense of the supplier. The fact that the goods have been inspected after the contractual delivery period will not have the effect of keeping the contract alive and this will be without any prejudice to the legal rights and remedies available to the purchaser under the terms & conditions of the contract.
- 8.6 The purchaser's/consignee's contractual right to inspect, test and, if necessary, reject the goods after the goods' arrival at the final destination shall have no bearing of the fact that the goods have previously been inspected and cleared by purchaser's inspector during pre-despatch inspection mentioned above.
- 8.7 Goods accepted by the purchaser/consignee and/or its inspector at initial inspection and in final inspection in terms of the contract shall in no way dilute purchaser's/consignee's right to reject the same later, if found deficient in terms of the warranty clause of the contract, as incorporated under GCC Clause 15.
- 8.8 Principal/ Foreign supplier shall also have the equipment inspected by recognised/ reputed agency like SGS, Lloyd or equivalent (acceptable to the purchaser) prior to despatch at the supplier's cost and furnish necessary certificate from the said agency in support of their claim.

9. Terms of Delivery

Goods shall be delivered by the supplier in accordance with the terms of delivery specified in the contract.

10. Transportation of Goods

10.1 Instructions for transportation of imported goods offered from abroad:

The supplier shall not arrange part-shipments and/or transshipment without the express/prior written consent of the purchaser. The supplier is required under the contract to deliver the goods under DDP at consignee site; the shipment shall be made by Indian flag vessel or by vessels belonging to the conference lines in which India is a member country through India's forwarding agents/coordinators. In case the forwarding agent/coordinators are unable to provide timely adequate space in Indian flag vessel or by vessels belonging to the conference lines, the supplier shall arrange shipment through any available vessel to adhere to the delivery schedule given in the contract.

In case of airlifting of imported goods offered from abroad, the same will be done only through the National Carrier i.e. Air India wherever applicable. In case the National Carrier is not available, any other airlines available for early delivery may be arranged.

Goods will be custom cleared by the supplier/ Indian agent and transported to the consignee's site as per the contract terms.

Instructions for transportation of domestic goods including goods already imported by the supplier under its own arrangement:

In case no instruction is provided in this regard in the SCC, the supplier will arrange transportation of the ordered goods as per its own procedure.

The goods shall be custom cleared by the Supplier/ Indian Agent and transported to the consignee's site as per contract terms. The supplier shall be responsible for safe and timely delivery of ordered goods under his own arrangement.

11. Insurance:

Unless otherwise instructed in the SCC, the supplier shall make arrangements for insuring the goods against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the following manner:

- i) in case of supply of domestic goods on Consignee site basis, the supplier shall be responsible till the entire stores contracted for arrival in good condition at destination. The transit risk in this respect shall be covered by the Supplier by getting the stores duly insured. The insurance cover shall be obtained by the Supplier and should be valid till 3 months after the receipt of goods by the Consignee.
- ii) in case of supply of the imported goods on DDP Basis, the supplier shall arrange and pay for marine/ air insurance making the consignee as beneficiary. The additional extended Insurance (local transportation and storage) would also be borne by the Supplier from the port of entry to the consignee site for a period including 3 months beyond date of delivery.

12. Spare parts

12.1 If specified in the List of Requirements and in the resultant contract, the supplier shall supply/provide any or all of the following materials, information etc. pertaining to spare parts manufactured and/or supplied by the supplier:

- a) The spare parts as selected by the Purchaser/Consignee to be purchased from the supplier, subject to the condition that such purchase of the spare parts shall not relieve the supplier of any contractual obligation including warranty obligations; and

- b) In case the production of the spare parts is discontinued:
- i) Sufficient advance notice to the Purchaser/Consignee before such discontinuation to provide adequate time to the purchaser to purchase the required spare parts etc., and
 - ii) Immediately following such discontinuation, providing the Purchaser/Consignee, free of cost, the designs, drawings, layouts and specifications of the spare parts, as and if requested by the Purchaser/Consignee.

12.2 Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spares for the goods so that the same are supplied to the Purchaser/Consignee promptly on receipt of order from the Purchaser/Consignee.

13. Incidental services

Subject to the stipulation, if any, in the SCC (Section – V), List of Requirements (Section – VI) and the Technical Specification (Section – VII), the supplier shall be required to perform the following services.

- i) Installation & commissioning, Supervision and Demonstration of the goods
- ii) Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
- iii) Training of Consignee's Doctors, Staff, operators etc. for operating and maintaining the goods
- iv) Supplying required number of operation & maintenance manual for the goods

14. Distribution of Dispatch Documents for Clearance/Receipt of Goods

The supplier shall send all the relevant despatch documents well in time to the Purchaser/Consignee

Unless otherwise specified in the SCC, the usual documents involved and the drill to be followed in general for this purpose are as follows.

A) For Domestic Goods, including goods already imported by the supplier under its own arrangement

Within 24 hours of despatch, the supplier shall notify the purchaser, consignee, and others concerned if mentioned in the contract, the complete details of despatch and also supply the following documents to them by registered post / speed post (or as instructed in the contract):

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Two copies of packing list identifying contents of each package;
- (iii) Inspection certificate issued by the nominated Inspection agency, if any.
- (iv) Certificate of origin;
- (v) Insurance Certificate as per GCC Clause 11.
- (vi) Manufacturers/Supplier's warranty certificate & In-house inspection certificate.

B) For goods imported from abroad

Within 24 hours of despatch, the supplier shall notify the purchaser, consignee, and others concerned if mentioned in the contract, the complete details of despatch and also supply the

following documents to them by airmail/ registered post / speed post (or as instructed in the contract).

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Original and four copies of the negotiable clean, on-board Bill of Lading/Airway bill, marked freight pre paid and four copies of non-negotiable Bill of Lading/Airway bill;
- (iii) Four Copies of packing list identifying contents of each package;
- (iv) Insurance Certificate as per GCC Clause 11.
- (v) Manufacturer's/Supplier's warranty certificate;
- (vi) Inspection Certificate for the despatched equipments issued by recognized/ reputed agency like SGS, Lloyd or equivalent (acceptable to the purchaser) prior to despatch
- (vii) Manufacturer's own factory inspection report;
- (viii) Certificate of origin
- (ix) Port of Loading;
- (x) Port of Discharge and
- (xi) Expected date of arrival.

15. Warranty

- 15.1 The supplier warrants comprehensively that the goods supplied under the contract is new, unused and incorporate all recent improvements in design and materials unless prescribed otherwise by the purchaser in the contract. The supplier further warrants that the goods supplied under the contract shall have no defect arising from design, materials or workmanship or from any act or omission of the supplier that may develop under normal use of the supplied goods under the conditions prevailing in India.
- 15.2 This **warranty shall remain valid for 2(Two) years** in general, after the goods or any portion thereof as the case may be, have been delivered to the final destination and installed and commissioned at the final destination and accepted by the Purchaser/Consignee in terms of the contract, **unless specified otherwise in the SCC.**
- a. No conditional warranty like mishandling, manufacturing defects etc. will be acceptable.
 - b. Warranty as well as Comprehensive Maintenance contract will be inclusive of all accessories and Turnkey work and it will also cover the following items:-
 - i. X-ray and CT tubes and high-tension cables.
 - ii. Helium replacement
 - iii. Any kind of motor
 - iv. Plastic & Glass parts
 - v. All kinds of sensors including oxygen sensors
 - vi. All kinds of coils, probes and transducers including ECG cable, BP transducers, SpO2 Probes, Ultrasound and Color Doppler Transducers/probes, BP Cuffs, Defibrillator internal paddles, chart recorders, ventilator reusable patient circuits, servo humidifier with chamber, electrodes and probes for blood gas analyser, MRI coils.
 - vii. All kinds of flat panel sensors and cassettes for Digital Radiography & Computer Radiography systems and patients handling trolleys, etc.
 - viii. Printers and imagers including laser and thermal printers with all parts.
 - ix. UPS including the replacement of Batteries.
 - x. Air-conditioners
 - c. Replacement and repair will be under taken for the defective goods.
 - d. Proper marking has to be made for all spares for identification like printing of installation and repair dates.

- 15.3 In case of any claim arising out of this warranty, the Purchaser/Consignee shall promptly notify the same in writing to the supplier. The period of the warranty will be as per G.C.C clause number 15.2 above irrespective of any other period mentioned elsewhere in the bidding documents.
- 15.4 Upon receipt of such notice, the supplier shall, within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis respond to take action to repair or replace the defective goods or parts thereof, free of cost, at the ultimate destination. The supplier shall take over the replaced parts/goods after providing their replacements and no claim, whatsoever shall lie on the purchaser for such replaced parts/goods thereafter. The penalty clause for non-rectification will be applicable as per tender conditions
- 15.5 In the event of any rectification of a defect or replacement of any defective goods during the warranty period, the **warranty for the rectified/replaced goods shall be extended to a further period as mentioned under clause 15.2** from the date such rectified / replaced goods starts functioning to the satisfaction of the purchaser.
- 15.6 If the supplier, having been notified, fails to respond to take action to repair or replace the defect(s) within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis, the purchaser may proceed to take such remedial action(s) as deemed fit by the purchaser, at the risk and expense of the supplier and without prejudice to other contractual rights and remedies, which the purchaser may have against the supplier.
- 15.7 During Warranty period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the installation for preventive maintenance of the goods
- 15.8 The Purchaser/Consignee reserve the rights to enter into **Annual Comprehensive Maintenance Contract** between Consignee and the Supplier for the period as mentioned in General Points for Technical Specifications, **Section VII (para-4)**, after the completion of warranty period.
- 15.9 The supplier along with its Indian Agent and the CMC provider shall ensure continued supply of the spare parts for the machines and equipments supplied by them to the purchaser for **10 years** from the date of installation and handing over.
- 15.10 The Supplier along with its Indian Agent and the CMC Provider shall always accord most favoured client status to the Purchaser vis-à-vis its other Clients/Purchasers of its equipments/machines/goods etc. and shall always give the most competitive price for its machines/equipments supplied to the Purchaser/Consignee.

16. Assignment

- 16.1 The Supplier shall not assign, either in whole or in part, its contractual duties, responsibilities and obligations to perform the contract, except with the Purchaser's prior written permission.

17. Sub Contracts

- 17.1 The Supplier shall notify the Purchaser in writing of all sub contracts awarded under the contract if not already specified in its tender. Such notification, in its original tender or later, shall not relieve the Supplier from any of its liability or obligation under the terms and conditions of the contract.
- 17.2 Sub contract shall be only for bought out items and sub-assemblies.
- 17.3 Sub contracts shall also comply with the provisions of GCC Clause 4 ("Country of Origin").

18. Modification of contract

- 18.1 If necessary, the purchaser may, by a written order given to the supplier at any time during the currency of the contract, amend the contract by making alterations and modifications within the general scope of contract in any one or more of the following:
- a) Specifications, drawings, designs etc. where goods to be supplied under the contract are to be specially manufactured for the purchaser,
 - b) Mode of packing,
 - c) Incidental services to be provided by the supplier

- d) Mode of despatch,
- e) Place of delivery, and
- f) Any other area(s) of the contract, as felt necessary by the purchaser depending on the merits of the case.

18.2 In the event of any such modification/alteration causing increase or decrease in the cost of goods and services to be supplied and provided, or in the time required by the supplier to perform any obligation under the contract, an equitable adjustment shall be made in the contract price and/or contract delivery schedule, as the case may be, and the contract amended accordingly. If the supplier doesn't agree to the adjustment made by the Purchaser/Consignee, the supplier shall convey its views to the Purchaser/Consignee within twenty-one days from the date of the supplier's receipt of the Purchaser's/Consignee's amendment / modification of the contract.

19. Prices

19.1 Prices to be charged by the supplier for supply of goods and provision of services in terms of the contract shall not vary from the corresponding prices quoted by the supplier in its tender and incorporated in the contract except for any price adjustment authorised in the SCC.

20. Taxes and Duties

20.1 Supplier shall be entirely responsible for all taxes, duties, fees, levies etc. incurred until delivery of the contracted goods to the purchaser.

20.2 Further instruction, if any, shall be as provided in the SCC.

21. Terms and Mode of Payment

21.1 Payment Terms

Payment shall be made subject to recoveries, if any, by way of liquidated damages or any other charges as per terms & conditions of contract in the following manner.

A) Payment for Domestic Goods Or Foreign Origin Located Within India.

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) On delivery:

- (i) 90 % payment of the contract price shall be paid within 24 hours on receipt of goods in good condition and upon the submission of the following documents complete in all respects. Bills shall be returned un-paid in case of any discrepancy. Delay in payment on account of above shall rest with the supplier.
- (ii) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (iii) Consignee Receipt Certificate as per Section XVII in original issued by the authorized representative of the consignee;
- (iv) Two copies of packing list identifying contents of each package;
- (v) Inspection certificate issued by the nominated Inspection agency, if any;
- (vi) Insurance Certificate as per GCC Clause 11;
- (vii) Certificate of origin.

b) On Acceptance:

Balance 10 % payment would be made against 'Final Acceptance Certificate' as per Section XVIII of goods to be issued by the consignees subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise.

B) Payment for Imported Goods:

Payment for foreign currency portion shall be made in the currency as specified in the contract in the following manner:

a) On delivery:

Ninety (90) % of the net CIP price (CIP price less Indian Agency commission) of the goods shipped shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country and upon submission of documents specified hereunder:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Original and four copies of the negotiable clean, on-board Bill of Lading/ Airway bill , marked freight pre paid and four copies of non-negotiable Bill of Lading/Airway bill;
- (iii) Four Copies of packing list identifying contents of each package;
- (iv) Insurance Certificate as per GCC Clause 11 and documents also to be submitted for payment of LC confirming that dispatch documents has already been sent within 24 hours to all concerned as per the contract;
- (v) Manufacturer's/Supplier's warranty certificate;
- (vi) Manufacturer's own factory inspection report and
- (vii) Certificate of origin by the chamber of commerce of the concerned country;
- (viii) Inspection Certificate for the despatched equipments issued by recognized/ reputed agency like SGS, Lloyd or equivalent (acceptable to the purchaser) prior to despatch.
- (ix) Consignee Receipt Certificate as per Section XVII in original issued by the authorized representative of the consignee

b) On Acceptance:

Balance payment of 10 % of net CIP price of goods would be made against 'Final Acceptance Certificate' as per Section XVIII to be issued by the consignees through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the Foreign Principal in a bank in his country, subject to recoveries, if any.

- c) Payment of custom duty amount with Custom Duty Exemption Certificate (CDEC), if applicable, customs clearance and handling charges, loading/ unloading, inland transportation, incidental costs till consignee site & incidental services (including installation & commissioning, supervision, demonstration and training) will be paid in Indian Rupees to the Indian agent at actual not exceeding the quoted rates after 100 % payment to the foreign principal.**

d) Payment of Indian Agency Commission:

Indian Agency commission will be paid to the manufacturer's agent in the local currency for an amount in Indian rupees indicated in the relevant Price Schedule (as per prevailing rate of exchange ruling on the date of Contract) and shall not be subject to further escalation / exchange variation. Payment shall be made after 100% payment to the Foreign Principal.

C) Payment of Turnkey, if any:

Turnkey payment will be made to the manufacturer's agent in Indian rupees indicated in the relevant Price Schedule (as per prevailing rate of exchange ruling on the date of Contract) and shall not be subject to further escalation / exchange variation. Payment shall be made to the Indian Agent after 100 % payment to the Foreign Principal.

D) Payment for Annual Comprehensive Maintenance Contract (CMC) Charges:

The consignee will enter into CMC with the supplier at the rates as stipulated in the contract. The payment of CMC will be made on six monthly basis after satisfactory completion of said period, duly certified by the consignee on receipt of bank guarantee for an amount equivalent to 2.5 % of the cost of the equipment as per contract in the prescribed format given in Section XV valid till 2 months after expiry of entire CMC period.

- 21.2 The supplier shall not claim any interest on payments under the contract.
- 21.3 Where there is a statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the Supplier at rates as notified from time to time.
- 21.4 Irrevocable & non-transferable LC shall be opened by ESIC/ Purchaser. However, if the supplier requests specifically to open confirmed LC, the extra charges would be borne by the supplier. If LC is required to be extended and/or amended for reasons not attributable to the purchaser/consignee, the charges thereof shall be borne by the supplier.
- 21.5 The payment shall be made in the currency / currencies authorised in the contract.
- 21.6 The supplier shall send its claim for payment in writing, when contractually due, along with relevant documents etc., duly signed with date, to the purchaser.
- 21.7 While claiming payment, the supplier is also to certify in the bill that the payment being claimed is strictly in terms of the contract and all the obligations on the part of the supplier for claiming that payment has been fulfilled as required under the contract.
- 21.8 While claiming reimbursement of duties, taxes etc. (like sales tax, excise duty, custom duty) from the Purchaser/Consignee, as and if permitted under the contract, the supplier shall also certify that, in case it gets any refund out of such taxes and duties from the concerned authorities at a later date, it (the supplier) shall refund to the Purchaser/Consignee forthwith.
- 21.9 In case where the supplier is not in a position to submit its bill for the balance payment for want of receipted copies of Inspection Note from the consignee and the consignee has not complained about the non-receipt, shortage, or defects in the supplies made, balance amount will be paid by the paying authority without consignee's receipt certificate after three months from the date of the preceding part payment for the goods in question, subject to the following conditions:
- (a) The supplier will make good any defect or deficiency that the consignee (s) may report within six months from the date of despatch of goods.
 - (b) Delay in supplies, if any, has been regularized.
 - (c) The contract price where it is subject to variation has been finalized.
 - (d) The supplier furnishes the following undertakings:

"I/We, _____ certify that I/We have not received back the Inspection Note duly receipted by the consignee or any communication from the purchaser or the consignee about non-receipt, shortage or defects in the goods supplied. I/We _____ agree to make good any defect or deficiency that the consignee may report within three months from the date of receipt of this balance payment.

22. Delay in the supplier's performance

- 22.1 The supplier shall deliver the goods and perform the services under the contract within the time schedule specified by the Purchaser/Consignee in the List of Requirements and as incorporated in the contract.
- 22.2 Subject to the provision under GCC clause 26, any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:

- (i) imposition of liquidated damages,
 - (ii) forfeiture of its performance security and
 - (iii) termination of the contract for default.
- 22.3 If at any time during the currency of the contract, the supplier encounters conditions hindering timely delivery of the goods and performance of services, the supplier shall promptly inform the Purchaser/Consignee in writing about the same and its likely duration and make a request to the Purchaser/Consignee for extension of the delivery schedule accordingly. On receiving the supplier's communication, the Purchaser/Consignee shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.
- 22.4 When the period of delivery is extended due to unexcused delay by the supplier, the amendment letter extending the delivery period shall, inter alia contain the following conditions:
- (a) The Purchaser/Consignee shall recover from the supplier, under the provisions of the clause 23 of the General Conditions of Contract, liquidated damages on the goods and services, which the Supplier has failed to deliver within the delivery period stipulated in the contract.
 - (b) That no increase in price on account of any ground, whatsoever, including any stipulation in the contract for increase in price on any other ground and, also including statutory increase in or fresh imposition of customs duty, excise duty, sales tax/ VAT, Service Tax and Works Contract Tax or on account of any other tax or duty which may be levied in respect of the goods and services specified in the contract, which takes place after the date of delivery stipulated in the contract shall be admissible on such of the said goods and services as are delivered and performed after the date of the delivery stipulated in the contract.
 - (c) But nevertheless, the Purchaser/Consignee shall be entitled to the benefit of any decrease in price on account of reduction in or remission of customs duty, excise duty, sales tax/ VAT, Service Tax and Works Contract Tax or any other duty or tax or levy or on account of any other grounds, which takes place after the expiry of the date of delivery stipulated in the contract.
- 22.5 The supplier shall not dispatch the goods after expiry of the delivery period. The supplier is required to apply to the Purchaser/Consignee for extension of delivery period and obtain the same before despatch. In case the supplier dispatches the goods without obtaining an extension, it would be doing so at its own risk and no claim for payment for such supply and / or any other expense related to such supply shall lie against the purchaser.

23. Liquidated damages

- 23.1 Subject to GCC clause 26, if the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 24.

During the above-mentioned delayed period of supply and / or performance, the conditions incorporated under GCC sub-clause 22.4 above shall also apply.

24. Termination for default

- 24.1 The Purchaser/Consignee, without prejudice to any other contractual rights and remedies available to it (the Purchaser/Consignee), may, by written notice of default sent to the supplier, terminate

the contract in whole or in part, if the supplier fails to deliver any or all of the goods or fails to perform any other contractual obligation(s) within the time period specified in the contract, or within any extension thereof granted by the Purchaser/Consignee pursuant to GCC sub-clauses 22.3 and 22.4.

- 24.2 In the event of the Purchaser/Consignee terminates the contract in whole or in part, pursuant to GCC sub-clause 24.1 above, the Purchaser/Consignee may procure goods and/or services similar to those cancelled, with such terms and conditions and in such manner as it deems fit and the supplier shall be liable to the Purchaser/Consignee for the extra expenditure, if any, incurred by the Purchaser/Consignee for arranging such procurement.
- 24.3 Unless otherwise instructed by the Purchaser/Consignee, the supplier shall continue to perform the contract to the extent not terminated.

25. Termination for insolvency

- 25.1 If the supplier becomes bankrupt or otherwise insolvent, the purchaser reserves the right to terminate the contract at any time, by serving written notice to the supplier without any compensation, whatsoever, to the supplier, subject to further condition that such termination will not prejudice or affect the rights and remedies which have accrued and / or will accrue thereafter to the Purchaser/Consignee.

26. Force Majeure

- 26.1 Notwithstanding the provisions contained in GCC clauses 22, 23 and 24, the supplier shall not be liable for imposition of any such sanction so long the delay and/or failure of the supplier in fulfilling its obligations under the contract is the result of an event of Force Majeure.
- 26.2 For purposes of this clause, Force Majeure means an event beyond the control of the supplier and not involving the supplier's fault or negligence and which is not foreseeable and not brought about at the instance of, the party claiming to be affected by such event and which has caused the non – performance or delay in performance. Such events may include, but are not restricted to, acts of the Purchaser/Consignee either in its sovereign or contractual capacity, wars or revolutions, hostility, acts of public enemy, civil commotion, sabotage, fires, floods, explosions, epidemics, quarantine restrictions, strikes excluding by its employees, lockouts excluding by its management, and freight embargoes.
- 26.3 If a Force Majeure situation arises, the supplier shall promptly notify the Purchaser/Consignee in writing of such conditions and the cause thereof within twenty-one days of occurrence of such event. Unless otherwise directed by the Purchaser/Consignee in writing, the supplier shall continue to perform its obligations under the contract as far as reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- 26.4 If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of Force Majeure for a period exceeding sixty days, either party may at its option terminate the contract without any financial repercussion on either side.
- 26.5 In case due to a Force Majeure event the Purchaser/Consignee is unable to fulfil its contractual commitment and responsibility, the Purchaser/Consignee will notify the supplier accordingly and subsequent actions taken on similar lines described in above sub-paragraphs.

27. Termination for convenience

- 27.1 The Purchaser/Consignee reserves the right to terminate the contract, in whole or in part for its (Purchaser's/Consignee 's) convenience, by serving written notice on the supplier at any time during the currency of the contract. The notice shall specify that the termination is for the convenience of the Purchaser/Consignee. The notice shall also indicate interalia, the extent to which the supplier's performance under the contract is terminated, and the date with effect from which such termination will become effective.

- 27.2 The goods and services that are complete and ready in terms of the contract for delivery and performance within thirty days after the supplier's receipt of the notice of termination shall be accepted by the Purchaser/Consignee following the contract terms, conditions and prices. For the remaining goods and services, the Purchaser/Consignee may decide:
- a) To get any portion of the balance completed and delivered at the contract terms, conditions and prices; and / or
 - b) To cancel the remaining portion of the goods and services and compensate the supplier by paying an agreed amount for the cost incurred by the supplier towards the remaining portion of the goods and services.

28. Governing language

- 28.1 The contract shall be written in English language following the provision as contained in GIT clause 4. All correspondence and other documents pertaining to the contract, which the parties exchange, shall also be written accordingly in that language.

29. Notices

- 29.1 Notice, if any, relating to the contract given by one party to the other, shall be sent in writing or by cable or telex or facsimile and confirmed in writing. The procedure will also provide the sender of the notice, the proof of receipt of the notice by the receiver. The addresses of the parties for exchanging such notices will be the addresses as incorporated in the contract.
- 29.2 The effective date of a notice shall be either the date when delivered to the recipient or the effective date specifically mentioned in the notice, whichever is later.

30. Resolution of disputes

- 30.1 If dispute or difference of any kind shall arise between the Purchaser/Consignee and the supplier in connection with or relating to the contract, the parties shall make every effort to resolve the same amicably by mutual consultations. The jurisdiction for the settlement of disputes will be at New Delhi, India.
- 30.2 If the parties fail to resolve their dispute or difference by such mutual consultation within twenty-one days of its occurrence, then, unless otherwise provided in the SCC, either the Purchaser/Consignee or the supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided the applicable arbitration procedure will be as per the Arbitration and Conciliation Act, 1996 of India. In the case of a dispute or difference arising between the Purchaser/Consignee and a domestic Supplier relating to any matter arising out of or connected with the contract, such dispute or difference shall be referred to the sole arbitration of an officer, appointed to be the arbitrator by the Director General of ESIC. The award of the arbitrator shall be final and binding on the parties to the contract subject to the provision that the Arbitrator shall give reasoned award in case the value of claim in reference exceeds Rupees One Lac (Rs. 1,00,000/-)
- 30.3 Venue of Arbitration: The venue of arbitration shall be New Delhi, India.

31. Applicable Law

The contract shall be governed by and interpreted in accordance with the laws of India for the time being in force.

32. General/ Miscellaneous Clauses

- 32.1 Nothing contained in this Contract shall be constructed as establishing or creating between the parties, i.e. the Supplier/its Indian Agent/CMC Provider on the one side and the Purchaser on the other side, a relationship of master and servant or principal and agent.
- 32.2 Any failure on the part of any Party to exercise right or power under this Contract shall not operate as waiver thereof.

-
- 32.3 The Supplier shall notify the Purchaser/Consignee of any material change would impact on performance of its obligations under this Contract.
- 32.4 Each member/constituent of the Supplier/its Indian Agent/CMC Provider, in case of consortium shall be **jointly and severally liable** to and responsible for all obligations towards the Purchaser/Consignee for performance of contract/services including that of its Associates/Sub Contractors under the Contract.
- 32.5 The Supplier/its Indian Agent/CMC Provider shall at all times, indemnify and keep indemnified the Purchaser against all claims/damages etc. for any infringement of any Intellectual Property Rights (IPR) while providing its services under CMC or the Contract.
- 32.6 The Supplier/its Agent/CMC Provider shall, at all times, indemnify and keep indemnified the Purchaser/Consignee against any claims in respect of any damages or compensation payable in consequences of any accident or injury sustained or suffered by its employees or agents or by any other third party resulting from or by any action, omission or operation conducted by or on behalf of the supplier/its associate/affiliate etc.
- 32.7 All claims regarding indemnity shall survive the termination or expiry of the contract

SECTION – V**SPECIAL CONDITIONS OF CONTRACT (SCC)**

The following Special Conditions of Contract (SCC) will apply for this purchase. The corresponding clauses of General Conditions of Contract (GCC) relating to the SCC stipulations have also been incorporated below.

These Special Conditions will modify/substitute/supplement the corresponding (GCC) clauses. Whenever there is any conflict between the provision in the GCC and that in the SCC, the provision contained in the SCC shall prevail.

For GCC Clause No. 15.8:

After completion of Warranty period Annual Comprehensive Maintenance Contract (CMC) to be quoted as mentioned in General Technical specifications Section VII (Para-4) for all the items except for Item Sl. No. 1, 16 & 46.

SECTION - VI
LIST OF REQUIREMENTS

Part I:

Sl. No.	Item Description	Total Qty.
1	Abdominal hystrectomy set	2
2	Ambulatory BP monitor	4
3	Automated Visual Field Analyzer with printer (Perimeter)	1
4	Automatic film processor	1
5	Baby warmer	4
6	Bilirubinometer	1
7	Bio-chemistry semi auto analyser	1
8	BIPAP	1
9	Biphasic Defibrillator	5
10	Blood cell counter-Fully Automated 3 Part Analyser	1
11	Blood gas analyser	1
12	Coagulation Analyser	1
13	Combined A & B Scan Machine	1
14	CPM	1
15	CR system (Computer Radiography System)	1
16	Delivery & caesarean set	2
17	Dental Unit Complete	1
18	Diagnostic X-ray Machine 300m.A	1
19	Diognostic X-ray Machine 500 m.A	1
20	ECG machine 12 channel	2
21	Electrolyte analyser	1
22	Emergency resuscitation Kit	2
23	Fully automated ELISA reader with washer	1
24	Fully Automated random access biochemistry analyser	1
25	Holter Recorder & Analyser	1
26	Infant Resuscitator	1
27	Laryngoscope (Fibreoptic)	1
28	MOBILE C-ARM IMAGE INTENSIFIER WITH DSA	1
29	Multipara bed side monitors with Central station	10(1)
30	Neonatal monitor	2
31	Oesophagoscope Adult	1
32	Operating Zoom microscope for ophthalmology	1
33	Patient Warming system	1
34	Phaco Emulsification Machine	1
35	Pharmaceutical Refrigerator for OT	3
36	Portable ventilator	1
37	Portable X-ray machine (Mobile)	2

Sl. No.	Item Description	Total Qty.
38	Pulse oximeter	7
39	Slit Lamp	1
40	Suction machine(portable)	2
41	Syringe infusion pump	5
42	TMT	1
43	Traction unit	1
44	UVB Chamber	1
45	Vacuum extractor	1
46	Vaginal hysterectomy set	2
47	Volumetric infusion pump	7
48	800mA X-ray machine with fluroscopy with IITV	1
49	O T Tables for General Surgery & Gynaecology	5
50	O T Tables for Orthopaedics	1

Part II: Required Delivery Schedule:

a) For Indigenous goods or for imported goods if supplied from India:

Within **60 days** from date of Notification of Award to delivery at consignee site. The date of delivery will be the date of delivery at consignee site (Tenderers may quote earliest delivery period).

b) For Imported goods directly from abroad:

Within **90 days** from date of opening of L/C. The date of delivery will be the date of delivery at consignee site (Tenderers may quote earliest delivery period).

Part III: Scope of Incidental Services:

Installation & Commissioning, Supervision, Demonstration, Trial run and Training etc. as specified in GCC Clause 13. Installation & Commissioning shall be completed within 15 days of handing over the site of installation, complete in all respect by the consignee. The date of handing over of the site has to be intimated by the supplier to the purchaser. The delay on the part of the supplier to install and commission the equipment will attract the provisions as contained in the liquidated damage clause.

Part IV:

Turnkey (if any) as per details in Technical Specification. The tenderer shall also specify the time schedule for completion of Turnkey work.

Part V:

Annual Comprehensive Maintenance Contract (CMC) as per details in General Technical Specifications para 4.

Part VI:

Required Terms of Delivery and Destination.**a) For Indigenous goods or for imported goods if supplied from India:**

Delivery required at Consignee Site.

Insurance (local transportation and storage) would be borne by the Supplier from warehouse to the consignee site for a period including 3 months beyond date of delivery

b) For Imported goods directly from abroad:

The foreign tenderers are required to quote their rates on DDP at consignee's site basis giving break up of the price as per the Proforma prescribed in the Price Schedule.

Custom clearance, handling, unloading & loading and transportation to the consignee's site shall be the responsibility of the supplier/ Indian agent.

Insurance (local transportation and storage) would be extended and borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery.

c) Destination/Consignee details are given in Section XXI**Part VII:****Inspection:****a) For Indigenous goods or for imported goods if supplied from India:**

At consignee site by the respective In-charge of the Hospital or his authorised representative.

b) For Imported goods directly from abroad:

Pre-despatch inspection at manufacturer's premises as per GCC clause 8.8 and inspection after arrival in India at consignee site by the respective In-charge of the Hospital or his authorised representative.

Section – VII

Technical Specifications

1. Abdominal Hysterectomy Set

- 1 Sponge Holder 4
- 2 Needle holder 3
- 3 Bab cock 2
- 4 Artery forceps 6.1/2" 10
- 5 Artery forceps 7.1/2" 4
- 6 Allis Forceps 6
- 7 Straight Hyst. Clamp 5
- 8 Curved clamp 5
- 9 Mosquitoes forceps 4.1/2" 6
- 10 Doyn's Retractor Big 1
- 11 Self Retaining Forceps (Balfour) 1
- 12 Deverse Small 1
- 13 Deverse Big 1
- 14 Towel clip 5
- 15 Steel Basin 2
- 16 S.S Bowl 3
- 17 Steel Tray Big 2
- 18 Myo curved scissor 8" 2
- 19 Straight scissor 6.5" 2
- 20 Tooth dissecting forceps 6.5" 1
- 21 Non Tooth dissecting forceps 6.5" 1
- 22 Metzenbaum curved scissor 8" 2

2. Ambulatory Blood Pressure Monitor

Monitor must be validated by BHS
Monitor must take less than 4 AA batteries
Must be able to programme at least 6 time intervals
Must have wWindows xp XP compatible software.
Monitor shall be able to measure down to 30mmhg for diastole
ABMP must be able to interface to computer using a serial or usb USB cable
Software must be able to email report as a pdf .
Software must be able to analyze data over 48hrs
Software must allow easy selection section of data to be for analysis
Should have US FDA certification
At least 5 sizes of cuff must be available for use with ABMP
Software will have inbuilt security with easily accessible log of users.

3. Automated Visual Field Analyzer with printer (Perimeter)

High quality goldman standard Imported automated full field perimeter

with bowl size 30cm. Computer monitor should be inbuilt with the perimeter.

Clinically validated normative databases(FDA approved).

- Maximim intensity 10,000Asb,Bowl illumination 31.5Asb
 - Floppy drive ,Internal hard disk drive with future upgradation to MOD
 - Stimulation duration 200ms,wavelenth Broad band visible light
 - Stimulus/Background colour White on White
 - Maximum temporal range 90Deg.Suitable for central 30 as well as full field testing
 - Central field test patterns 30-2,24-2,10-2,Macula
 - Peripheral field test pattern 60-4,Nasal Step
 - Thresold test strtegies full thresold,Fast Pac,SITA,SITA Fast,SITA Standard
 - Screening field test P-60,FF-80,FF-120, FF-240,Nasal Step for periphery .
 - Screening test strategies Two zone,Three Zone and Quantify Defects
 - Custom Test
 - Stimulus Size I-V as per goldmann standards
 - Glaucoma hemifield test,Heijl –Krakau blind spot moniter
 - Video eye monitering,Trial Lens Holder,
- Touch screen on CRT as well as Keyboard & Mouse
- Motorised chinrest, Motorised table with Laser Jet Printer

OPTIONAL

Glaucoma progression analysis software

4. Automatic film processor

Heavy Duty Automatic Film Processening Unit

Automatic X-Ray Film processor for processing of all standard medical X-Ray and Imaging Films including Ultrasound,. Mammography and CT scan films. It should have the following features:-

1. Processor should be able to take the maximum film width and infeed width shall be 45 cm and minimum of 10cm x 10cm .
2. High capacity through put of apprx.200 Films/Hours of size 14" x17"
3. Transport speed adjustable between 1.5 minutes to 5.0 minutes (in the step of ranging between 5-8 seconds for perfect quality control).
4. The developing and fixing tanks should be of 10 litres capacity for proper developing & washing. Washing tank should be of somewhat higher capacity.
5. Replenishment should have the individual programmes, manual & automatic with tank capacity of 25 Litres.
6. The roller transport system should have squeeze roller and the main drive should stop automatically when not in use.
7. The container assembly should be monoshell and made of material which is non corrosive and of latest technology.
8. Thermostatic controlled temperature of developer and it should have range between 28Cto 40 C.
9. Multiple program memory should be there preferred for different imaging Applications. At least four program memory must be there.
10. There should be at least four infrared sensors for film detecation.
11. Replenishment rate should be adjustable, which should be mentioned
12. Dryer temperature range should be around 70 C for better drying.
13. Water consumption during use should not exceed 2Lit per minute. Lower consumption shall be preferred.
14. It should have anti- oxidation programme option in between adjustable range of 10-90 minute intervals and should have anti oxidation and anti-crystallization mode.

15. It should have separate replenishment pumps for developer & fixer to facilitate selection of different replenishment rates.
16. Film output should be possible through the well (wall mounted). The processor should operate on 230 volts, 50Hz ., 10 Amperes. Firm to quote an appropriate voltage stabilizer. Details about availability of spares & service should be given.
17. Audible alarm with display of error message to be available.
18. The company should provide two sets of chemicals for initial operation of the machine.
19. Space requirement and installation requirement and power requirement for the equipment should be mentioned.

5. Baby Warmer

Infant warmer to be used in neonatology.

The unit should conform all relevant international, national and local standards.

Specifications

Temperature control:

- Range 30-38° C
- Skin range 25 – 42 °
- Increment 0.1 °
- Display Digital

Control Unit (to be supplied with.)

- Automatic heat control type
- Set point mechanism
- Heater Indicator.

Alarms (Audible and Visual)

- High air temperature
- Sensor disconnect
- Power Failure

Alarm in manual mode: every 15 minutes with automatic shutoff

The warmer should includes:

- Self- check features
- Breaks for casters
- Skin sensor
- Supplemental humidity
- Protection against breaks and bursts of radiant and light source
- Spares and accessories
- Service and users manuals

Accessories:

- No. of hand ports 6
- No. of tubing ports 6
- No. of oxygen inlet port 1
- Backup thermostat

Examination Light 50 W Halogen

Radiant heat source Quartz tube 600w

- Phototherapy lights
- Resuscitation equipment packages
- X-Ray cassette holder

6. Bilirubinometer

Microprocessor Controlled

MAIN CHARACTERISTICS: -

- Auto zero Function
- Handles hemolysis and turbidity easily
- Easy set of the sample tube
- Alarm lamp informs user of abnormalities
- Flexible power source
- Easy lamp replacement

TECHNICAL SPECIFICATION: -

- Filters :461nm & 551nm
- Measurement Range: 0-30 mg/dl (Total Bilirubin)
- Correcting hemolysis: 0-250 mg/dl HbCV
- Measuring accuracy :+5%
- Sample Volume
- Alarm Display : 3-1/2, 7-segment red LED
- Sample container : Hematicrit capillary tube

- Light Source : 6V, 1.5A tungsten lamp
- Photocell : Silicon photocell
- Power supply 90-240 VAC 50/60Hz, 35w
- Dimensions :Approx.280mm (w) x230mm (D) x120mm(H)

Suitable centrifuge to be quoted

7. Bio-Chemistry Semi-Auto Analyzer

Automatic flow-through Cell Analyzer
 Easy to use high quality precision Analyzer
 Monochromatic and Bichromatic measurements
 (340 - 630) °
 Interference filters 340-405-500-546-578-630nm
 Bandwidth 8 nm.
 Optical measuring pass of 10 nm.
 Measures in absorbances, Concentration kinetics,
 Fixed Time, Rate and Differential MULTI-STANDARDS
 Zero set: fully automatic
 Fully programmable directly from the keyboard
 Storage capacity - 100 complete tests.
 Reading volume need: 500 ul
 Disposable cuvette programmable with flow through
 Cell: 500- 1000ul
 Built in printer
 Thermostated block by peltier effect at 25-30-37 °C
 for 10 cuvettes.
 External interface to be connected to computer.
 WITH STANDARD ACCESSORIES

8. BIPAP

Non invasive ventilation/ bipap with LCD display of parameters with following features

- IPAP range 6-30 cms
- EPAP range 4-20 cms
- Respiratory rate can be set to 4-40 bpm
- Spontaneous / CPAP/ spontaneous with time/ timed
- Connectors for et ET tube available for direct connection
- Should have CPAP, PSV ST, PCV, PACV modes
- Should have facility to set target tidal volume
- Inspiratory time should be possible to set allowing critical patients to breathe out.

The rise time from EPAP to IPAP can be set and varied

The system should be supplied with operator's manual, ultra mirage mask (full mask) , hose pipe and power cable

9. Biphasic Defibrillator

- Biphasic
 - Waveform display (min 3)
 - External energy selection from 2 J to 200 J, biphasic
 - Charging time less than 8 seconds @ 200 J (with a charged battery)
 - Synchronizer and cardio version
 - Unique disarm button (in addition to automatic time delay)
 - Should come with high resolution monitor.
 - Should come with 3 lead ECG that can be measured from cables, adult external paddles, paediatric paddles.
 - Heart rate: 20 to 300 bpm BPM with user selectable alarms.
 - Should come with external pacing, demand and asynchronous modes
 - Should display CPR in real time
 - Should have large internal memory that stores and prints 25 ECG events.
 - Battery to last not less than 2 hours of continuous ECG monitoring or 30 full energy discharges.
 - Battery indicator on display and self test on battery
- Accessories needed
- Paddles with remote energy selection, charge and discharge buttons on paddles.
 - Pediatric paddles adapters (set of 2)
 - A/c 240 v V 50 hz charger and mains power source. Adult paddles and test paddles.
 - 3 lead patient cable
 - One spare battery
 - Roll of 50mm recording paper x 10 rolls
 - 5 oz tube of defibrillation gel x 5 tubes
 - ECG cable with leads- 2 sets.
 - External disposable pacing pads-2 sets
 - Operation manual, service manual complete.
- Should have USFDA Approval

10. Blood cell counter-Fully Automated 3 Part Analyser

Measures 18 parameters including differential leucocyte with only 20 microlitre of blood

Backlit and touch screen LCD display.

All 18 parameters alongwith three histograms are displayed

In built thermal printer, and facility of external also.

Automatic sample prob wipe.

Automatic prob parking in side the machine

Automatic orifice cleaning with back flush.

Possibility of shifting discriminators manually for particle size analysis.

Should have zero back ground count at any time.

Should have two seperate chambers for WBC & RBC measurement.

Should have long life led, sensors & maintenance free membrane valves.

Should have throughput 60 test / hour

Should have reagent level monitoring system/ alarm.

Should have enhanced WBC/platelets measuring range to accommodate high WBC/Platelets counts in cases of leukamia/blood bank samples

Should have local service station for maintenance of equipment

Should have power back up system in case of power failure.

Per sample cost should be lowest.

11. Blood Gas Analyzer

• A fully automated pH/Blood gas/electrolyte analyzer measuring the following parameters: - pH, PCO₂, PO₂, Barometric pressure.

Na, K, Ca, Cl

Co-oximetry: ct Hb, CCO Hb, Met Hb, Sulf Hb, Haematocrit and Barometric pressure.

• Sample volume should be approximate 100 µl for all parameters.

• All calibration and cleaning cycles should be fully automated with user selectable calibration items.

• Calibration should be performed by liquid calibration for all parameters.

• The electrodes provided should be zero maintenance including the reference electrode.

• The system should have on board data manager to store all patient results, QC data and calibrations.

• The system should have a closed waste system and monitored continuously. Also all the system reagents should be monitored continuously.

• A power fail protection for 20 min. to take all calibration and programmed data.

• The analyzer should have a colour LCD screen to access all the system software and to display

the patient's results. With alphanumeric key board/touch screen.

- A built in thermal printer should be provided to print out patient results.
- The system should work in discrete testing, ie, selectable parameter testing.
- Should be supplied with consumable, reagents and QC agents for 1000 tests, as per the user requirements so that they do not expire.
- Should not preferably use special gases

12. COAGULATION ANALYZER (COAGULOMETER)

Fully automated, open random access, 4 detector channel Blood Coagulation Analyzer having clotting, chromogenic and immunological assay channels. Clotting detection methods by recording change of viscosity and/or change of light transmittability and permitting both chromogenic and immunogenic assays

Parameters:

Provision for simultaneous random analysis for at least 15 parameters such as PT, APTT, Fibrinogen,

Protein S , Protein C, APC resistance, ProC, Antithrombin III, Heparin, Plasminogen Activator inhibitor,

D-Dimer Plus, von Willebrand Antigen, Lupus Anticoagulant, and coagulation factors VIII & IX .

Should have throughput of at least 100 tests per hour for PT.

Should be provided with a compatible printer.

Should have a intra assay reproducibility of:

Normal:

PT<2.0%

APTT<2.0%

Fibrinogen<4%

Patient:

PT<2.0%

APTT<2.0%

Fibrinogen<5%

Additional Accessories required :1) All standard accessories 1 set extra.

2) Compatible Online UPS with1 hour back up.

13. Combined A & B Scan Machine

The following requirements must be met

- High resolution dedicated A and B, ophthalmic scanning unit B scan will cross vector.
- The system should consist of fourth generation microcomputer and high speed digital electronics, with highest

resolution monitor.

Technical features:

- A-scan
- Three a scan modes
- Auto biometric, manual biometric, diagnostic

- Complete IOL program capabilities include SRK1 SRK11 SRK. T Hollady or Binkhorest formulas.

- Save in memory capacity at least 45 cases for a-scan images and corresponding IOL data.
- 10mhz solid probe

- The unit should incorporate, audio feed back for probe alignment.
- B-scan
- 256 gray levels
- User definable, DGC curve
- Pre & post processing capabilities.
- Volume, distance and area/ perimeter measurement
- Selectable a-vector for simultaneous A/B display.
- Annotation/arrow placement
- Archiving of at least 150 patients in a single data file with an unlimited number of data files possible.
- Complete IOL calculation capability with IOL data storage.
- B-scan sector angle at least 55°
- Standard accessories
- Should include :
- Console with 7" display
- Alphanumeric keyboard
- Trackball
- Foot pedal
- 7&10 MHz, A-B scan probe
- A scan calibration cylinder
- Probe holders etc.
- Vendors may quote other accessories
- Standard accessories should include :
- Console with 7" display
- Alphanumeric key board
- Trackball
- Foot pedal
- 100 & 12.5 MHz, A-B scan probe
- A scan calibration cylinder
- Probe holders etc

14. CPM Unit

The lower extremity Continuous Passive Motion machine including accessories

Range of Motion

Hyperextension - 10° to flexion 120°

Speed: 30°/minute to 160°/minute

Dimensions: 94 x 33cm (L x W) approx.

Voltage: 100 to 240 volts, 50/60Hz
Patient Sizing: Full leg 71 to 104cm Tibia 38 to 58cm, Femur 33 to 46cm
It should conform to the international standards ISO 13485, CE certified.
The Upper extremity Continuous Passive Motion machine including accessories
Elbow
CPM with base including accessories
Range of motion:
Flexion: 135°
Extension: 0°
Pronation/Supination: 90°
Speed (cycle of 0°-135°): 3 min 50 sec to 1 min
Patient height: 140 to 190cm
Dimensions: 72 x 65 x130cm approx.
Voltage: 110/240 volts 50/60Hz
Upper extremity CPM including accessories
Provides increased ROM and greater patient comfort
With improved control features such as the ability to make quick adjustments, pause features
A visual biofeedback mechanism, and progressive protocols for consistency
The machine can be used for isolated or synchronised movement training
Hand and wrist CPM should have following accessories
Allows simultaneous motion of the three phalanges, matching the physiological spiral of flexion
and allowing the formation of a true composite fist
Versatile and cost effective: accommodates the fingers, thumbs and wrist (either right or left hand ; from the smallest to the largest adult)
Easily portable for use in the hospital room, therapy
It should conform to the international standards ISO 13485, CE certified.
Patient safety switch.
Timer - 0 to 99 min.
Hold period - 0 to 90 sec.
Buzzer - buzz after preset time adjusted.

15. CR System (COMPUTER RADIOGRAPHY SYSTEM)

Specifications for State of the art Latest Generation Computed Radiography (CR) system for high resolution Digital radiography

Features Remarks

Technical Requirements - CR system configuration shall include:

- a) Imaging plates (IP)
- b) Image reader system
- c) CR workstations
- d) RIS interface

- e) Remote ID and Preview stations
- f) Accessories and consumables
- g) Laser Imager

CR Compatible imaging plates

Following sizes & qty. are required as standard (Unit rate of each CR plate to be Quoted separately for any additional requirement)

- a) 35 cm x 35 cm - 12Nos.
- b) 24 cm x 30 cm - 12 Nos.
- c) 18 cm x 24 cm - 12 Nos.
- d) 15 cm x 30 cm - 6 nos.
- e) 35 cm x 43 cm – 6 nos.
- f) 30 cm x 37.5 cm – 6 nos.

Image plate storing Rack- one (Price to be quoted separately)

Image reader shall meet the Functional requirements:

- a) Various image-processing protocols available for the respective regions of the body
- b) IP processing rate should be about 90 plates / hour
- c) Mechanism for accepting exposed Imaging Plates with out patient demographics, for Causality /Trauma workflow requirement
- d) Mechanism for Re-routing the newly acquired Images to the preconfigured CR workstation
- e) Capability of retrieving (Service Intervention) at least last 10 scanned images, as part of
- e) Capability of retrieving (Service Intervention) at least last 10 scanned images, as part of contingency plan.
- f) Capability for quick check of the image and exam data of at least the last 4 Imaging Plates scanned at the x- ray room
- g) Protocol for verifying the connectivity status of configured image destinations
- h) Spatial resolution of the digital image shall preferably be 2kx2kx16 bits for optimal resolution.

Identification and Preview

System Functional requirements:

- a) Capability of interfacing to HL7, Non-HL7, Proprietary, DICOM Work list or user defined Windows/DOS /Linux based interface protocols to HIS/RIS.
- b) Please specify whether you have tested interfacing with HL7-DICOM Bridge.
- c) Mechanism for retrieving Demographics of at least last 10 patients identified on a particular Identification Terminal.
- d) Customizable Graphic User Interface (GUI) in Identification station with facility of selecting DICOM print & Storage destination.
- e) Indication of Over Exposure on the preview module.
- f) Mechanism for User release from Preview terminal in case of Auto-routing Images to Predefined DICOM Destinations.

- g) Customizable Graphic User Interface (GUI) for Preview terminal.
- h) Solution for storing patient demographic data for multiple exams in RIS/non RIS environment.
- i) It should be possible to put a custom configurable data field in the demographic information of the patient linked with the image.

Software

System should include the following Software applications:

Please list all the optional software(s) which are available with you for enhancing the workflow

and service in the Digital Radiology environment for the following

1. Advanced Processing Software
2. Application Software
3. Connecting Software
4. Visual Output Software
5. Quality Monitoring Software

The system should include the following SW applications as standard:

Full Leg/Full spine image processing.

Quality Control software.

1. Software, which enables to see in the preview terminal the deviation from normal exposure and with the details of the deviation on the CR workstation.
2. Software masking of the collimation areas.
3. Special attention should be placed on pediatric applications.
4. Software for storing images on any DICOM 3 (or newer versions) compliant stations.
5. Software for printing on any DICOM printer.

CR Workstation

System configuration requirements:

- a) Accept images from CR Reader without any loss of data
- b) Capable of Archiving & Printing selected image to a standard DICOM destination in DICOM 3.0 Format.
- c) Storing images in the local disk for pre-defined period.
- d) Mechanism for accepting New images when the local disk is full
- e) Should include 21" antiglare flicker free TFT/LCD color monitor (1.2K X 0.78K resolution)
- f) Should include 21" Monochrome antiglare flicker free Medical Grade TFT/LCD monitor with at least 2k X 2k resolution.
- g) DVD Burner
- h) 240 GB or more on board storage

System Functional requirements:

- a) Support DICOM Work list or user defined Windows/Dos based interface to HIS/RIS

-
- b) Mechanism for retrieving Demographics of atleast last 10 patient identified on that Terminal.
- c) Customizable Graphic User Interface with facility of selecting DICOM print & storage destination.
- d) Indication of Over Exposure on the preview module.
- e) Mechanism for User release in case of Auto-routing Images to Pre-defined DICOM Destinations.
- Functional requirement for CR workstation:
- a) Built in routine for using predefined image processing parameters for image quality enhancement.
- b) Mechanism for storing the Patient image based on name, date, exam, etc.
- c) Capability of storing user defined image processing parameters.
- d) Capability of overwriting predefined image parameter with user-defined parameters & storing these two images separately.
- e) Correcting typographically in Patient Demographic module, in case the RIS connection was down and manually data entry was done.
- f) Capability of changing W/I, Flipping, Rotating, Zooming, Collimating Annotating incoming image.
- g) Auto-routing incoming image to predefined DICOM Store (SCP storage) or Print Destination (SCP Print Destination)
- h) Mechanism for printing Multiple Images in one film, with the possibility of slide and True Size printing
- i) Compatible DVD Writer along with relevant software to be quoted separately.
- Laser Imager System Configuration requirements:
- Print Images from CR Workstation
- a) Capable of Printing Images in DICOM 3.0 format
- b) Mechanism to print images 14x 17,11X14, 8 x 10 film sizes simultaneously.
- c) Resolution should be 500 dpi or more
- d) Capable of handling mammography plates.
- Functional requirement for Laser Imager:
- a) Capable of Printing images in High quality
- b) Mechanism for printing images in 14 x 17,11X14 and 8 x 10 film sizes simultaneously.
- c) Mechanism for Printing Multiple Images in one film, with the possibility of slide printing.
- Provision for Distributed CR System should be present. Please quote price separately for additional workstation image reader preview stations and image planes.
- Warranty/AMC: All items and accessories should covered under warranty

Please list all the Optional software's, which are available with you for enhancing the workflow and services in the Digital Radiology environment.

Price to be Quote separately for additional laser imagers.

Price for On line UPS with one hour back up for complete system should be quoted.

System should have CE/FDA approval

Review station at key areas – qty 04 nos. (in OPD, OR, DOCTOR'S room etc.)

(Unit Price of review stations to be quoted separately)

PC based DVD reader image manipulating software and high definition monitor (1.2K ×0.78K) (approx).

Acceptance tests as per International Standard should be carried out at manufacturing facility as well as installation site (including all Safety and QA tests)

16. Delivery and Ceasarian set

B.P.Handle No.3	1
B.P.Handle No.4	1
Dissecting Forceps Tooth 15cm	2
Dissecting Forceps Plain 15cm	1
Scissor Mayo St. 18cm	1
Scissor Mayo St. & Curved 16cm	1 Eac h
Scissor Episectomy	1
Needle Holder 18cm	2
Placenta Puncture Forceps	1
Sponge Holder 20cm	2
Sponge Holder 26cm	1
Anterior Vaginal Retractor	1
Artery Forceps Curved 16cm	2
Cord Clips	1
Sim's Speculum Medium	1
Sim's Speculum Large	1
Towel Clip	4
S. S. Bowl 6 cm	2
Big Kidney Tray	1
Allis FORCEPS 8 “	2
Mosquito Artery cd	2
Ceasarian set	
B.p handle no-3	1
B.p handle no-4	1
Dissecting forceps fine toothed 14cm	1
Dissecting forceps adson plain 12.5cm	1

Dissecting forceps adson toothed 12.5cm	1
Dissecting forceps Atragrip 20cm	1
Dissecting forceps Atragrip15cm	1
Dissecting forceps plain 15cm	1
Dissecting forceps toothed 20cm	2
Cord clip	1
Towel clip backhus	5
Allis forceps 16cm	6
Allis forceps 20cm	2
Babcork forceps 16cm	2
Babcork forceps 20cm	2
Artery forcpe st 16cm	8
Artery forcpe cd 20cm	4
Artery forcpe cd 16cm	12
Needle holder 18cm	1
Needle holder 20cm	1
Needle Holder15cm	1
scissors mayo cd 16cm	1
Scissors mayo st 20cm	1
Episectomy scissors	1
McIndo scissors cd 23cm	1
Suture cutting wave scissors 18cm	1
Uterine scissors (sims) 20cm	1
kocher's clamp cd	2
Kocher's clamp st	4
Green Armitage forceps	6
Sponge holder 26cm	4
Yankurs suction tip	2
Frizer suction tip cannula NO-3	1
C' shaped retractor	3
Doyans retractor	2
Morries retractor	1
Doyen's skin clip	1
Uterine sound	
Bladder sound	
lengenback retractor (right angle)	
Allis forceps 20cm	4

Aneurysm needle	1

17. Dental Unit Complete

1. Dental Chair

The chair should be designed to provide good ergonomics, hygiene and aesthetics. The design also

enables the operator to be close to the patient so as to provide optimum vision of the operating field and

safe control of all component devices.

1.1 Fully motorized, hydraulically driven which give smooth start when switch is activated.

1.2 Supplied with 8 button footswitch for user friendly.

1.3 The backrest should be thin. Choice of Back rest

1.31 Contoured back rest.

1.32 Slim back rest or

1.33 Slim back rest with arms slings

1.4 Fixed small arm rest, for easy slide in/out for patient to provide support to get up from the chair. Should have the facility for Pivotal Arm rest.

1.5 Height range should be from 14" to 29".

1.6 Base plate should be Cast Aluminum. Plus a tough urethane coating that enhances corrosion resistance and protects treatment room floors

1.7 Upholstery can be cleaned with disinfectant Solution.

1.8 Chair should have safety brake system while going down for patient exit position.

1.9 Chair should have multipurpose double articulating head rest for ease of adjustment for pediatric patients & should be reversible for wheel Chair patient.

1.10 Chair should have minimum 4 programs. Two patient entry programs, one rinse program & one patient exit program.

1.11 Should have integrated 80 Watt power supply for Fibre optic Handpieces, Piezo Scaler, electric motor etc.

2. Dental Unit

2.1 Should be side delivery system. Should rotate to 270o

2.2. Handpiece control block should be flow-through water design to eliminate stagnant water.

2.3 Built-in anti-retraction valves and flush valve system for infection control.

2.4 Autoclavable Quick Disconnect 3 in 1 water syringe.

2.5 2 nos 3 hole tubing for Air Turbine and Air/Micro motor with straight & Contra Handpieces.

2.6 All water tubes should be AlphaSun. An antimicrobial additive that becomes part of the tubing material itself and inhibits the ability of bacteria to attach to the tubing surface by ionic silver exchange

2.7. Ultrasonic Piezo scaler

2.8.1 Wide power range from 2VA to 26 Va

2.8.2 Auto power control adjusts instantaneously to resistance at the tip

2.8.3 deleted.

2.8.4 Wide procedural applications from prophy to endodontics.

2.8 Doctor's Touchpad should have dual operator modes.(Optional)

2.9 Tray holder

3. Cuspidor

3.1 High and low Air suction. The suction system should be with separating tank and automatic self

cleaning separating tank to allow continuous draining.

3.2 Autoclavable

saliva ejector.

3.3. Autoclavable High volume evacuator

3.4 Autoclavable syringe.

3.5 Suction should be Air Ventury (should work on compressed air)

3.6 Spitton bowl should be of poly urethane material with cup fill and bowl rinse timers.

3.7 Clean water bottle System.

4. Operating light.

4.1 Light head includes two axial movements – Horizontal, Vertical adjustment.

4.2 Should have normal auto mobile halogen bulb

4.3 Feathered-edge, balanced-intensity light pattern

4.4 Optically designed reflector

4.5 Two-position intensity switch with high and low settings.

4.6 Color temperature to be 5,000 K.

4.7 Light Intensity :, Low 18,000 lux, High 24,000 lux.

4.8. BTU's per hour should be 325.

5. Doctor's Stool

5.1 Cast-aluminium base with five tile casters

5.2 Two-way adjustable lumbar support

5.3 Integral gas cylinder for height adjustment.

5.4. Height range 470mm – 635mm

6. Assistant's stool

6.1 Cast-aluminium base with five tile casters.

6.2 Height adjustable torso support with height adjustable foot-ring

6.3 Integral gas cylinder for height adjustment.

6.4. Height range 572mm – 737 mm.

7.1 Contra angled hand piece - 1 no.

7.2 Straight hand piece - 1 no.

7.3 Contra angled micro motor - 1 no.

7.4 Fibre optic hand piece - 1 no.

7.5 Light cure unit - 1 no.

7.6 Compressor - 1 no.

8. Certification (for all the above)

8.1 US-FDA approved.

8.2 CE-approved

8.3 ISO-approved

8.4 EMC test certificate

8.5 BSI Service Quality certificate

18. Diagnostic X-ray Machine 300m.A

Diagnostic Xray unit suitable for radiography with motorized table top complete with all essential accessories as detailed below :

1. Xray Generator

Full wave rectified 24 KW Xray Generator capable of delivering an output of 300mA at 100 kV

- Minimum exposure time of 10 millisecond

Digital display of kV, mA and mAS

- Radiographic kV selection from 45kV to 125 kV

2. Xray Tube (one nos.)

- Nominal voltage 125 kV

- Rating of 20/40 kW with rotating anode of BEL Make.

Anode braking device for increasing the life of x-ray tube.

Preheating of small & large filament of the x-ray should be there to increase the life of the x-ray tube.

3. Examination Table

With motorized table

- Very high ratio Grid .

Examination table: Instead of motorized table, it has to be a Bucky table(with high grid ratio) with motorized movement of table top with liberal movement in both horizontal (both ways) and vertical direction (both head end and foot end)

Power requirement for operation of the machine to be quoted by the firm.

the machine should have inbuilt voltage compensation and the machine should be able to operate without the need of additional external voltage stabilizer.

4. Radiographic Wall stand

Floor mounted Radiographic wall stand

- Vertical movement of at least 100 cm
- Should be capable of taking Cassette of sizes 5"x7" to 14" x17"

Very high ratio Grid

5. Column stand

Floor mounted column stand.

- Rotation of column stand full 360deg.
- X ray tube rotation +/- 180deg.

6. Accessories

Sinus cone

Mastoid cone

Dental cone

Power requirement, Space requirement and installation requirement for the equipment should be mentioned.

AERB/ BARC approved.

19. Diagnostic X-ray Machine 500m.A

A. Generator:

1. Generator should be high frequency/inverter type for constant output.
2. Output 50 kw or more.
3. Kv range 40 kv – 120 kv.
4. Output at 100 kv should be 500 ma or more.
5. It should have automatic exposure control device.
6. It should have digital display of kv & mas.
7. Anatomical programming radiography should be possible.

8. It should have over loading protection.

B. X – ray tube and collimator:

1. The x-ray tube should be rotating anode high speed, compatible with the generator and must have dual

focus. Focal spots of following sizes:

Large focus: 1.2/2.0 mm or better.

Small focus: 0.6/1.0 mm or better.

Tube with anode heat storage capacity 150 kHu or more.

2. Motorized collimator having additional filters (for dose reduction) and auto shut provision for the light.

C. X – ray table:

1. Horizontal table with floating table top.

2. It should have transverse + 10 cm or more and longitudinal movements + 35 cm or more with electromagnetic brakes.

3 It should be thin table top.

4. It should be provided with bucky which can hold all standard sizes of cassettes upto 14"x17".

5. Bucky should have a grid ratio 10:1 or more with 65 lines per cm.

D. Vertical bucky stand:

1. It should have provision to do chest radiography with grid.

E. Essential accessories: the following essential accessories to be provided with the unit.

1. Lateral cassette holder – one.

Should be supplied with following:

- Whole body (upper and lower) lead apron (standard make) – 8 nos.
- Lead screen (with lead glass window)– 2 nos.
- Lead lining for walls and door
- Apron & glove hanger

Heavy duty multiple glove and apron hanger wall or floor mounted.

Stainless steel construction. Holds two pairs of gloves and two lead aprons.

Dimensions: 76 x 46 x 25 cm approximately.

All necessary aerb / barc approvals.

20. ECG Machine 12 channel

- Simultaneous acquisition of up to 12 leads.
- Real time continuous recording of 3, 6 and 12 channel.
- Recording speeds of 10, 25 or 50 mm/sec
- Extensive ECG quality control by AC Noise Filtering and Baseline.
- A4 size reports for convenient reading and filing.
- Colour coded keys for ease of operation.
- Convenient battery operation for greater mobility.
- Versatile report formats and speed options to provide auto reports or rhythm reports.
- User configurable filters.

- Preview signal quality prior to printing. Saving time and paper.
- Keyboard entry for patient ID information.
- Capability to generate any number of ECG copies possible for filing and distribution.
- Adult and paediatric analysis programs std.
- Automatic interpretations of ECG data.
- Availability of adult, paediatric and neonatal accessories.
- Accessories e.g., stand, cables, electrodes etc. Should be quoted separately.
- ECG Paper roll for 1000 patients
- Service and operation manual complete.

21. ELECTROLYTE ANALYZER

Fully automated ph/blood gas/electrolytes analyzer, measures the following parameters
ph

pco2 and po2 and barometric pressure, na, k, ca, cl, ct, thb, hcl

Analyses per hour should not be less than 36

Sample volume should not exceed 110 µl for all parameter

All calibration and cleaning cycle should be fully automated with user selectable
calibration

times

The electrodes provide should be zero maintenance including the reference electrode.

The system should have onboard data management to store all patient result and
calibrations.

A power fault protection for 20 minute to keep all calibration and programmed data

The system should have a screen to access all system software and to display the
patient

results.

A built –in printer should be provided to print out patient result

The system should have a patient data, new reagents and new electrodes. Startup kit
1000

samples, essential spares.

The system should work on discrete testing, i.e. Selectable parameter testing.

Regarding stand by mode and use of reagent during non functioning status

Optional use of only Ca⁺ or only Na⁺ & K⁺ to save reagents.

22. Emergency Resuscitation Kit

1. To have Retromolar Intubation fiberscope for unexpected difficult airways.
 - a. Tip Distal Bending 40°.
 - b. To be movable eyepiece
 - c. To have a light source connection
 - d. With length 40-42cms and dia 5-6 cms.
 - e. ET tube holder should be provided

-
- f. Should take min. 5.5 size of ET tube
 - 2. Portable LED light source should be provided
 - i. with illumination not less than 50000 Lux
 - ii. should run on two 3v photo batteries
 - iii. burning life should be more than 100 minutes
 - iv. ergonomically designed and can be connected to both the fibrescopes
 - v. life of LED should be close to 50000 hrs
 - 3. One Laryngoscope with rechargeable battery pack and blade with fibreoptic mechanism should be provided to be used on both adult and pediatric patients with charger.
 - 4. Other accessories like, magill forceps should be provided.
 - 5. Should have Emergency Cricothyroidotomy for pediatric and adult
 - i. disposable blades
 - ii. dialator
 - 6. Should have Combitube size 37Fr.
 - i. with complete kit
 - 7. Should have Intubating Laryngeal Mask Airways with Following Components:
 - a. ILMA Sizes 3 & 4.
 - b. ILMA Tubes ID 7mm & 7.5mm.
 - c. Tube Stabilizing rod
 - d. Cuff deflator
 - 8. Should have Laryngeal Mask Airways
 - i. sizes 1,2 and 4
 - 9. Handy and strong brief case/bag should be provided to keep all the instruments safe.
 - 10. Set of disposable percutaneous tracheotomy kit for adult and pediatric.
 - 11. Should have standard AMBU bag for pediatric and adult.
 - 12. Mechanical suction pump with suction catheter and stomach tubes.
 - 13. Should have Aluminum Oxygen reservoir 2 Liter with oxygen tube and catheter.
 - 14. Oxygen pressure reducer, regulable 0-15 liter with coupler for respirator.
 - 15. Ventilating bag
 - 16. Lubricant
 - 17. Blood pressure meter, bosco K-II
 - 18. Stethoscope
 - 19. Rescue blanket gold/silver
 - 20. Infusion system.

23. Fully automated ELISA reader with washer

1. Description of Function

1.1 ELISA Reader is required to Read the Colour Density known as OD(Optical

Density) in ELISA (Enzyme Linked Immuno-Sorbent Assay.)Plates.

2 Operational Requirements

2.1 Only ELISA Reader is required.

3 Technical Specifications

3.1 OPTICAL SYSTEM

Digital light control

8 measurement channels including 1 reference.

Single and dual wavelength measurement with facility for kinetic measurement

8 s maximum measurement time for single and dual wavelength and 5

s(+/_1Sec.) for kinetic

Measurement Range 400-700nm

Indication Range 0-2.999 abs

Accuracy Plus/Minus 2% or Plus/Minus 0.005 abs

Resolution 0.001 abs

Inbuilt Filters: Narrow band interference

Should have the following filters – 405, 450, 492(+/_2nm), 540, 620 (+/-10nm) and 690 nm

Should measure end point, curves and kinetic.

3.2 SOFTWARE:

Storage of immediately preceding measurement At least 15 user programmable tests permanently stored

Time programmable between each measurement. Agitation programmable before each reading

Bidirectional printer interface.

Data memory through computer

Built in Windows based software programming software.

3.3 MEASUREMENT MODES

Plate shaking mode for sample mixing (selectable speed and time)

Flexible blank mode setting

Matrix Modes: Matrix -/x/t, Matrix-/0-0 (Range),

Matrix-/f/(Floating cut off)

Difference Mode: Absorbance of each well in even numbered subtracted from those of odd numbered columns

Curve fit Modes: LIN/LIN.LIN/LOG.LOG/LOG or auto curve transformation

with ability to add the standard curve; 8 to 12 way string orientation or kinetic modes

Table of optical densities, Delta DD, Graphic, Reaction rate/V-Max

3.4 Adjustable for different micro plate geometrics

3.5 Halogen Lamp 20 - 40 W.

3.6 16 digit alphanumeric fluorescent display

3.7 Membrane keyboard.

3.7 Technical Specifications for washer

3.1 Auto strip washer for 96 well plates / strips

3.2 1 x 8 strips/ 1x12 strips.

3.3 Dispensable wash volume 50 - 300 µl.

3.3a Residual wash Volume <0.5µl

3.4 Aerosol Shield for user safety.

3.5 In built shaking facility

4 System Configuration Accessories, spares and consumables

8-12 channel manifold, all tubing sets, wash, rinse and waste bottles

Maintenance kit to be provided.

4 System Configuration Accessories, spares and consumables

4.1 System as specified-

4.2 Halogen Lamps : 2

4.3 Printer inbuilt or external to be supplied along with 10 Rolls/Z Fold

4.4 Dust cover.

5 Environmental factors

5.1 The unit shall be capable of being stored continuously in ambient temperature of 0 -50deg C and relative humidity of 15-90%

5.2 The unit shall be capable of operating continuously in ambient temperature of 10 - 40deg C and relative humidity of 15-90%

6 Power Supply

6.1 Power input to be 220-240VAC, 50Hz fitted with Indian plug

6.2 Resettable over current breaker shall be fitted for protection

6.3 Suitable voltage corrector/stabilizer

6.4 Suitable UPS with maintenance free batteries for minimum one-hour back-up should be supplied with the system.

7 Standards and Safety

7.1 Comprehensive training for lab staff and support services till familiarity with the system.

7.2 Should be compliant to ISO 13485: Quality systems - Medical devices - Particular

requirements for the application of ISO 9001 applicable to manufacturers and service providers that perform their own design activities.

7.3 Should be FDA or CE or ISI approved product

8 Documentation

8.1 User/Technical/Maintenance manuals to be supplied

8.2 Certificate of calibration and inspection from factory.

8.3 List of Equipments available for providing calibration and routine maintenance

support as per manufacturer documentation in service / technical manual.

8.4 List of important spare parts and accessories with their part number and costing.

8.5 Log book with instruction for daily, weekly, monthly and quarterly maintenance checklist.

The job description of the hospital technician and company service engineer should be clearly spelt out

24. Fully Automated random access biochemistry analyser

Multi parameter and discrete analyzer including electrolyte with computer and printer.

- Capable of analyzing all routine bio-chemistry analysis including electrolytes (ISE – Module)
- Throughput up to 320 test/hour, on 24 hours basis.
- Capability up to 40 samples/run and up to 20 different at any one time
- On board refrigeration for reagents.
- Integrated computer system for data management and storage data.
- Sample volume can be adapted to paediatric samples (micro analysis)
- Calibration and maintenance requirement is minimum.
- To be supplied with 90 Litre/hour multi-stage with 3 different grades of water (3 separate filter) and deionised.
- Tests to be analyzed by pre-analyzer, LFT, liquids, total protein, calcium, phosphorus, magnesium amylase,, sodium, potassium, bicarbonate, chloride urea, creatine and glucose etc.
- Battery backup minimum for 30 mins.

With complete accessories as cuvette, startup kit, consumables, cup 1000 etc.

Stat sample facility.

On-board laundry system.

Minimum 10 filters ranges from 300nm to 700nm next different 20 bio-chemistry reagent can be placed at reagents segment.

True random selection possible with self detection of test.

continue sample loading facility with self detection of test.

25. Holter Recorders and Analyzers

1 It should have FDA clearance and CE certification

2 It should meet and exceed the requirements of ANSI / AAMI EC 38

Holter Recorders

- 1 It should be small and lightweight recorder
- 2 It should be capable of simultaneous real time acquisition of 3 channel and 12 channel recording.
- 3 Recorder should have LCD display to preview ECG waveforms during the patient hook up and have lead quality check function.
- 4 it should have capability of continuous recording for 24 hours and 48 hours / 7 days.
- 5 It should have at least 24 hours with single battery.
- 6 The recorder should be water resistant.
- 7 Recorder should be battery operated and have single AA/AAA battery or rechargeable in built sealed battery with recharging unit supplied.
- 8 It should have internal memory for 99 full disclosure readings.
- 9 it should have advanced signal processing algorithms to provide superior accuracy in beat detection, labeling and noise rejection.
- 10 It should be a 12 channel analyzing system.
- 11 it should have capacity to download 24 hours data through USB port / standard connection port in less than 3 minutes.
- 12 Device should be defibrillator protected.
- 13 Recorder should have 1000s/sec/channel digital sampling rate for standard recording and internal storage.
- 14 Recorder should have compact flash card memory card.
- 15 it should be capable of pacemaker spike detection.
- 16 it is preferable if it supports voice recording capability.
- 17 It should include 2 sets of electrodes and patient cable to enable 3 and 12 channel recording.
- 18 It should come in 2 sets of patient hook up pouch and hook up accessories.
- 19 Should include compact flash card as required for each recorder with adequate capacity to store entire recording cycle.
- 20 It should include cable for connecting and downloading data to PC.
- 21 It should include compatible flash card reader.
- 22 It should include users guide and technical manual as well as quick guide and patient hook up instruction posters for easy reference and use.
- 23 It should include 100 sets of patient event recording diary and pens.
- 24 It should have two years warranty on recorder / memory card and 6 months for cable.

Analysis software + Hardware.

- 1 It should have multiple scanning options like retrospective, prospective and superimposition scanning modes, event and template review and be customizable.
- 2 It should have facility to convert final report into PDF and XML format that enables connectivity to data management systems.
- 3 It should have trend graphs for HR, RR interval, RR variance, 12 lead ST, SVPB, VPB etc and provide graphic display options.
- 4 It should have different beat classifications.
- 5 It should have various rhythm analysis and ST measurements.
- 6 It should have 3 and 12 lead ST segment measurement analysis.
- 7 It should be capable of Pacemaker analysis including atrial,

ventricular and dual chamber pacing and under / over sensing and capture failure.

8 It should support customizable report format including patient data, 24 hour profile, selective printing of rhythm strips and trends, summary statistics in tabular and narrative format and support user defined acronyms for comments.

9 The PC + Printer provided should be Intel Pentium IV 2.5 GHz and above and include minimum hard disk of 250 GB / DVD+RW drive or Blue Ray disc writer / USB port/4GB RAM/10"TFT color monitor, Laser printer. Should have one year on site warranty.

10 It should have capacity to archive into hard disk and DVD+RW / High density DVD / Blue Ray disk.

11 It should include licensed operating software for ready to use status along with reliable anti virus software. These should have back up provided on CD / DVD.

12 It should include external data back up / archiving facility and hardware.

13 All trainer manuals should be supplied and on site installation to complete satisfactions of end user.

14 Training of 2 technicians to operate and use the equipment and software to be provided at the company expense.

15 Assurance of 95% uptime and prompt provision of service back up for hardware and software with replacement to maintain operational equipment as required.

26. Infant Resuscitator

Manometer Range: 20 to 80 cm H₂O (-20 to 7.8 kPa)

Maximum Pressure relief should be 5 to 70 cm h₂O at 8 LPM

Peak inspiratory Pressure (PIP) should be 5 to 70 cm H₂O at 8 LPM

Positive-end expiratory pressure (PEEP) should be 1 to 9 cm H₂O at 8 LPM.

Gas inlet flow range:-

5 LPM (min) to 15 LPM (max) If the gas inlet flow rate increases from 5 to 15 LPM, the peak inspiratory

pressure typically increases approximately 8cm H₂O (0.8 kPa).

Delivered oxygen Concentration: - Up to 100% depending on gas supply

Operating time (400 L cylinder):- @ 8 LPM 50 minutes

Recommended body weight: - Up to 10 kg (22 lb)

Should be supplied with all related tubing's and consumable (Preferable reusable) for at least 50 cases (application)

27. Laryngoscope (Fibreoptic)

Field of View:75 Deg

Depth of Field:3-50 mm
Diopter: +2-8 Dptr
Tip Deflection:up/down 130 Deg
Distal Diameter:4.8mm
Insertion Tube Diameter:4.9mm
Diameter of Instrument Channel:2.2 mm
Working Length:580 mm
Standard Accessories.
Biopsy Forcep with Window-1 No
Cleaning Brush-1 No
Cylinder Cleaning Brush-1 No
Rubber Inlet Seals-10 Nos
O-Ring Set-1 No
Suction Cylinder Cl.Cap-1 No
Eyepiece Cap-1 No
ETO Ventinc Cap- 1 No
Halogen Light Source-1 No
Leakage Tester-1 No

28. MOBILE C-ARM IMAGE INTENSIFIER WITH DSA

Features - Generator

Microprocessor controlled High Frequency generator with 2.5kW or More with integrated beam filters

to reduce patient skin radiation dose.

Collimator: IRIS or multi leaf

X Ray mode (kV & mA range):

kV- range : 40 - 110 kV

Fluoroscopya)

Fluoroscopy should not exceed 5 mA .

b) Pulsed Fluoroscopy with last Image Hold

Radiography –

Radiographic mode for cassette exposures: minimum of 20mA

Image Intensifier:

9"or More Triple Mode Image Intensifier with Hi – resolution CCD Camera

Image Processing:

a) Minimum 12 bit Digital Fluoroscopy Imaging Unit with dedicated video pipe-line processor

b) Archival memory CD/DVD mode.

c) Detachable Cassette holder for film recording.

d) Complete Hi end and latest computer system with required licensed software for image capture,

storage, post process, retrieval, print, transfer and patient data storage.

Image Display:

Two 18" TFT/ LCD High resolution, high contrast and flicker free Monochrome Monitors of at least

1024 X 1024 matrix .

Soft Tissue filters to be provided for better visualisation of soft tissues.

System Functionality:

Vertical ,Horizontal and Orbital Travel should be available

C arm rotation +/- 130 degree or more

The System should be DICOM ready

Accessories:

a)Wrap around light weight vinyl Lead Aprons with 0.5 mm lead equivalence certified by BARC or AERB or ISO : 6 (Six Nos.)

The system should perform DSA with acquisition of 6 frames per second or more, real time

and peak hold , road mapping, annotation, re-masking and multi image display.

Warranty (as specified in the tender document)

1. Comprehensive WARRANTY for complete system including x-ray tubes and all vacuumatic items, II and all accessories.

CMC for complete system including x-ray tubes, all vaccumatic items, II and all accessories.

Acceptance tests as per International Standard should be carried out at manufacturing

facility as well as installation site (including all Safety and QA tests)

29. Multipara Bedside Monitor with Central Monitoring Station

Patient monitor system should be of modular type and capable of monitoring adult, pediatric & neonatal patients.

Monitor should have 17" independent flat panel display.

Touch screen user interface .

Module rack / housing should be independent and shall be able to be placed near to the patient.

Should be capable of 8 traces display.

Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard: ECG, NIBP, SpO₂, invasive pressures (2), temperatures (2)

Should be compatible with Capnography, Cardiac output,4 channel direct EEG, BIS and prices to be offered as optional for each module separately.

ECG should have capability for 3, 5 and / or 10 lead monitoring and should have built in arrhythmia monitoring on all leads.

Inbuilt ST segment analysis and arrhythmia detection for all the leads should be possible.

Haemodynamic and drug dose calculations should be available.

Arrhythmia should be grouped based on classifications – and should show no of arrhythmias occurred.

Respiration should be available with Cardio Vascular Artifact filter.

ICP monitoring should be possible.

Alarm parameter should flash red in the presence of high priority alarms (e.g. ventricular fibrillation and asystole) and flash yellow in the presence of medium or low priority alarms (e.g. noisy signal, etc.)

24 hours trend data should be displayed.

All monitors including central station should have similar user interface for usage among all clinicians.

Monitor shall provide the capability to interact with alarms at remote bedsides.

Monitor shall provide the capability to receive and display real-time waveforms, trended data and alarm status from other bedside or telemetry units on the patient monitoring network.

Monitor shall provide the capability enter patient information at the bedside or central monitor.

On-screen keyboard for entering this data is preferable. Should have USB ports to connect mouse, key board, bar code scanner.

Alarm limit status (ON/OFF) must be indicated on-screen for each parameter and actual parameter alarm settings must be displayed on-screen when alarms are on.

Position of the displayed waveforms must be user configurable.

Waveform color changing should be user configurable.

Monitor shall permit the optional ability to receive and display information from other patient devices such as ventilators, infusion pumps and other standalone devices.

All modules should be compatible with all monitors quoted.

Bed to bed communication between the monitors should be possible with out a central station.

Networking to central station should be possible and price of central station should be offered as optional

Patient monitoring network shall use standard TCP/IP protocol and be capable of residing on hospital's network infra-structure.

Should be compatible with HIS and should be HL7 compliant.

Monitor should provide remote viewing of real time waveforms through internet.

Patient monitoring network shall be able to support up to 1,000 monitoring nodes.

Should be supplied with necessary accessories for adult , pediatric and neonatal accessories.

Accessories and spares

1. ECG / respiration: 5 lead ECG cable and lead wire set per monitor
 2. NIBP: Adult: 2 sizes and Pediatric 2 sizes and neonatal, 1 size per monitor
 3. SPo2 Sensor: Adult sensor with cable, pediatric sensor with cable and neonatal sensor with cable per monitor
 4. IBP: Include 10 nos of disposable pressure transducer with bracket and interface cable per monitor
 5. Temperature: Skin and nasopharyngeal probes per monitor.
- The equipment should be CE & US FDA Approved.

30. NEONATAL MONITOR

Patient monitor system should be of modular type and capable of monitoring neonatal patients.
Monitor should have 17" independent flat panel display.
Touch screen user interface

Module rack / housing should be independent and shall be able to be placed near to the patient.
Should be capable of 8 traces display.

Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard: ECG, NIBP, SpO₂, invasive pressures (4), temperatures (2).

ECG should have capability for 3, 5 and / or 10 lead monitoring and should have built in arrhythmia monitoring on all leads.

Inbuilt ST segment analysis and arrhythmia detection for all the leads should be possible.

Haemodynamic and drug dose calculations should be available.
Arrhythmia should be grouped based on classifications – and should show no of arrhythmias occurred.

Respiration should be available with Cardio Vascular Artifact filter.
OCRG should be available for monitoring neonates.
ICP monitoring should be possible.

Alarm parameter should flash red in the presence of high priority alarms (e.g. ventricular fibrillation and asystole) and flash yellow in the presence of medium or low priority alarms (e.g. noisy signal, etc.)

24 hours trend data should be displayed.

All monitors including central station should have similar user interface for easy usage among all clinicians.

Monitor shall provide the capability to interact with alarms at remote bedsides.

Monitor shall provide the capability to receive and display real-time waveforms, trended data and alarm status from other bedside or telemetry units on the patient monitoring network.

Monitor shall provide the capability enter patient information at the bedside or central monitor.

Alarm limit status (ON/OFF) must be indicated on-screen for each parameter and actual parameter alarm settings must be displayed on-screen when alarms are on.

Position of the displayed waveforms must be user configurable.
Waveform color changing should be user configurable.

Monitor shall permit the optional ability to receive and display information from other patient devices such as ventilators, infusion pumps and other standalone devices.
All modules should be compatible with all monitors quoted.

Bed to bed communication between the monitors should be possible without a central station
Networking to central station should be possible.

Patient monitoring network shall use standard TCP/IP protocol and be capable of residing on hospital's network infra-structure.

Should be compatible with HIS and should be HL7 compliant.

Patient monitoring network shall be able to support up to 1,000 monitoring nodes.

Should have CE and FDA certifications.

Should be upgradeable for EEG, BIS, SvO₂, EtCO₂ and Cardiac Output and price to be quoted separately

Accessories and spares

1. ECG / respiration: 3 lead ECG cable and lead wire set
2. NIBP: Neonatal :2 sizes per monitor
3. SPO₂ Sensor: Neonatal sensor with cable per monitor
4. IBP: Include 10 nos of disposable pressure transducer with bracket and interface cable per monitor
5. Temperature: Skin and nasopharyngeal probes per monitor.

31. Oesophoscope Adult

Field of View: 100 Deg

Depth of Field:3-100 mm

Diopter: +2-8 Dptr

Tip Deflection:up/down -210/120 deg

Tip Deflection:right/left -120/120 deg

Distal Diameter:9.8mm

Insertion Tube Diameter:9.8mm

Diameter of Instrument Channel:2.8 mm

Insertion Working Length:1050mm

Total Length:1395 mm

Standard Accessories.

Biopsy Forcep with Window-1 No

Cleaning Brush-1 No

Cylinder Cleaning Brush-1 No

O-Ring Set for Suction Valve-1 No

Check Valve Set-20 Pcs

A/W Channel Cleaning Adapter-1 No

A/W Suction Channel Cleaning-1

Eyepiece Cap-1 No

Eyeshield-1 No

Bite Block-1 No

Silicon Oil-1 No

ETO Gas Ventinc Cap- 1 No

Halogen Light Source-1 No

Leakage Tester-1 No

32. Operating zoom microscope for ophthalmology

MICROSCOPE:

- Compact microscope body with high quality & complete apochromatic Optics with 1:6 zoom ratio. Magnification factor 0.4X to 2.4X.

- Focusing range 50mm, Objective lens f= 200mm, 65mm diameter.
- Binocular tube: Tiltable tube with integrated image inverter without any external attachment.
- Eyepices: 10X with +8D to-5 D compensator.
- Deepview: Depth of field management system for optimal depth perception & maximum light transmission.

ILLUMINATION :

- Stereo Coaxial Illumination system for unique detail recognition, high contrast & stability of Red reflex even with strongly pigmented decentered and ametropic eye.
- Retina Protection Device and contrast enhancement aperture.
- Integrated 408nm UV barrier filter/ Blue blocking filter/ fluorescence filter.

X Y COUPLING:

- Motorized foot controlled X-Y coupling with automatic re-centering and X-Y inversion facility. X-Y Range should be at least 60mm x 60mm adjustable range.
- Stereo co observation attachment with 360 Rotation -2 joints .

SUSPENSION SYSTEM:

- 14 function wireless foot control, Motorized foot controlled Zoom and focus with re-centering of focussing position through foot control. Image inversion facility on foot control.
- High quality floor stands with long spring balance suspension arm with effective length of 1Metre or more having load bearing capacity of atleast 14Kg or More.
- Stand should have touch screen LCD display with programming facility for setting the speed of XY, Zoom and focus, Foot Pedal.
- Stand should have cold light fiber Optic illumination 12v 100w Halogen lamp with in built lamp housing with two lamps, with automatic Lamp changeover facility.

CCTV ATTACHMENT:

- 1CCD Camera with camera control unit, control unit should be integrated in the in the floor stand. Video o/p S video & analog through the stand & programming through LCD display in the floor stand.

NETWORKING:

- Ethernet interface for microscope i incl. 10 m cable

WIDE ANGLED VIEWING SYSTEM:

- Wide angled Non Contact observation/ viewing system (autoclave able) with field of viewing 120deg.(minimum). With independent focussing

33. Patient Warming System

1. The convective air patient warming system should have a basic warming unit and disposable blankets.
2. The convective air patient warming system should have fast warming reaching 38 °C with in 30 sec.

3. The warming system should have temperature range settings of 30 °C to 34 °C, 36 °C to 40 °C and 42 °C to 46 °C.
 4. The warming system should have an automatic step down facility. After 45 min temperature will come down from high mode to medium mode.
 5. Should be CE / FDA certified
 6. Should have Hepa filter of 0.05 micron filtration efficacy.
 7. Multiple mounting options: Cart, Bedrail, IV Pole and floor.
 8. Machine should come with the stainless steel movable trolley for mobile purpose.
 9. Machine should have auto power cut facility to control the set pressure and sensors to prevention patient burn.
- Machine should have hour meter to understand total run time.

Blankets

1. The blankets for convective air patient warming system should be compatible with the basic warming unit.
 2. The blankets should be lighter and resistance to puncture and fluids.
 3. The Blankets should latex free, made of 2 ply material – non -woven outer layer and polyethylene inner layer. They should be precision dye cut to have an even airflow and smooth surface.
- Blanket sizes – Adult upper body, Adult lower body, Adult Full Body, Cardiac Blanket, Pediatric blanket.(10 Each)

34. Phaco Emulsification Machine

Graphic User Interface based on 8,4" color LCD and touch screen

Voice feedback for function selections-English.

Dual Linear footswitch, selection among 6 pre-programmed modes

surgeon can store up to 10 user programs

I / A :

Vacuum level range programmable from 5 to 500mmHg (step 5mmHg) Closed system vacuum reading,(reading through a sterile silicone membrane), Reusable tubing, Reusable tubings are steam autoclavable up to 50 times or more

Irrigating pressure regulated by height of I V pole, Continuous irrigation-controlled by footpedal and key on touch screen.

Flow rate range 25 levels, programmable from 2 to 50cc/min (step 2cc/min)

Rise time programmable on the following 25 levels: 0.5, 0.52, 0.55, 0.57, 0.6, 0.63, 0.65, 0.70, 0.75, 0.77, 0.85, 0.90, 0.95, 1.05, 1.15, 1.25, 1.4, 1.5, 1.8, 2.1, 2.5, 3.0, 4, 6, 12 sec.

Panel or linear Vacuum control by the foot pedal

Ultrasound:

Operating frequency approx 40 Khz

Stroke range 5 to 100 micron, step of 5 micron controlled by "Adaptive Power Control", panel or Linear stroke control by the foot pedal

Operating modes : Continuous, Pulsed from 1 to 40 Hz, Single Burst, Multiple Burst, Continuous Burst- Available Pulse rates : 1,2,3,4,5,6,7,8,10,13,16,20,25,32,40Hz

Handpieces : 4 piezoelectric hand piece with 4 / 40 UNC thread for the U/S tip and 1/4-32 UNEF for the sleeve. (four crystals, titanium, ultra light) Handpiece natural titanium color, steam autoclavable up to 600 times or more, dishwasher-safe

U/S Tips: Coaxial Phaco with 4/40 UNC thread Incision size range from 2.2 mm upto 3.2 mm, Coaxial Phaco, Co-MICS, MMICS

19 GA, Color : gray (natural titanium),20 GA, Color : Blue, Co-MICS with 4/40 UNC thread

SLEEVES RANGE :

Silicone sleeve for 19 GA tips, Color : gray, thread 1/4-32 UNEF, Silicone sleeve for 20 GA tips, Color: blue, thread 1/4-32 UNEF

Vitrectomy:

Cutters reusable guillotine- Steam autoclavable. Adjustable port from 0.2mm up to 0.7mm, Pneumatic cutter Cutting range from 60 to 700 cut / min. Available cut rates : 60, 70, 80, 90, 100, 110, 130, 150, 170, 190, 220, 250, 280, 320, 370, 420, 480, 540, 620, 700 cut/min

Compressed air from integrated air compressor, Operating pressure 2.0 +0/-0.1 bars (28 +0/-1.5 PSI)

Panel or Linear cut rate control by footpedal

Diathermy:

Type : Bipolar Max power 7W @ 450hm, Power adjustment 5 to 100 % step 5, Operating frequency 2Mhz, Panel or Linear power control by foot pedal

Reusable diathermy forceps / Reusable diathermy pencil / Diathermy Bipolar Cable- steam autoclavable up to 50 times

Input voltage: 100/120/220/230-240 V(A.C.) selected, Mains frequency 50/60 Hz

35. Pharmaceutical Refrigerator for O.T.

Good quality double steel walled cabinet
Capacity 500 litre

Forced air circulation to maintain chamber temp. uniformity.

Temp. in range of 2 to 8 degree centigrade

Low noise level

Automatic defrosting & digital temp.display

Audiovisual alarm for low & high temp.

Input voltage 220 to 240 volts

Inbuilt electronic voltage stabilizer

36. Portable Ventilator

The Ventilator should be portable.

The ventilator should be capable of ventilating from pediatric patients to adults.

The modes should include Assist Control, SIMV in Volume control and Pressure control modes.

It should also have CPAP and PSV available.

The tidal volume should range from 50ml to 2000ml.

Peak Inspiratory Pressure - 5 to 55 mbar

Pressure Support – 5 to 55 mbar

Insp. Time – 0.3 to 2.4 secs

Breath Rate – 0-60 BPM

Insp. Sensitivity – 1 to 5

Exp. Sensitivity – 5 – 95%

Flow Pattern should be Square, descending and Sinusoidal

PEEP should be 0.5mbar to 20mbar

Rise Time should be from 1 to 4 with increments of 1.

I:E ratio should be 1:4 to 1:1

I / T should be 20% to 50%

The Apnea time should be 1- 60 secs and backup rate of 5 to 40 BPM.

Vt SIGH should be in single to double multiplier of Vt.

The ventilator should have bright display with backlight

The ventilator should have Waveform display of Pressure Vs. Time and Flow Vs. Time.

Ventilator should be able to be used invasively through an artificial airway, or noninvasively through a mask or other noninvasive interface.

The ventilator should have choice to select the type of breathing circuit (Paed. OR Adult) for Circuit volume compensation.

The ventilator should have internal battery back up for at least 11 hours with Real-time battery life indicator

The ventilator noise level should be < 30 dBA at 1 m.

The ventilator should have Automatic atmospheric adjustment

The ventilator should have recording facility of patient data trends through USB for 3 months to 1 year time with following parameters;

Monitoring: pressure, inspired flow, exhaled flow and leak measurement

Trends: leaks, VTI, VTE, Rate, I/T, M. Vol, P MAX and PEEP measurements

37. Portable X-ray unit (Mobile)

2.5 kw HF General purpose, single tank diagnostic mobile

X-ray Equipment, suitable for radiography of standing, sitting or recumbent patients as well as for patients in bed or operation table & as stand by unit.

Comparising of High Frequency X-ray generator

2.5 with the following out put-
 Output: 2.5 kw as per Is 7620
 Kv Range : 40 kv-100kv in 24 steps
 mA Range :16mA-60mA, mAs Range : 0.32mA's-200mAs
 Exposure Time : Min 20 m sec

The unit will be calibrated within the permissible limit of x-ray tube

X -RAY TUBE

rotating anode x-ray tube having 1.4*1.4mm focal spot

Collimato:-

Brightfield light beam collimator with 100w halogen lamp and auto shut-off facility

Power Supply:-

Single phase 195v-265v, 50Hz power supply with line resistance<0.8 ohms

Mobile Stand

Light weight,low height mobile stand with easy manoeuvrability
 AERB/ BARC approved

38. Pulse Oximeter

- It monitors from -neonate to adult.
- To track oxygen saturation even during low perfusion condition.
- Set seconds alarm management technology avoids unnecessary alarm & distraction during monitoring of severe motion patients.
- Display for better monitoring.
- Dual colour LED for differentiating normal mode & alarm mode.
- Internal battery provides minimum of 8 hours backup.
- 10 segment LED bar graph for pulse strength monitoring.
- Facility for computer interface & printer connectivity
- Individual alarm settings for Sao2 & pulse rate
- Compatible with the complete range of disposable & reusable sensors.
- Built-in battery charger with charger indication.

39. SLIT LAMP

Magnifications : 8x, 12x, 20x
 Field of view : 45 mm – 10 mm(or better)
 Eyepiece magnification : 10x high-eyepoint eyepieces, \pm 8D compensation of ametropia
 Width of slit image : 0 – 14 mm, continuously adjustable(or better)
 Length of slit image : in steps: 0.5 / 3.5 / 8 / 14
 1 – 14 mm, continuously adjustable
 Angle of slit image : \pm 90°, continuous
 Decentration of slit image : variable, click stop at 0°

Swivel range of slit prism : 180°, scale for angular difference, Click stop at 0°

Angle of incidence : 0°, horizontal

Filters : blue, green (red-free), and diffusing screen, Swing-in-type; UV protection filter, Heat-absorbing filter.

Free working distance : 73mm

Travel of instrument base : vertical: 30mm, X-axis: 110 mm, Y-axis: 90mm

Vertical travel of chin rest : 58 mm

Light Source : 6V, 10W Halogen Lamp, continuously adjustable brightness

Optional Accessories

1. Applination Tonometer
2. Motorised Table

40. Suction machine (portable)

Noiseless suction unit should have fast vacuum build up

Suction capacity should not be less than 50 ltr/ min.

Vacuum should not be less then -90 kPa/-675 (mmHg)

The unit should have vacuum display on the machine

Fitted on mobile stand with On/Off facility.

Mechanical overflow protection system.

Suction system should have piston/cylinder.

Twin Bottle capacity 3 Ltr. (unbreakable) with switchover facility.

The manufacturer should have CE & FDA approval Certificate.

41. Syringe Pump

Should be able to adopt syringes from 5 to 60ml horizontal

Programming : Should be ml/hr, ml/hr x volume limit & time x volume limit

Flow rate should be 450 ml/hr with 0.1ml/hr increment (independent of syringe capacity for all syringes)

Volume Limit should be 999.9 ml

Must have variable Occlusion Pressure at least 3 levels

KVO should be 1ml/hr or less

Bolus rate should be 450 ml/hr , minimum 1 ml

Volume x Time infusion mode limit : 0.1 to 999.9 ml and time 1 min to 99 hrs and 59 min.

Should have Check facility to recall programmed parameters

Special Functions Like : Clear Volume

Should have pre alarms such as End of Battery, End of Infusion, & visual & sound alarms like

Syringe positioning error, occlusion, infusion completed, syringe error, programming error,

Block error, clamp error , low battery etc

Flow rate deviation should not be more than +/- 1% of the programmed flow rate.

Battery Rechargeable with battery life 3hrs , Should be able to work on 230 Volts

Should have clear LED display.Syringe placement should be top loading.

Sholuld have check facility to recall programme parameter.

The product should have GMP, CE & ISO 13485 certifications

42. TMT

Stress testing system should be complete with PC, Software, Tread Mill, Patient acquisition module and necessary cables.

System should be based on windows platform with 17" color monitor having minimum resolution of 1280 x 1024, 80 GB HDD, CD-RW, Mouse, UPS for analyser.

System should acquire and analyze 12 leads and should store the full 12 lead resting / stress ECG

Should show a pictogram and indicate which electrode has a bad contact before resting / Stress recording.

Should be able to review retrospective average beat and ST values.

Should have user settable protocols.

Patient module should have USB connectivity.

System should provide standard Full interpretation of Supine ECG with reasoning.

Should provide display of real time 12 lead diagnostic quality ECG waveform, average complexes beat of all 12 leads with superimposed comparison along with ST level & slope and ST trend graph.

Automatic detecton, display, storage and review of arrhythmia, heart rate. Should have running trends of ST available on screen during the test process.

Should have ability of comprehensive auto-measurement package including RR intervals, P-wave duration, PQ interval and QRS width.

System should have high quality filters for muscle and baseline noise without influence on the ST results.

Should have alarm levels for ST, HR and Blood Pressure.

Should have keyboard short cuts.

Capability to pause ECG screen view to find a past ECG event without loosing track of current real time ECG.

Capability to insert event marks during stress.

Capability to change ST points during stress.

Should have manual measurement cursors for P-wave, QRS and T-wave. Should allow to measure width (P, QRS, T-wave) and intervals (PR, QT etc.,)

Should have alarm levels for ST, HR and Blood Pressure.

Should be able to export QRS intervals from 12 lead Rhythm ECG, Stress / Resting ECG report in word / RTF format and shuld be able to export raw ECG data.

Printing:

Should be able to print strips during stress or after.

Strips to be printed in 3 channel,s, 2x6 channels, 12 channels.

Print reports should be with summary table, event markers and stage information.

Should have average beat report with ST measurements.

Should have Trend graphs

System should provide multiple and customizable printing formats as per users choice on A4 size normal plain printing paper.

Heavy duty treadmill - imported. Noise free with speed ranging from 0.8 to 19 kph and have an elevation range of 0 to 25%.

Treadmill should provide smooth and safe operation.

Should have automatic belt alignment.

Treadmill should have 175 kgs weight(approx.) capacity with all metal chasis.

Should have emergency stop button.

Other Technical specs:

ECG sampling rate: 1000 Hz
ECG input voltage: 16mm Vpp
Noise voltage: 20uV
10 lead patient cable
Treadmill interface
Signal frequency range: 0.05Hz to 250 Hz.
Windows operating system.
PC Pentium latest workstation.
All consumables required for installation and standardisation of system to be part of the system.
Should have US FDA and CE certifications

43. Traction Unit

Should Have Continuous/Intermittent, harmonized/ Intermittent Progressive modes
Traction force should be adjustable from 2.5 Kgs to 91 Kgs.
Should have computerized automatic compensation.
Should Have Timer 1 to 99 minutes.
Speed of progressive modes. Pull should be adjustable.
Should have the option for wall mounting or fixation on electrical height adjustment Traction table.
Should have the possibility of applying cervical traction in sitting position on Traction table.
Couch:
3 section electrically height adjustable couch.
Head section adjustable from -750 to +600.
Leg section adjustable from 00 to +750.
Head section with a breathe hole and plug.
Rolling lumber section (lockable).
Castors are standard.
Patient safety switch.

44. UVB Chamber

The unit has 24 NBUVB Narrow Band UVB Tubes. (311nm)
Tubes from Phillips Holland 100W, 6Ft.
Special UV chokes for maximum life for the tubes.

The Unit is provided with imported mirror (from Italy) type reflectors.
Cabinet made of high quality Steel and powder coated.

In-built Multi Sensor dosimeter with Dynamic Range Angular Sensitivity (DRAS) Technology.

Cumulative hour meter is provided in embedded system with password.
Four line Liquid Crystal Display for the control panel.
Feather touch key pad provided.
Lock and Key provided for the control panel.

Personalized phototherapy patient data card is provided (100 nos.)

One year warranty against any manufacturing defects except for tubes.

Computer calculated arrangements of the tubes, enables an even illumination of the body in the treatment field.

All safety features provided with trippers and independent control for each panel.

Small space required (elegant & compact).

Large door, very easy to open from inside.

Plenty of space in the cabin for the patient.

Cooling system provided for each panel.

Easy to service and assemble (modular design).

The unit is provide with talking system.

Personalized Phototherapy Patient Data Cards Provide with the units.

Minimal heat development due to low current consumption.

Power Required 6KVA.

45. Vacuum Extractor

1. Noiseless suction unit have fast vacuum built up
2. Vacuum should have maximum -90 k pa/-675 mm Hg, Suction capacity 50ltr/min/1kg/cm² & bottle capacity 3ltrs
3. Suction system should have piston / cylinder (self lubricating)
4. Should have mechanical overflow protection system .
5. Set of silicon cups 50mm & 60mm-2each
6. Set of bird cups, stainless steel 40mm, 50mm & 60mm-2each
7. Machine can be operated on 220-240 V AC single phase
8. Should be provided with a gauge to display of vacuum generated
9. Should have vacuum release valve.

46. Vaginal Hysterectomy Set

B.P.Handle No.3	1
B.P.Handle No.4	1
Diss. Forceps Plain 16cm	1
Diss. Forceps Plain 20cm	1
Diss. Forceps Tooth 20cm fine	1
Diss. Forceps Tooth 20cm	1
Diss. Forceps Tooth 16cm fine	1
Diss. Forceps Attri. 20Cm	1
Diss. Forceps Fine Plain 20cm	1
Artery Forceps St.16cm	2
Artery Forceps Curved.16cm	6

Artery Forceps Mosquito curved	5
Artery Forceps Mosquito curved	5
Artery Forceps Mosquito St.	2
Heaany's Clamp	6
Valsalum Forceps Curved	2
Allis Forceps 20cm	6
Allis Forceps 16cm	6
Babcock Forceps 20cm	2
Babcock Forceps 16cm	2
Tanaculum Forceps 25cm	3
Artery Forceps Long Curved 23cm	2
Kochers Clamp 20cm St.	2
Kochers Clamp 20cm Cu.	4
Hysterectomy Clamp 23cm	6
Needle Holder 20cm	1
Needle Holder 20cm	2
Needle Holder 15cm	1
Needle Holder 16cm	1
Scissors Mayo St. 15cm	1
Scissors Mayo St. 18cm	1
Scissor Mayo Curved 17cm	1
Scissor Mayo Curved 18cm	1
Scissor Mayo Curved 20cm	1
Sponge Holder 10" St.	2
Sponge Holder 10" Cd.	2
Avard Vaginal Spaculam Handle	1
Avard Vaginal Spaculam Weight	1
Avard Vaginal Spaculam Blade	1
Avard Vaginal Spaculam Blade	1
Sim's Spaculam	1
Sim's Spaculam Blade Guide	1
Sim Speculam Medium	1
Sim Speculam Large	1
Uterus Holding Forceps	1
Anterior Vaginal Wall Retractor	1
Jackson Retractor	2
Suction Cannula Trizer No.4	1
Towel Clip	6
S S Bowls 10 cm	2
S S Kidney tray 12"	1

47. Volumetric Infusion Pump

Flow rate should be from 0.1 to 990 ml/hr with 0.1ml/hr increment

It should have choice of Micro & Macro drops

Volume Limit should be more than 9900ml

It should use conventional tubing (4 mm), PVC any Indian Make

Immediate display of choice of programming after purge

Must have function for immediate forward & reverse of screen/programming parameters

Programming :

ml/hr,

time x volume limit.

Weight x concentration x dose

Must have Occlusion Pressure from 0 to 120 kPa

KVO should be 0.1 ml to 3ml/hr

Bolus rate should be more than 900 ml/hr

Must have Special functions like Titration , Fluid Balance

Time limit from 1 min to 99hrs & 59 minutes.

Should have pre alarms – Ultrasonic sensor type

Visual / sound alarms like End of Battery Power , Incorrect flow rate, Door open, No drops, Occlusion , programming error, Running in KVO, Fluid balance etc.

Flow rate deviation with standard tubing should not be more than +/- 6% of the programmed flow rate.

Battery Rechargeable more than 5 hrs of operation

Should be able to work on 220 Volts

The Unit should be GMP, CE & ISO 13485 certified.

48. High Frequency 800 mA X- Ray Machine

1. GENERATOR :

Radiographic Parameters

Maximum mA Output :800mA

Maximum kV Output :125kV

Fluoroscopic Parameters

Maximum mA Output : 3mA

Maximum kV Output :90 kV

Exposure Time

From 5 Millisecond to 5 Seconds with digital mAs integrator

The generator should be provided with Rapid Anode Braking Device (to increase tube life)

Automatic computation & display mAs value & kV digitally to be provided .

2. X-RAY TUBE :

The unit should be complete with : 2 Nos. of 30/50kW x-ray tube with high speed anode rotation of higher than 8500 rpm each with a pair of H.T Cable for under couch and over coach radiography.

3. TABLE:

Vertical to Trendelenberg position

Bucky with adjustable cassette tray with a grid ratio of 10:1 and 100 lines per inch

Motorized Under Couch Collimator and Manual Over Couch collimator.

The Spot Film Device should include the following :

- 4 in 1 on 8"X10" Film
- 2 in 1 on 10"X12" Film
- 1 in 1 on 14"X14" Film
along with a suitable grid

Foot Rest, Compression Device foot Switch hand Grip, Skull Cone, Shoulder Rest, Strap Radiation flaps

4. COLUMN STAND :

- Floor to Ceiling with Counter Balancing System.

5. IITV

The entire system should consist of the following :

- Image Intensifier (II) capable of being coupled with the existing 500mA X-Ray machine and fluoroscopic tables.
- High resolution CCD camera to capture the image from the output phosphor of the II.
- Trolley mounted 17" Monitor (44")

Minimum specifications of the Image Intensifier :

- Nominal entrance field diameter of 23 cm with dual field selection capability.
- Output window size 20-25 mm.
- Vertical & Horizontal image orientatino reversal switches mounted on II.
- The assembly should have `All metal' construction and provide magnetic and lead shielding.

Minimum specifications of the CCD camera.

- Should be of PAL Systems with minimum 750 x 580 pixels.
- Video Standard 50 Hz. 625 lines interlaced.
- Video gain should have automatic gain control.
- Circular blanking facility.
- Video output – 1 Vpp composite video.
- Last image hold facility must be present.

Minimum specification of the monitor.

- Minimum 43 cm. 17" diagonal with circular mask.
- Local controls for image contrast and brightness adjustment.
- Should be cart/ trolley mounted and movable anywhere in the fluoroscopy room.

The Image Intensifier should be fully counter balanced consisting of ceiling counterpoise system with rails.

Accessories :

- Voltage stabilizer servo controlled type of 75 KVA capacity.
- Lead Aprons – 4 nos.
AERB/ BARC approved.

Item No 49**O T Tables for General Surgery & Gynaecology**

- Should be Mobile, modular operating table for any surgery, electro-hydraulic drive, with corded hand control, with upper back plate, with leg plates, with head rest, with SFC padding of 80 mm. Compatible with c-arm image intensifier.

Should have following features:

- Battery (1 week/ 60-70 operations on Single Charge) and mains operation.
- Stable base construction with 4 double swivel castors for easy motion and manoeuvring (locking of the double swivel castors via a foot pedal)
- Base cover and cover for the override control panel made of CrNi steel, resistant to impact, breakage and disinfectants
- SFC padding (Special Foam Core) with multi-layer construction for exceptional comfort during patient positioning and utmost patient safety
- Safe patient positioning due to optimum covering of table top joints
- Column casing made of CrNi steel
- Supporting bars for the seat and back section, leg plate sockets, joint covers and side rails made of CrNi steel
- Operating table top subdivided into 5 sections: head rest, upper back plate, lower back plate, seat plate & leg plates
- If there are any malfunctions or if the corded hand control is defective, the operating table can be controlled using the override panel. Entire table top without crossbars, allowing intra-operative fluoroscopy
- Operating elements: corded hand control, override panel, foot switch (Optional)

Should meet following Electrical Standards:

- Special-design, maintenance-free batteries with a capacity for about a week's use in the operating room
 - Battery charge level: electronic monitoring with optical indicator
- Recharging of the batteries and supply of the operating table by means of a mains cord
- Nominal mains voltage /220/230-240 V frequency 50/60 Hz
 - IEC 60601-1

Should meet following Technical data:

- Length of table top : 2075 mm
- Width of seat/lower back plate without side rails : 510 mm
- Width of seat/lower back plate with side rails: 560 mm
- Max. Patient Weight (with Accessories)

Should have following Electro hydraulic adjustments:

Via corded hand control or override control panel, with status indicators for battery charge level and base brake feature, located at the column head.

Adjustment options via corded hand control or IR remote control:

- Height with padding: 600 – 950 mm
- Trendelenburg/reverse Trendelenburg: + 25° / - 25°

- Lateral Tilt :15°
- Lower back plate: 60° / 50°
- Leg plates: 0° / 95 °
- Longitudinal Shift: 290 mm
- Body Elevator: 100 mm
- 0-position (horizontal alignment of the table top)

Should have following Manual adjustments:

- Standard back plate : + 90° / - 45°
- Standard head rest : + 45° / - 45°
- Base braking (lock / unlocking)

Should be supplied with Basic equipment:

- Corded hand control
- Head rest, inclinable, SFC padding
- Retrofit kit, side rails, for head rest for subsequent lateral mounting
- Arm Boards pair
- Anesthesia Screen
 - Clamps or Adaptors
 - Leg plates with SFC padding
 - Gopel knee crutch with clamp pair
 - Side supports pair

Item No 50

O T Tables for Orthopaedics with 'C' arm Compatibility

Electro-hydraulic battery operated mobile operating and extension table for orthopaedic interventions including open and intra-medullary nailing and shall be fully compatible with C-Arm Image Intensifier. Should include back plate adjustment, with SFC padding, including. basic equipment - 2 traction bars short and long, 1 rotation-tilt clamp, 2 elongation bars, 1 rotation and traction stirrup clamp, 1 screw tension device, 1 counter traction post for neck of femur, 1 foot plate support, 1 side rail elongation.

Should have the following features:

- Battery-powered or mains operated electrically controlled hydraulic drive.
- Base covers and column casing, table top frame, traction bars and basic accessories made of CrNi steel.

Operating table with easy-rolling castors for longitudinal and lateral movement.

Operating table top subdivided into 9 sections: head rest , 3-section back plate with 2 detachable shoulder segments, seat plate with 2 detachable buttock supports, pair of leg plates, detachable.

- 2 traction bars beneath the seat plate with 2 separately adjustable pivot joints each for extension of the extremities with unimpeded use of the image intensifier.
- Guide rails beneath the seat plate allowing X-ray cassette insertion from either the right- or the left-hand side for the hip area.
- Detachable pads made of special foam core, 80 mm thick, radio translucent and electrically conductive on plastic supporting plates.

Should meet the following Electrical standards:

Special-design batteries with a capacity for approx. 2 weeks (50 - 80 operations).

- Battery charge level: electronic monitoring with optical indicator and acoustic signal
- Recharging the batteries via mains supply 100 - 240 V AC (adjustable),
- 50-60Hz, via mains cable
- Class II, Type B; the enclosure leakage current meets the requirements of the patient leakage current for CF conditions according to EN 60601-1.

Should meet following Technical data:

- Length of table top incl. head rest : 1965 mm
(With standard back plate)
- Width of seat / back plate without side rails : 530 mm

Permissible load approx : 180kg

Should have following Electro hydraulic adjustments:

Control of the electro hydraulic motions from outside the sterile area via cordless IR remote control optional hand control.

- Height without padding : 800 – 1175 mm
- Trendelenburg / reverse Trendelenburg : + 25° / - 30°
- Lateral tilt : 20°
- Back plate: + 60 °
- 0-position (horizontal alignment of the table top)

Should supply the following with Basic table:

- IR remote control comprising: cordless, battery-powered hand control with film keypad and mobile charging station (to recharge the battery of the hand control) with mains cable; wall mounting should be possible.
- Cable-connected hand control, detachable, with film keypad; for connection to the operating table via flexible coiled cable
- Head rest
- Foot plates, width adjustable, with Velcro strap and heel pad, pair
- Arm board
- Anaesthesia screen
- All necessary clamps or adaptors
- Accessory stand

Should supply the following with Extension accessories:

FEMUR NAILING IN SUPINE POSITION:

- a) Counter traction post for femur: 1 No.

FOR TIBIA & FIBULA NAILING

- a) Counter traction post for tibia : 1 No.
b) Guide bar : 1 No.
c) Condoyle fixation : 1 No.

FOR TOTAL HIP END PROSTHESIS

- a) Fixtures for body support : 2 Nos.
b) Back- buttock support with pad : 1 No
c) Lateral support with pad : 1 No.

KNEE REPLACEMENTS

- a) Motorised Knee positioning device
including radial clamp : 1 No

SPINE IN KNEE ELBOW POSITION

- a) Positioning device for spinal surgery, height adjustable and horizontal adjustment : 1 No.
b) Trolley : 1 No.

- c) Sitting bracket with lateral supports : 1 No
- d) Radial setting clamps : 2 Nos.

CERVICAL SPINE SURGERIES

- a) Connection Fixture : 1 No
- b) Horse shoe shaped head rest (two pc.) : 1 No.
- c) Guiding roller for head side traction : 1 No.

INTRAMEDULLARY NAILING OF HUMERUS

- a) Humerus positioning device : 1 No.
- b) Countertraction post for humerous : 1 No
- c) Weinberger hand traction device : 1 No.

OPERATIONS ON THE HAND.

- a) Hand operating table : 1 No.
- b) Radial setting clamp. : 1 No.

GENERAL TECHNICAL SPECIFICATIONS

GENERAL POINTS:

1. Warranty:

- a) Comprehensive Warranty as stated in GCC clause No. 15 (in Section – IV) for complete equipment (including X ray tubes, HT Cable, Probes, Electrodes, Detectors, Batteries for UPS, other vacuumatic parts wherever applicable) and Turnkey Work from the date of satisfactory installation, commissioning, trial run & handing over of equipment to Hospital/ Dispensaries/ Institute/ Medical College.
- b) 98% up time Warranty of complete equipment with extension of Warranty period by double the downtime period on 24 (hrs) X 7 (days) X 365 (days) basis.
- c) All software updates should be provided free of cost during Warranty period.

2. After Sales Service:

- a) After sales service centre should be available at the city of Hospital/Institution/Medical College on 24 (hrs) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 8 hrs. The service should be provided directly by Tenderer/Indian Agent. Undertaking by the Principals that the spares for the equipment shall be available for at least 10 years from the date of supply.
- b) The service provider should have the necessary equipments recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual.

3. Training:

On Site training to Doctors/ Technicians/ staff is to be provided by Principal/ Indian Agents (if they have the requisite know-how) for operation and maintenance of the equipment to the satisfaction of the consignee.

4. Annual Comprehensive Maintenance Contract (CMC) of subject equipment with Turnkey:

- a) **The cost of Annual Comprehensive Maintenance Contract (CMC)** which includes preventive maintenance including testing & calibration as per technical/ service /operational manual of the manufacturer, labour and spares, after satisfactory completion of Warranty period **to be quoted for next 5 years on yearly basis** for complete equipment (including X ray tubes, HT Cable, Probes, Electrodes, Detectors, Batteries for UPS, other vacuumatic parts wherever applicable) and Turnkey (if any). The supplier shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, but at least once in six months during the CMC period
- b) The cost of CMC to be quoted along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
- c) Cost of CMC will be added for Ranking/Evaluation purpose.
- d) The payment of CMC will be made on six monthly basis after satisfactory completion of said period, duly certified by end user on receipt of bank guarantee for 2.5 % of the cost of the equipment as per Section XV valid till 2 months after expiry of entire CMC period.
- e) There will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- f) During CMC period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.
- g) All software updates should be provided free of cost during CMC.

- h) Failure of the above [4. e) to 4. g)] by the supplier, may lead to the forfeiture of the Bank Guarantee for Annual CMC.
- i) The payment of CMC will be made as stipulated in GCC Clause 21.

5. **Turnkey:**

Turnkey is indicated in the technical specification of the respective items, wherever required. The Tenderer shall examine the existing site where the equipment is to be installed, in consultation with HOD of Hospital/Institution/Medical College concerned. Turnkey details of each Hospital/Institution/Medical College are given at the end of Technical Specification. The Tenderers to quote prices indicating break-up of prices of the Machine and Turnkey Job of each Hospital/Institution/Medical College. **The Turnkey costs may be quoted in Indian Rupee will be added for Ranking Purpose.**

The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

The Turnkey Work should completely comply with AERB requirement, if any.

Section – VIII

Quality Control Requirements

(Proforma for equipment and quality control employed by the manufacturer(s))

Tender Reference No.

Date of opening

Time

Name and address of the Tenderer:

Note: All the following details shall relate to the manufacturer(s) for the goods quoted for.

- 01 Name of the manufacturer
 - a. full postal address
 - b. full address of the premises
 - c. Email ID
 - d. telephone number
 - e. fax number
- 02 Plant and machinery details
- 03 Manufacturing process details
- 04 Monthly (single shift) production capacity of goods quoted for
 - a. normal
 - b. maximum
- 05 Total annual turn-over (value in Rupees)
- 06 Quality control arrangement details
 - a. for incoming materials and bought-out components
 - b. for process control
 - c. for final product evaluation
- 07 Test certificate held
 - a. type test
 - b. BIS/ISO certification
 - c. any other
- 08 Details of staff
 - a. technical
 - b. skilled
 - c. unskilled

Signature and seal of the Tenderer

Section – IX

Qualification Criteria

01. The Tenderer must be a Manufacturer or its authorized Agent.
02. (a) The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, atleast 100% of the quoted quantity of the similar equipment meeting major specification parameters which is functioning satisfactorily. The foreign Manufacturer satisfying the above criteria should also have supplied and installed in last **Five** years from the date of Tender Opening, at least 50% (or one No. where the schedule of requirement is one no.) of quoted quantity of similar model which is functioning satisfactorily any where outside the country of manufacture.
02. (b) The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria 02 (a) should have supplied and installed in last **Five** years from the date of Tender Opening, atleast 50% (or one No. where the schedule of requirement is one no.) of the quoted quantity of similar equipment which is functioning satisfactorily, any where in India of the same manufacturer.

Note:

1. In support of 2 (a) & 2 (b), the Tenderer shall furnish Performance statement in the enclosed Proforma 'A'.

The manufacturer as well as the Tenderer/ Indian Agent shall furnish Satisfactory Performance Certificate in respect of above, duly translated in English and duly notarized, alongwith the tender. The performance certificate should be in accordance with the requirement of clause 2 (a) / 2 (b) as stated above.

2. The Tenderer shall furnish a brief write-up, packed with adequate data explaining and establishing his available capacity/capability (both technical and financial) to perform the Contract (if awarded) within the stipulated time period, after meeting all its current/present commitments. The Tenderer shall also furnish details of Equipment and Quality Control in the enclosed Section VIII.
3. Notwithstanding anything stated above, the Purchaser reserves the right to assess the Tenderer's capability and capacity to perform the contract satisfactorily before deciding on award of Contract, should circumstances warrant such an assessment in the overall interest of the Purchaser.
4. The Purchaser reserves the right to ask for a free demonstration of the quoted equipment at a pre determined place acceptable to the purchaser for technical acceptability as per the tender specifications, before the opening of the Price Tender.

PROFORMA 'A'
PROFORMA FOR PERFORMANCE STATEMENT

(For the period of last five years)

Tender Reference No.: _____

Date & Time of opening: _____

Name and address of the Tenderer: _____

Name and address of the manufacturer: _____

Order placed by (full address of Purchaser/Consignee)	Order number and date	Description of ordered goods and services	Quantity of ordered goods	Value of order (Rs.)	Date of completion of Contract		Remarks indicating reasons for delay if any	Have the goods been functioning Satisfactorily (attach documentary proof)**
					As per contract	Actual		
1	2	3	4	5	6	7	8	9

Signature and seal of the Tenderer

** The documentary proof will be a certificate from the consignee/end user with cross-reference of order no. and date in the certificate duly notarised certification authenticating the correctness of the information furnished. If at any time, information furnished is proved to be false or incorrect, the earnest money and or performance security furnished will be forfeited. Such certificates from a third party or middleman other than actual end user will not be accepted. The satisfactory performance implies working satisfactorily without any complaint since the date of installation, commissioning & handing over to the end user as per the standard format enclosed.

Section – X TENDER FORM

Date _____

**To,
Head (P & CD)
HLL Lifecare Limited
Procurement and Consultancy Division
B-14A, Sector -62, Noida -201307, Uttar Pradesh**

Ref. Your TE document No. _____ dated _____

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. _____, dated _____ (if any), the receipt of which is hereby confirmed. We now offer to supply and deliver _____ (Description of goods and services) in conformity with your above referred document.

If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC clause 5, read with modification, if any, in Section - V – “Special Conditions of Contract”, for due performance of the contract.

We agree to keep our tender valid for acceptance as required in the GIT clause 20, read with modification, if any in Section - III – “Special Instructions to Tenderers” or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any

(Signature with date)

(Name and designation) Duly authorised to sign tender for and on behalf of

SECTION – XI PRICE SCHEDULE**A) PRICE SCHEDULE FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA**

1	2	3	4	5							6
				Price per unit (Rs.)							
Item Sl. No.	Brief Description of Goods (with make & model)	Country of Origin	Quantity (Nos.)	Ex - factory/ Ex - warehouse /Ex-showroom /Off - the shelf (a)	Excise Duty (if any) [%age & value] (b)	Sales Tax/ VAT (if any) [%age & value] (c)	Transportation, loading/ unloading and Incidental costs till consignee's site (d)	Insurance charges for a period including 3 months beyond the date of delivery (e)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at the Consignee's site (f)	Unit Price (at Consignee Site) basis (g) =a+b+c+d+e+f	Total Price (at Consignee Site) basis (Rs.) 4 x 5(g)

Total Tender price in Rupees: _____

In words: _____

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted separately as per Section-XI – Price Schedule C

Name _____

Business Address _____

Place: _____

Signature of Tenderer _____

Date: _____

Seal of the Tenderer _____

B) PRICE SCHEDULE FOR GOODS TO BE IMPORTED FROM ABROAD

1	2	3	4	5											
				Price per unit (Currency)										Unit price on DDP basis at consignee's site	
				Gross FOB price at sea/air port of Lading (inclusive of Agency Commission)	Amount and percentage of Agency Commission **	Net FOB (excluding Agency Commission) (a-b)	Insurance & Freight	Net CIP by Air/ Sea at the port of entry (c+d)	Custom Duty amount as % of Net CIP (amount with CDEC as applicable) **	Custom Clearance & Handling Charges **	Loading/ Unloading, inland transportation, insurance as per Clause 11 of GCC & incidental cost till consignee's site **	Installation commissioning, supervision. Demonstration & training at the consignee's site **	In foreign currency	In Indian Rupees	
(a)	(b)	(c)	(d)												(e)

** to be quoted in Indian Currency

Total price at Consignee's site

(A) In foreign currency : column (4 x e) _____ (In figures and words) plus
 ** (B) In Indian Rupees : column 4 x (b+f+g+h+i) Rs _____ (In figures and words)

Note: -

1. The Tenderer will be fully responsible for the safe arrival of the goods at the consignee site in good condition as per terms of contract.
2. The bidders break up of prices under various columns is for comparison of prices up to delivery of goods at consignee's site for tender evaluation.
3. The quoted price should be supported with original proforma invoice from the foreign manufacturers. The proforma invoice should indicate the percentage of agency commission included in the FOB prices. Indian Agent to be paid in Indian Currency.
4. All the components of the DDP price will be paid by the tenderer. The purchaser will make the payment of DDP price after receipt of goods at consignee's site in good condition as per payment terms in the contract.
5. The prices quoted in foreign currency in column (e) shall be converted in Rupees at the selling rate of exchange applicable on the date of tender opening. The customs duty amount so worked out as percentage of net CIP value in rupees will be taken for evaluation and comparison of tenders
6. The charges for Annual CMC after warranty shall be quoted separately as per Section-XI – Price Schedule C

Name _____

Business address _____

Signature of Tenderer _____

Seal of Tenderer _____

Place: _____

Date: _____

C) PRICE SCHEDULE FOR COMPREHENSIVE MAINTENANCE CONTRACT AFTER WARRANTY PERIOD

1	2	3	4					5
Item Sl. No.	Brief Description of the Goods	Quantity (Nos.)	Comprehensive Maintenance Contract Cost for Each Unit year wise*.					Total Comprehensive Maintenance Contract Cost for 5 (or as specified) Years [3 x (4a+4b+4c+4d+4e)]
			1 st	2 nd	3 rd	4 th	5 th	
			a	b	c	d	e	

* After completion of Warranty period

NOTE:-

1. In case of discrepancy between unit price and total prices, THE UNIT PRICE shall prevail.
2. The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual and labour, after satisfactory completion of Warranty period may be quoted for next 5 (or as specified) years on yearly basis for complete equipment and Turnkey (if any).
3. The cost of CMC may be quoted along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
4. Cost of CMC will be added for Ranking/Evaluation purpose.
5. The payment of CMC will be made as per clause GCC clause 21.1 (D).
6. The uptime warranty will be 98 % on 24 (hrs) x 7 (days) x 365 (days) basis or as stated in Technical Specification of the TE document.
7. All software updates should be provided free of cost during CMC period.
8. The stipulations in Technical Specification will supersede above provisions
9. The supplier shall keep sufficient stock of spares required during Comprehensive Maintenance Contract period. In case the spares are required to be imported, it would be the responsibility of the supplier to import and get them custom cleared and pay all necessary duties.

Name_____

Business Address_____

Signature of Tenderer_____

Seal of the Tenderer_____

Place: _____

Date: _____

D) PRICE SCHEDULE FOR TURNKEY

Item Sl. No.	BRIEF TURNKEY DESCRIPTION OF GOODS	CONSIGNEE CODE	Turnkey price

Note: -

1. The cost of Turnkey as per Technical Specification (Section VII) may be quoted on lump sum along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
2. Cost of Turnkey will be added for Ranking/Evaluation purpose.
3. The payment of Turnkey will be made as per clause GCC clause 21.1 (c).
4. The stipulations in Technical Specification will supersede above provisions

Name_____

Business Address_____

Place: _____

Signature of Tenderer_____

Date: _____

Seal of the Tenderer_____

**SECTION – XII
QUESTIONNAIRE****Fill up the Section XX – Check List for Tenderers and enclose with the Tender**

1. The tenderer should furnish specific answers (alongwith mention of relevant page nos. of tender) to all the questions/issues mentioned in the Checklist. In case a question/issue does not apply to a tenderer, the same should be answered with the remark “not applicable”
2. Wherever necessary and applicable, the tenderer shall enclose certified copy as documentary proof/ evidence to substantiate the corresponding statement.
3. In case a tenderer furnishes a wrong or evasive answer against any of the question/issues mentioned in the Checklist, its tender will be liable to be ignored.

SECTION – XIV
MANUFACTURER’S AUTHORISATION FORM

To,

Head (P & CD)

HLL Lifecare Limited
Procurement and Consultancy Division
B-14A, Sector -62, Noida -201307, Uttar Pradesh

Dear Sir,

Ref. Your TE document No _____, dated _____

We, _____ who are proven and reputable manufacturers of _____ (*name and description of the goods offered in the tender*) having factories at _____, hereby authorise Messrs _____ (*name and address of the agent*) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We further confirm that no supplier or firm or individual other than Messrs. _____ (*name and address of the above agent*) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also hereby extend our full warranty, CMC as applicable as per clause 15 of the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document.

We also confirm that the price quoted by our agent shall not exceed than that which we would have quoted directly.

Yours faithfully,

[Signature with date, name and designation]
for and on behalf of Messrs _____

[Name & address of the manufacturers]

Note: 1. This letter of authorisation should be on the letterhead of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
2. Original letter may be sent.

SECTION – XV

BANK GUARANTEE FORM FOR PERFORMANCE SECURITY/ CMC SECURITY

To
Head of Hospital/Institute/Medical College of ESIC

WHEREAS _____ (Name and address of the supplier) (Hereinafter called “the supplier”) has undertaken, in pursuance of contract no _____ dated _____ to supply (description of goods and services) (herein after called “the contract”).

AND WHEREAS it has been stipulated by you in the said contract that the supplier shall furnish you with a bank guarantee by a scheduled commercial bank recognised by you for the sum specified therein as security for compliance with its obligations in accordance with the contract;
AND WHEREAS we have agreed to give the supplier such a bank guarantee;

NOW THEREFORE we hereby affirm that we are guarantors and responsible to you, on behalf of the supplier, up to a total of. _____ (Amount of the guarantee in words and figures), and we undertake to pay you, upon your first written demand declaring the supplier to be in default under the contract and without cavil or argument, any sum or sums within the limits of (amount of guarantee) as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

We hereby waive the necessity of your demanding the said debt from the supplier before presenting us with the demand.

We further agree that no change or addition to or other modification of the terms of the contract to be performed there under or of any of the contract documents which may be made between you and the supplier shall in any way release us from any liability under this guarantee and we hereby waive notice of any such change, addition or modification.

This guarantee shall remain valid till 2 (two) months beyond the warranty period, i.e up to ----- (indicate date)

.....
(Signature with date of the authorised officer of the Bank)

.....
Name and designation of the officer

.....
Seal, name & address of the Bank and address of the Branch

SECTION – XVI
CONTRACT FORM - A

CONTRACT FORM FOR SUPPLY, INSTALLATION, COMMISSIONING, HANDING OVER, TRIAL RUN, TRAINING OF OPERATORS & WARRANTY OF GOODS

(Address of the Purchaser's/Consignee's office issuing the contract)

Contract No _____ dated _____

This is in continuation to this office's Notification of Award No _____ dated _____

1. Name & address of the Supplier: _____
2. Purchaser's TE document No _____ dated _____ and subsequent Amendment No _____, dated _____ (if any), issued by the purchaser
3. Supplier's Tender No _____ dated _____ and subsequent communication(s) No _____ dated _____ (if any), exchanged between the supplier and the purchaser in connection with this tender.
4. In addition to this Contract Form, the following documents etc, which are included in the documents mentioned under paragraphs 2 and 3 above, shall also be deemed to form and be read and construed as integral part of this contract:

- (i) General Conditions of Contract;
- (ii) Special Conditions of Contract;
- (iii) List of Requirements;
- (iv) Technical Specifications;
- (v) Quality Control Requirements;
- (vi) Tender Form furnished by the supplier;
- (vii) Price Schedule(s) furnished by the supplier in its tender;
- (viii) Manufacturers' Authorisation Form (if applicable for this tender);
- (ix) Purchaser's Notification of Award

Note : The words and expressions used in this contract shall have the same meanings as are respectively assigned to them in the conditions of contract referred to above. Further, the definitions and abbreviations incorporated under clause 1 of Section II – 'General Instructions to Tenderers' of the Purchaser's TE document shall also apply to this contract.

5. Some terms, conditions, stipulations etc. out of the above-referred documents are reproduced below for ready reference:

- (i) Brief particulars of the goods and services which shall be supplied/ provided by the supplier are as under:

Item Sl. No.	Brief description of goods/services	Accounting unit	Quantity to be supplied	Unit Price	Total price	Terms of delivery

Any other additional services (if applicable) and cost thereof: _____

Total value (in figure) _____ (In words) _____

- (ii) Delivery schedule
- (iii) Details of Performance Security
- (iv) Quality Control
 - (a) Mode(s), stage(s) and place(s) of conducting inspections and tests.
 - (b) Designation and address of purchaser's inspecting officer
- (v) Destination and despatch instructions
- (vi) Consignee, including port consignee, if any
- (vii) Warranty clause
- (viii) Payment terms
- (ix) Paying authority

**(Signature, name and address
 of the Purchaser's/Consignee's authorised official)
 For and on behalf of** _____

Received and accepted this contract

(Signature, name and address of the supplier's executive
duly authorised to sign on behalf of the supplier)

For and on behalf of _____

(Name and address of the supplier)

(Seal of the supplier)

Date: _____

Place: _____

CONTRACT FORM – B
CONTRACT FORM FOR COMPREHENSIVE MAINTENANCE CONTRACT

Comprehensive Maintenance Contract No. _____ **dated** _____
 Between _____

(Address of Head of Hospital/Institute/Medical College)
 And _____

(Name & Address of the Supplier)

Ref: Contract No _____ **dated** _____ (Contract No. & date of Contract for supply, installation, commissioning, handing over, Trial run, Training of operators & warranty of goods)

In continuation to the above referred contract

- a) The Contract of Comprehensive Maintenance is hereby concluded as under: -

1	2	3	4					5
Item Sl. No.	BRIEF DESCRIPTION OF GOODS	Quantity (Nos.)	Comprehensive Maintenance Contract Cost for Each Unit year wise*.					Total Comprehensive Maintenance Contract Cost for 5 (or as specified) Years [3 x (4a+4b+4c+4d+4e)]
			1 st	2 nd	3 rd	4 th	5 th	
			a	b	c	d	e	

Total value (in figure) _____ (In words) _____

- b) The CMC commence from the date of expiry of all obligations under Warranty i.e. from _____ (date of expiry of Warranty) and will expire on _____ (date of expiry of CMC)
- c) The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance and labour, after satisfactory completion of Warranty period may be quoted for next 5 (or as specified) years as contained in the above referred contract on yearly basis for complete equipment (including X ray tubes, Batteries for UPS, other vacuummatic parts, _____ & _____) and Turnkey (if any).
- d) There will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- e) During CMC period, the supplier shall visit at each consignee's site for preventive maintenance including testing and calibration as per the manufacturer's service/ technical/ operational manual. The supplier shall visit each consignee site as recommended in the manufacturer's manual, but at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.
- f) All software updates should be provided free of cost during CMC.
- g) The bank guarantee valid till _____ [(fill the date) 2 months after expiry of entire CMC period] for an amount of Rs. _____ [(fill amount) equivalent to 2.5 % of the cost of the equipment as per contract] shall be furnished in the prescribed format given in

Section XV of the TE document, along with the signed copy of CMC within a period of 21 (twenty one) days of issue of CMC failing which the proceeds of Performance Security shall be payable to the Purchaser/Consignee.

- h) If there is any lapse in the performance of the CMC as per contract, the proceeds CMC bank guarantee for an amount of Rs. _____ (equivalent to 2.5 % of the cost of the equipment as per contract) shall be payable to the Consignee.
- i) **Payment terms:** The payment of CMC will be made against the bills raised to the consignee by the supplier on six monthly basis after satisfactory completion of said period, duly certified by the HOD concerned. The payment will be made in Indian Rupees.
- j) **Paying authority:** _____ (name of the consignee i.e. Hospital/ Institute /Medical College's authorised official)

**(Signature, name and address of
Hospital/Institute/Medical College's authorised official)
For and on behalf of _____**

Received and accepted this contract

(Signature, name and address of the supplier's executive
duly authorised to sign on behalf of the supplier)

For and on behalf of _____

(Name and address of the supplier)

(Seal of the supplier)

Date: _____

Place: _____

SECTION – XVII

CONSIGNEE RECEIPT CERTIFICATE

(To be given by consignee’s authorized representative)

The following store (s) has/have been received on said to contain basis in good condition:

1) Contract No. & date : _____

2) Supplier’s Name : _____

3) Consignee’s Name & Address with telephone No. & Fax No. : _____

4) Name of the item supplied : _____

5) No of cartons received which are said:

Which are said to contain the items (List of items in each carton to be given.)

: _____

6) Date of Receipt by the Consignee : _____

7) Name and designation of Authorized Representative of Consignee : _____

8) Signature of Authorized Representative of Consignee with date : _____

9) Seal of the Consignee : _____

SECTION – XVIII
Proforma of Final Acceptance Certificate by the Consignee

No _____

Date _____

To

M/s _____

Subject: Certificate of commissioning of equipment/plant.

This is to certify that the equipment(s)/plant(s) as detailed below has/have been received in good conditions along with all the standard and special accessories and a set of spares (subject to remarks in Para no.02) in accordance with the contract/technical specifications. The same has been installed and commissioned.

- (a) Contract No _____ dated _____
- (b) Description of the equipment(s)/plants: _____
- (c) Equipment(s)/ plant(s) nos.: _____
- (d) Quantity: _____
- (e) Bill of Loading/Air Way Bill/Railway
 Receipt/ Goods Consignment Note no _____ dated _____
- (f) Name of the vessel/ Transporter: _____
- (g) Name of the Consignee: _____
- (h) Date of commissioning and proving test: _____

Details of accessories/spares not yet supplied and recoveries to be made on that account.

Sl. No.	Description of Item	Quantity	Amount to be recovered
---------	---------------------	----------	------------------------

The proving test has been done to our entire satisfaction and operators have been trained to operate the equipment(s)/plant(s).

The supplier has fulfilled its contractual obligations satisfactorily ## or

The supplier has failed to fulfil its contractual obligations with regard to the following:

He has not adhered to the time schedule specified in the contract in dispatching the documents/drawings pursuant to 'Technical Specifications'.

He has not supervised the commissioning of the equipment(s)/plant(s) in time, i.e. within the period specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment(s)/plant(s).

The supplier as specified in the contract has not done training of personnel.

The extent of delay for each of the activities to be performed by the supplier in terms of the contract

is _____.

The amount of recovery on account of non-supply of accessories and spares is given under Para no.02.

The amount of recovery on account of failure of the supplier to meet his contractual obligations is _____ (here indicate the amount).

Signature

Name:

Designation with stamp

Explanatory notes for filling up the certificate:

- He has adhered to the time schedule specified in the contract in dispatching the documents/drawings pursuant to 'Technical Specification'.
- He has supervised the commissioning of the equipment(s)/plant(s) in time, i.e. within the time specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment(s)/plant(s).
- Training of personnel has been done by the supplier as specified in the contract
- In the event of documents/drawings having not been supplied or installation and commissioning of the equipment(s)/plant(s) having been delayed on account of the supplier, the extent of delay should always be mentioned in clear terms.

SECTION – XIX

DETAILS OF SHIPPING ARRANGEMENT FOR LINER CARGOES IN RESPECT OF C&F/CIF/TURNKEY/F.O.R CONTRACTS FOR IMPORTS

1. SHIPMENT FROM PORTS OF U.K INCLUDING NORTHERN IRELAND (ALSO EIRE), FROM THE NORTH CONTINENT OF EUROPE (GERMANY, HOLLAND, BELGIUM, FRANCE, NORWAY, SWEDEN, DENMARK, FINLAND AND PORTS ON THE CONTINENTAL SEABOARD OF MEDITERRANIAN (I.E. FRENCH WESTERN ITALIAN PORTS), TO PORTS IN INDIA.

The Seller should arrange shipment of the goods by vessels belonging to the member lines of the India-Pakistan-Bangladesh Conference. If the Seller finds that the space on the 'Conference Lines' vessels is not available for any specific shipment, he should take up with India-Pakistan-Bangladesh Conference, Conferity House, East Grinstead, Sussex (UK), for providing shipping space and also inform the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSCHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159)

The Seller should arrange shipment through the Government of India's Forwarding Agents, M/s Schenker & Co., 2000-Hamburg (Cable: SCHENKER CO., HAMBURG) OR obtain a certificate from them to the effect that shipment has been arranged in accordance with instructions of the Ministry of Surface Transport, (TRANSCHART), New Delhi.

2. SHIPMENT FORM PORTS OF U.K. INCLUDING NORTHERN

Goods under this contract would be shipped by the national shipping companies of the Contracting Parties operating bilateral shipping service and vessels under the flag of third countries in accordance with the Agreement between the Government of German Democratic Republic and the Government of the Republic of India in the Field of Merchant Shipping signed on 9.1.1979, as amended up-to-date.

3. ISHIPMENT FROM ADRIATIC PORTS OF EASTERN ITALY AND YUGOSLAVIA

The seller should arrange shipment of the goods by vessels belonging to the following Indian member lines;

1. The Shipping Purchaser of India Ltd.
2. The Scindia Steam Navigation Co., Ltd
3. India Steamship Co., Ltd

For the purpose of ascertaining the availability of suitable Indian vessels and granting dispensation in the event of their non-availability, the Seller should give adequate notice about the readiness of each consignment from time to time at least six weeks in advance of the required position to M/s Schenker & Co. 2000 HAMBURG (Cable: SCHENKER CO., HAMBURG) and also endorse a copy thereof to the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSCHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159)

The seller should arrange shipment through the Government of India's Forwarding Agents M/s Schenker & Co. 2000 HAMBURG (Cable: SCHENKER CO., HAMBURG) or obtain certificate from them to the effect that shipment has been arranged in accordance with the instructions of the Ministry of Surface Transport, (TRANSCHART), New Delhi.

4. SHIPMENT FROM POLAND & CZECHOSLOVAKIA

- (i) IMPORTS FROM POLAND

Shipment under this contract would be made by the National flag lines of the two parties and vessels of the third flag conference lines, in accordance with the agreement between the Govt. of the Republic of India and the Govt. of the Polish People's Republic regarding Shipping Co-operation dated 27.6.1960 as amended up-to-date.

(ii) IMPORTS FROM CZECHOSLOVAKIA

Goods under this contract would be signed by the National flag lines of the two parties and vessels of the third flag conference lines, in accordance with the Agreement Co-operation in shipping between India and Czechoslovakia signed on 3.11.1978 and ratified on 19.12.1979, as amended up-to-date.

Shipping arrangement should be made by the Sellers in consultation with Resident Representative of the Indian Shipping Lines in Gdynia, Co., Morska Agencja W. Gdyniul, Pulaskiego 8, P.O. Box 246, Gdynia (Poland) – Telex : MG PL. 054301, Tel.: 207621, to whom details regarding contract number, nature of cargo , quantity, port of lading, discharging, name of Government consignee, expected date of readiness of each consignment etc. should be furnish at least six weeks in advance of the required position, with a copy thereof endorsed to the Shipping Co-ordination Officer, Ministry of Surface Transport, (Chartering Wing), New Delhi, (Cable: TRANSHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159)

5. SHIPMENT FROM U.S.S.R

Shipment under this contract should be made in accordance with the agreement between the Government of the Republic of India and the Government of U.S.S.R on Merchant Shipping 1976, as amended up-to-date, by vessels of Indo-Soviet shipping Service.

6. SHIPMENT FROM JAPAN

The shipment of goods should be made of India vessels to the maximum extent possible subject to the minimum of 50%.

The Seller should arrange shipment of the goods in consultation with the Embassy of India in Japan, Tokyo to whom details regarding contract number, nature of cargo, quantity, port of loading/discharge, name of Govt. consignee, expected date of readiness of each consignment etc. should be furnished at least six weeks in advance of the required position.

Note: The copies of such contracts are to be endorsed both to the Attached (commercial) embassy of India in Japan, Tokyo, and the shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi.

7. SHIPMENT FROM AUSTRALIA, ALGERIA, BULGARIA, ROMANIA, EGYPT

The Seller shall arrange shipment of the goods by Indian flag vessels to the maximum extent possible subject to a minimum of 50 %. For the purpose of ascertaining the availability of suitable Indian vessels, the seller shall give adequate notice of not less than six weeks about the readiness of each consignment to the Shipping Purchaser of India Ltd., SHIPPING HOUSE, 245, Madame Cama Road, Bombay – 400 021 (CABLE: SHIPINDIA BOMBAY) and also endorse a copy thereof to the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159)

8. SHIPMENT FROM PAKISTAN

The shipment of cargoes should be made by Indian vessels to the maximum extent possible subject to a minimum of 50 %.

Shipment arrangement should be made by the sellers in consultation with M/s Mogul Line Ltd., 16-Bank Street, Fort, Bombay – 400023 (Cable: MOGUL BOMBAY; Telex: 011 – 4049 MOGUL), to whom, details regarding contract number, nature of cargo, quantity, port of lading discharging, name of government consignee, expected date of readiness of each consignment etc. should be furnish at least six weeks in advance of the required position, with a copy thereof endorsed to the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSCHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159)

9. SHIPMENT FROM U.S ATLANTIC & GULF PORTS

The Seller should arrange shipment of the goods by vessels belonging to the member lines of the India – Pakistan – Bangladesh – Ceylon and Burma Outward Freight Conference. If the Seller finds that the space of the 'Conference Lines' vessels is not available for any specific shipment he should take up with India – Pakistan- Bangladesh – Ceylon and Burma Outward Freight Conference, 19, Rector Street, New York, N.Y. 10006 USA, for providing shipping space and also inform the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSCHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159)

10. SHIPMENT FROM ST. LAWRENCE AN EASTERN CANADIAN PORTS

The Seller should arrange shipment of the goods by vessels belonging to the following shipping lines;

1. The shipping Purchaser of India Ltd.
2. The Scindia Steam Navigation Co., Ltd

If the Seller finds that the space in the vessels of these Lines is not available for any particular consignments, he should inform the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSCHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159) immediately so that dispensation from the shipping lines concerned to use alternative lifting may be sought.

11. SHIPMENT FROM WEST COAST PORTS OF U.S.S CANADA AND OTHER AREAS NOT SPECIFICALLY MENTIONED ABOVE

The Seller should arrange shipment of the goods by Indian vessels to the maximum extent possible subject to a minimum of 50 %. For the purpose of ascertaining the availability of suitable Indian vessels and granting dispensation in the event of their non-availability, the Seller should furnish the details regarding contract number, nature of cargo, quantity, port of lading, discharging, name of government consignee, expected date of readiness of each consignment etc. to the Shipping Co-ordination Officer, Ministry of Surface Transport, New Delhi, (Cable: TRANSCHART, NEW DELHI, Telex: VAHAN IN – 031 – 61157, 61158, 61159) at least six weeks in advance of the required position.

B) BILLS OF LADING:

- (i) C.I.F./C&F/TURNKEY SHIPMENTS

The Bills of lading should be drawn to indicate Shipper and 'Consignee' as under:

SHIPPER: The C.I.F (C&F)/TURNKEY SUPPLIERS concerned.

CONSIGNEE: As per consignee's particulars in the contract (The name and address of the 'Port Consignee' and 'Ultimate' both should be indicated).

(ii) F.O.R SHIPMENTS

The Bills of lading should be drawn indicating shipper Consignee as under:

SHIPPER: The F.O.R suppliers Concerned

CONSIGNEE: Supplier's Indian Agent on order

Note:

1. Moreover the name of the 'Purchaser' and 'Ultimate' Consignee should appear in the body of the Bills of Lading as the 'Notify' or as a remark.
2. Two non-negotiable copies of the Bills of Lading indicating the freight amount and discount, if any allowed, should be forwarded to The Shipping Co-ordination Officer, Ministry of surface Transport (Chartering Wing), New Delhi after the shipment of each consignment is effected.
3. The seller should avoid the use of over-aged vessels for the shipment of the goods under the contract and if so used the cost of additional. Insurance, if any, shall be borne by the seller.

SECTION – XX

CHECKLIST

Name of Tenderer:

Name of Manufacturer:

Sl. No.	Activity	Yes/ No/ NA	Page No. in the Tender document	Remarks
1.	Have you enclosed EMD of required amount for the quoted schedules?			
2(a).	Have you enclosed clause-by-clause technical compliance statement for the quoted goods vis-à-vis the Technical specifications?			
2(b).	In case of Technical deviations in the compliance statement, have you identified and marked the deviations?			
3.	Have you kept validity of 120 days from the Techno Commercial Tender Opening date as per the TE document?			
4(a).	Have you enclosed duly filled Tender Form as per format in Section X?			
4(b).	Have you enclosed Power of Attorney/ Authorisation in favour of the signatory?			
5.	Have you submitted manufacturer's authorization as per Section XIV?			
6.	Have you submitted the certificate of incorporation?			
7(a).	In case of Indian Tenderer, have you furnished Income Tax Account No. as allotted by the Income Tax Department of Government of India?			
7(b).	In case of Foreign Tenderer, have you furnished Income Tax Account No. of your Indian Agent as allotted by the Income Tax Department of Government of India?			

Sl. No.	Activity	Yes/ No/ NA	Page No. in the Tender document	Remarks
8.	Have you intimated the name and full address of your Banker (s) along with your Account Number			
9.	Have you furnished Annual Report (Balance Sheet and Profit & Loss Account) for last three years prior to the date of Tender opening?			
10.	Have you submitted the Quality Control Requirements as per Proforma given in Section VIII of TE document?			
11.	Have you accepted delivery period as per TE document?			
12.	Have you accepted the terms of delivery as per 'DDP at consignee site basis'?			
13.	Have you accepted the warranty/CMC as per TE document?			
14.	Have you accepted all terms and conditions of TE document?			
15.	Have you fully accepted payment terms as per TE document?			
16(a)	Have you submitted prices of goods, turnkey (if any), CMC etc. in the Price Schedule as per Section XI?			
16(b)	Have you submitted satisfactory end user performance certificate as per the Proforma for performance statement in Sec. IX of TE document in respect of all orders?			
17.	Have you submitted copy of the order(s) against the above end user certificate (s)?			
18.	Have you furnished documents establishing your eligibility & qualification criteria as per TE documents?			

N.B.

1. All pages of the Tender should be page numbered and indexed.
2. The Tenderer may go through the checklist and ensure that all the documents/confirmations listed above are enclosed in the tender and no column is left blank. If any column is not applicable, it may be filled up as NA.
3. It is the responsibility of tendered to go through the TE document to ensure furnishing all required documents in addition to above, if any.

(Signature with date)

(Full name, designation & address of the person duly authorised sign on behalf of the
Tenderer)

For and on behalf of

(Name, address and stamp of the tendering firm)

Section – XXI**Consignee addresses**

Consignee Code	Consignee Address	Telephone No. & Fax No.
Bapu Nagar	Medical Superintendent ESIC Hospital Bapunagar, Ahmedabad	(079) 22743935- 22745770, 22741866

NB: The purchaser/consignee will ensure timely issue of CDEC, Octroi Exemption Certificates, Road Permits & Entry Tax Exemption Certificates, wherever applicable, to the suppliers.