

SPECIFICATION OF MULTIPLE VITAL SIGN MONITOR (MULTIMONITOR)

1. It should monitor SPO2, pulse rate, ECG and NIBP with provision of suitable audio visual Alarms. Accessories for these parameters to be provided for both adult & paedric patients.
2. It should have valid CE or equivalent certification.
3. It should function on rechargeable internal battery also.
4. It should have bright and highly visible screen display.
5. Should have capability to monitor at least 6 real time wave forms along with related numerical parameters on a single configurable screen.
6. It should have capability to monitor the above parameters in adult and Pediatric patients.
7. It should have a graphical/tabular trending facility.
8. Should submit relevant evidence of IEC or equivalent safety certification.
9. Wall mounting facility device to be provided.
10. 20 disposable ECG electrodes are to be provided.

Dr. S.K. Aggarwal
HOD-Medicine
(ESIH-Basai)

Dr. Uma Verma
HOD-Medicine
(ESIH-Rohini)

Dr. R.K.Duggal
Spl.-Medicine

Dr. Subhash Chhokra
IHS