

Amendment No. 1 Dated 03.02.2011
Ref: Tender No. HLL/PCD/ESIC-44/10-11

The following amendments are issued in respect of list of requirements:-

Tender No.: HLL/PCD/ESIC-44/10-11

In the list of requirement 1 item of Rohini added:

S.NO.	ITEM DESCRIPTION	TOTAL QNTY.	EMD AMOUNT Rs.
1	Air condition Mortuary Cabinet (04 bodies)	1(Rohini)	10000

TECHNICAL SPECIFICATION

Air Condition Mortuary Chamber Fully Automatic (04) Bodies

- **Capacity : Four Bodies in two rows**
- **Temperature Range : -2 degree to 5 degree C**
- **Controller : Microprocessor Based**
- **Interior Panel : Made of Stainless Steel**
- **Outer Panel : Made of Stainless Steel**
- **Door : Hinged Insulated Doors with magnetic gaskets**
- **Insulation : Polyurethane foam with thickness of 80 – 100 mm**
- **Body Trays : Stainless Steel with Telescopic Mechanism**
- **Compressor : Heavy Duty low noise & minimal vibration**
- **Condenser : Efficient with Auto – condensing evaporating system**
- **Air Circulation: Forced Air Circulation**
- **Alarms : Audio – Visual Alarms to warn high or low temperature**
- **Supplies : 220 V/ 50Hz**
- **Illumination : Door operated illumination system in the chamber**
- **Accessories : Temperature recorder, data logger, voltage safety System, with battery backup.**