

I.G. ESI HOSPITAL, JHILMIL, DELHI-95.

MEDICAL STORE

DM(H)-JMD-46(17)/X-ray/2007

Dated: 17/11/2009


LIMITED TENDER FOR MISCELANEOUS ITEMS OF X-RAY DEPARTMENT.

To,

M/s-----

(To be returned duly completed latest by 18-11-2009, till 1.00 P.M. otherwise tender shall not be accepted).

Medical Superintendent, I.G. ESI Hospital, Jhilmil, Delhi invites Limited tender for Miscellaneous items of **X-RAY Department**, Terms and Conditions and specifications detailed at page no. 1 to 3 enclosed. Tender should be submitted in sealed envelope & subscribed a "Tender for (Name of the Equipment)" in the office of M.S. I.G. ESI Hospital, Jhilmil, Delhi in the tender box placed in the chamber of Dy. Medical Superintendent on dated 17/11/09 from 9 A. M. to 3 P. M. and on dated 18-11-2009 till 1.00 P. M. In case the Tender is sent by post it must reach in the office of M.S. by dated 18/11/2009 up to 1.00 P.M. Proof of postage won't be considered as a claim for timely submission of tender. The tender shall be opened on dated 18/11/2009 at 2.30 P. M. in Hospital Committee Room in presence of tenderers or their representative who wish to be present. In case 18/11/2009 is declared holiday then bid will be opened on next working day at the same time and venue.



Dy. Medical Superintendent.

TERMS AND CONDITIONS

1. Sealed Limited tenders are invited from the reputed firms as well as authorized distributors/stockiest for supply of miscellaneous items of X-ray Deptt. the specifications for same are enclosed as per ANX-I.
2. Any conditional acceptance is liable to rejection of tender.
3. EMD of Rs. ~~3000=00~~ only shall be deposited by the tenderer (except Govt. Agencies) by enclosing Demand Draft/Bankers Cheque only in favour of 'ESI Fund A/C No. 1', IG ESI Hospital, Jhilmil, Delhi which shall be refundable to bidders without any interest after finalization of tender
4. **All pages of tender document submitted should be numbered & signed by authorized signatory of participating firm with seal.**
5. Working demonstration/sample of the quoted item shall be provided whenever asked for.
6. Successful tenderer shall supply the material as per supply order on the rates quoted by them and approved by the competent authority of this hospital.
7. Taxes, if applicable, shall be as notified by Govt.
8. Tenderer must provide Delhi address, if any, along with the telephone no. and fax no. with tender for all correspondence.
9. Rates should be quoted for **F.O.R. destination including for installation charges.**
10. The rates should not be higher than quoted in any other organization/institution (Certificate may be enclosed).
11. A certificate about satisfactory performance of the equipment duly authenticated by other existing users of the equipment
12. The price of AMC/CMC for five year after expiry of warranty/guarantee period should be given in the tender (if applicable)
13. Any other accessories required for equipment may also be quoted.
14. Catalogue/literature make of model should be enclosed.
15. Validity of quotation shall be for one year. Firm quoting validity less than one year will be rejected.
16. Warranty period will be commencing from the date of installation.
17. Payment shall be released after satisfactory installation/inspection of the item.
18. The successful tenderer shall supply the articles within six weeks from the date of issue of supply order.
19. **Tenders in sealed envelope super scribed as "Limited Tender for supply of miscellaneous items of X-Ray Deptt can be dropped in the tender box kept in chamber of Dy. Medical Superintendent of this Hospital, on 17/11/2009 between 9.00 A.M. to 3.00 P.M. and on dated 18/11/2009 upto 1.00 P.M. No tender thereafter shall be accepted. The tender shall be opened on 18/11/09 at 2.30 PM in the Committee Hall of this hospital in presence of tenderers or their representatives who wish to be present. In the event of 18/11/09 being declared a holiday, the tender will be opened on next working day at same time and place.**
20. If the tender is sent by post/given personally it must be received in office of Medical Superintendent on dated 18/11/09 by 1.00 PM Proof of postage won't be considered as evidence for timely submission of tender.
21. All disputes subject to Jurisdiction of Delhi only.
22. M.S. reserves the right to accept or reject any or all tenders without assigning any reason thereof.



(DR.Renuka Kaul Nigam)
DY.MEDICALSUPERINTENDENT

6. Cassette pass Box (1 No.)

- Should be of mild steel of 16/18 gauges with 4 door model. The doors should be lined with 1 to 1.5 mm lead lining.
- The base and sides of the pass box should be lined with vinyl mat of 5 mm thickness to protect cassettes from impact while passing through the box.

7. Chest stand (1 No.)

- Stand should be sturdy strongly jointed with the base. The base should be provided with level adjusters and should not sag.
- The cassette holder should be made of aluminum and mild angle iron and should have heavy duty clamps/knobs and other devices with long durability.

8. View box (2 No.)

- For single 14" x 17" X-ray film.
- Made of suitable opaque acrylic sheet.
- Viewing sheet of good quality for uniform spread of light.
- With two tube lights of 2 feet each.
- Switch below viewing screen placed horizontally.
- Film should be held by Roller clips firmly.


DY. MEDICAL SUPERINTENDENT

specification of the required items (X-ray Department)

1. Hangers with clips (63 Nos. of different sizes)

- Should be made of stainless steel rods, rust free with ends of rods covered with vinyl rubber knobs.
- Should have strong joints in between rods making smooth hard pressed T
- Should have strongly joints clips with perfect locking cups which does no allow the film to be pulled out without opening clips & not pricking fingers & without needing the film to be pulled out without opening clips & not pricking fingers & without needing adjustment at all.

Qty. required – 14" x 17" = 07 Nos.

14" x 14" = 20 Nos.

12" x 15" = 14 Nos.

10" x 12" = 11 Nos.

8" x 10" = 09 Nos.

6.5" x 8.5" = 02 Nos.

2. Cassette with Screen (18 Nos. of different sizes)

- Should be push lock type (PLT)
- Should have reinforced corners and steel hinges and locks to survive fall on corner from 3 feet height without destroying the cassette.
- Should have good quality felt and desing to cover diagnostically important area.
- Should be pre fitted with screens of speed K R 4 (factory fitted)

Qty. required - 14" x 17" = 03 Nos.

14" x 14" = 03 Nos.

12" x 15" = 05 Nos.

10" x 12" = 03 Nos.

8" x 10" = 03 Nos.

6.5" x 8.5" = 01 Nos.

3. Lead coat (3 Nos.)

- Should be light weight, of large size, uniformly (non manually) mixed lead vinyl sheets of even thickness, flexible and not prone to cracking and pin holes. The lead coat should be made of more than 2 vinyl sheets.
- Should have lead equivalent of 0.5 mm
- Appropriate hanger should be provided with lead coat for storage.
- Should be approved by B.A.R.C/A.E.R.B

4. Lead divider (2 Nos.) of different sizes.

- Should be made of lead sheet covered with suitable metal sheets for heavy duty handling.
- Should be suitable for film sizes 17" x 14" and 12" x 12" (one piece for each size.)

5. Film Clips (60 Nos.)

- Should be off stainless steel with strong tension & rust free
- Should not prick fingers.
- Should have two sharp teeth to hold the film strongly without needing adjustment.