



कर्मचारी राज्य बीमा निगम  
Employees' State Insurance Corporation  
क. रा. बी. नि. अस्पताल E.S.I.C. HOSPITAL  
के. के. नगर, चेन्नई K. K. NAGAR, Chennai - 600 078  
Dir. : 24893714, Fax : 24891094  
Website : [www.esic.nic.in](http://www.esic.nic.in) , [www.esichennai.org](http://www.esichennai.org)  
E-Mail: [deanmckkn.tn@esic.in](mailto:deanmckkn.tn@esic.in)



Tender No: 512-U-P&B/Equipments-Paediatrics-2018-2019

Dated: 11.07.2018

**CORRIGENDUM**

**Sub: Due date for submission of bid is extended and modification/Additional in the specifications in respect of the e-tender Ultrasound with 2D Echo Colour Doppler (1No.) (e-Tender ID:1450) -Reg**

In reference to E-Tender for procurement, installation and commissioning of **Ultrasound with 2D Echo Colour Doppler (1No.) (e-Tender ID:1450)** for ESIC PGIMSR Medical College and Model Hospital Chennai-78, the competent authority has approved the following additions /modifications and changes in due date for submission of bid are detailed below:

**SPECIFICATIONS FOR ULTRASOUND WITH 2D ECHO COLOUR DOPPLER (1NO.) (E-TENDER ID:1450)**

S.NO	Specifications as floated in e-tender	Specifications as Amended during the Pre Bid Meeting
1	Monitor color display LCD 13 inches or more diagonally	<b>Monitor color display LCD 21 inches or more diagonally</b>
20	System should have two transducer ports on-board.	<b>System should have three active transducer ports.</b>
25	System should be supplied with phased array transducers of frequency 3-11 MHz with footprint of 17 x 13 mm	<b>System should be supplied with phased array transducers of frequency 3-8 MHz for paediatric echo</b>
•	System should be supplied with convex array transducers of frequency 4-9 MHz	<b>System should be supplied with phased array transducers of frequency 5-11 MHz with small footprint for neonatal echo</b>
•	Ultrasound probes should include convex linear probes for echo and small parts	<b>System should be supplied with micro convex transducers of frequency 4 -8 MHz for neonatal brain and paediatric abdomen</b>
<b>During Pre bid meeting: the additional changes/Specifications also included for the equipment</b>		<b>All the above probes with +/- 1 MHz frequency range are acceptable. Wider frequency range is preferable</b>
		<b>System should have 20 lakh or more digital processing system</b>
		<b>System should have US FDA and CE European approval</b>

The competent authority has also approved for extending the **due date** for submission of bids and the details are given below:

FOR	<ol style="list-style-type: none"><li>1.The interested bidders should submit the e-tender through online only at e- procurement portal <a href="https://esictenders.eproc.in">https://esictenders.eproc.in</a> till 12.07.2018 Upto 11.00 A.M.</li><li>2) The period for submission of e-Tender EMD payment in the mode of Demand draft to this Hospital : 12.07.2018 upto 11.00 AM</li><li>3)Date and time of opening of Technical Bid: 12.07.2018 at 11.30AM.</li></ol>
TO BE READ AS	<ol style="list-style-type: none"><li>1.The interested bidders should submit the e-tender through online only at e- procurement portal <a href="https://esictenders.eproc.in">https://esictenders.eproc.in</a> till 01.08.2018 Upto 11.00 A.M.</li><li>2) The period for submission of e-Tender EMD payment in the mode of Demand draft to this Hospital : 01.08.2018 upto 11.00 AM</li><li>3)Date and time of opening of Technical Bid: 01.08.2018 at 11.30AM.</li></ol>

All other terms and conditions of the said e-tender remains same. All bidders are requested to check further notifications / updates if any , on the websites : [esictenders.eproc.in](https://esictenders.eproc.in), [www.esic.nic.in](http://www.esic.nic.in), [www.esichennai.org](http://www.esichennai.org).

*S. Ananya*  
11/7/18

DEAN  
ESIC PGIMSR MEDICAL COLLEGE  
AND MODEL HOSPITAL, CHENNAI